



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

## CYRIX

HEALTHCARE PVT LTD

**No. :** 227794

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility : <u>D.M.C</u></p> <p>Address : <u>Chambakkara</u> <u>Ernakulam</u></p> <p>Ph : <u>9744166574</u></p>	<p>Call Registration Date : <u>08/08/2024</u></p> <p>Caller ID : <u>128200</u></p> <p>Date of Visit : <u>09/08/2024</u></p> <p>Asset No. : <u>0742687</u></p> <p>EQPT Name : <u>Weighing Machine</u></p> <p>Manufacture : <u>SKNOL</u> Model : <u>7227R</u></p> <p>S. No. : <u>NA</u> Dept. : <u>Sub center</u></p>
--	---

**Service Classification :** Breakdown Call  PMS  Calibration  Cust.Training

**Problem Identified :** Equipment not working

**Action Taken :** checked and found that the main board, load sensor and battery defective. Need to Replace the Spares for further checking and working condition of the equipment. Engined the spare with OEM

Completed  Date : 09/08/2024 Time : 12:00pm Spare Required

Spare Replaced  Requested


No.	Description	Qty.	Part Number	PR Number
1.	<u>NA</u>	<u>NA</u>	<u>NA</u>	<u>NA</u>
2.				
3.				

Cyrix Engineer	Date	Start Time	End Time
<u>Jithin.s.kumar</u>	<u>09/08/2024</u>	<u>11:00 AM</u>	<u>12:00PM</u>





**Customer Remark** NIL      Completed       Pending

**Service Engineer Name :** Jithin.s.kumar  
**Signature :** [Signature]  
**Date :** 09/08/2024  
**Contact Number :** 7012355543

**Customer Name :** Renu Raghavan  
**Signature :** [Signature]  
**Date :** 09/08/24  
**Contact Number :** 9744166574  
**Designation :** B SO  
**Hospital Seal :**





PEROFORMA RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER) BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)	
Name of Hospital: UPHC Chambakkara	Name of District: Ernakulam
Name of Equipment: Weighing Machine	Equipment ID & Barcode: 128200& 0742687
Make: SKNOL	Model: 7227R
Serial Number: NA	Warranty details: NA
Date of purchase/ Year of manufacture/ Installation Date: 22-02-2021	Present status of the equipment: Not Working
Date of breakdown: 08-08-2024 (Toll Free) (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: NA
Action taken: Checked and found that main board, load sensor and battery defective. Need to replace these spares for further checking and working condition of the equipment. Enquired spare with OEM	
Recommendations for repair (required service details): Not recommending for repair	
Cost of spares (specify parts and cost): NA	
# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA	Asset Value: 1550/-
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix Service report attached.	
Reasons for recommending the equipment as BER: Checked and found that main board, load sensor and battery defective Machine installed on 22-02-2021 and covered up to 3+ years. Quotation not attached since the spares are not available in the market. So recommending the equipment for condemnation.	
 Name & Signature of CYRIX Authority with date 17/10/24	
Remarks of Junior Consultant (Biomedical) NHM: main board, load sensor and battery damaged. Yearwise BER criteria met. But not attached Quot/ Spare engineering details. BER submitted on 16/10/24 after 30 days. Not recommended as per Kmscl BER criteria. Final decision up to Kmscl.	
Recommended for BER (Yes/ No): No, Submitted after 30 days / no Quot attached	
Date: 17/10/24	
Date: Medical Officer (i/c)	 
	 <b>CONSULTANT</b> <b>Bio-Medical Engineer</b> <b>National Health Mission (NHM)</b> <b>Aroyyakulam Ernakulam</b> Signature of Superintendent /

Medical officer in charge  
UPHC Chambaakkara



# STOCK REGISTER

WEIGHING MACHINE & HEIGHT MACHINE


Date	Name Received from or Issued to	Inv. No or issue	Rate		Receipts	Issues	Balance	Remarks
			Rs.	P.				
21/6/2016	NHM (Infant weighing machine)				1		1	
21/6/2016	Issued to JPHNS					1	0	
20/2/19	NHM Adult				2		2	
1/3/19	Medi Tools (Mechanical Counters Model 125) Adult	2228	1280	1	1		1	
17/2/21	Issued to Kzhilcatti				0	1	0	
1/3/19	Medi Tools (Stature Meter)	2228	400	1	1		1	
2/5/19	Issued to PHC					1	0	
29/2/2020	Welsa medicine & Wellness centre Paediatric weighing machine	140/19/ 8P/6878	5199	1	1		1	
1/3/2020	Issued to JPHNS					1	0	
22/2/21	Sevana Surgical (Stature Meter)	1326	385/-		5		5	
22/2/21	Issued to JPHNS					5	0	
22/2/21	Sevana Surgical (Infant Weighing machine)	1326	1550/-				5	
			Total					
			7750/-					
22/2/21	Issued to JPHNS					5	0	



Medical officer in charge  
UPHC ChampaKKara

*(Signature)*





Medical officer in charge  
UPHC Champakkara

*[Handwritten signature]*