



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. :
227795

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility : UPHC</p> <p>Address : Chambakkare Ernakulam</p> <p>Ph : 9744166574</p>	<p>Call Registration Date : 08/08/2024</p> <p>Caller ID : 128267</p> <p>Date of Visit : 09/08/2024</p> <p>Asset No. : 0742579</p> <p>EQPT Name : Weighing Machine</p> <p>Manufacture : Sknol Model : Ncr</p> <p>S. No. : Ncr Dept. : OP</p>
---	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Equipment not working

Action Taken : checked and found that the display and main board defective. Need to replace these spares for further checking and working condition of the equipment. Requested the spare with OEM

Completed Date : 09/08/2024 Time : 11:00AM Spare Required

Spare Replaced Requested

No.	Description	Qty.	Part Number	PR Number
1.	NA	NA	NA	NA
2.				
3.				

Cyrix Engineer	Date	Start Time	End Time
Arthun-s. Kumar	09/08/2024	10:00AM	11:00AM


Customer Remark : N/A
 Completed Pending

<p>Service Engineer Name : Arthun-s. Kumar</p> <p>Signature : [Signature]</p> <p>Date : 09/08/2024</p> <p>Contact Number : 7012355547</p>	<p>Customer Name : Renu Raghavan</p> <p>Signature : [Signature]</p> <p>Date : 09/08/24</p> <p>Contact Number : 9744166574</p> <p>Designation : DSO</p> <p>Hospital Seal :</p>
---	---



Medical officer in charge
IIPHC Chambakkara

PEROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: UPHC Chambakkara	Name of District: Ernakulam
Name of Equipment: Weighing Machine	Equipment ID & Barcode: 128267& 0742579
Make: SKNOL	Model: NA
Serial Number: NA	Warranty details: NA
Date of purchase/ Year of manufacture/ Installation Date: 22-02-2021	Present status of the equipment: Not Working
Date of breakdown: 08-08-2024 (Toll Free) (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: NA
Action taken: Checked and found that the display and main board defective. Need to replace these spares for further checking and working condition of the equipment. Enquired spare with OEM	
Recommendations for repair (required service details): Not recommending for repair	
Cost of spares (specify parts and cost): NA	
# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA	Asset Value: 1550/-
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrinx Service report attached.	
Reasons for recommending the equipment as BER: Checked and found that the display and main board defective Machine installed on 22-02-2021 and covered up to 3+ years. Quotation not attached since the spares are not available in the market. So recommending the equipment for condemnation.	
Name & Signature of CYRIX Authority with date <i>Arthun S. Kumard</i> <i>Arthun</i> 17/10/24	
Remarks of Junior Consultant (Biomedical) NHM: Main display board defects reported BER year wise criteria met. But not attached Quot/ spare enquiry. submitted on 16/10/24 after 30 days. Not recommended as per RUSCI BER criteria.	
Recommended for BER (Yes/ No): No after 30 days / no Quot/ spare checked	CONSULTANT Bio-Medical Engineer National Health Mission (NHM) Aroyyakeralam Ernakulam Signature of Superintendent /
Date: 17/10/24	
Date: Medical Officer (i/c)	Signature of Superintendent / 

Medical officer in charge
UPHC Champakkara

STOCK REGISTER

WEIGHING MACHINE & HEIGHT MACHINE

Date	Name Received from or Issued to	Inv. No or Issue	Rate		Receipts	Issues	Balance	Remarks
			Rs.	P.				
21/6/2016	NHM (Infant weighing machine)				1		1	
21/6/2016	Issued to JPHNS					1	0	
20/2/19	NHM Adult				2		2	
1/3/19	Medi Tools (Mechanical Counlers Model 125) Adult	2228	1280	1	1		1	
17/2/21	Issued to Vazhikatti				0	1	0	
1/3/19	Medi Tools (Statue Melia)	2298	400	1	1		1	
2/3/19	Issued to PHC					1	0	
29/2/2020	Welsa Medicine & Wellness Centre Paediatric Weighing Machine	140/19/ 8P/618-18	5199	1	1		1	
1/3/2020	Issued to JPHNS					1	0	
22/2/21	Sevana Surgical (Statue Melia)	1326	385/-		5		5	
22/2/21	Issued to JPHNS		Total 1925				5	
22/2/21	Sevana Surgical (Digital Weighing Machine)	1326	1550/-				5	
22/2/21	Issued to JPHNS		Total 7750/-				5	



Medical officer in charge
UPHC Champaakkara

[Handwritten signature]



Medical officer in charge
UPHC Champakkara