



**BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAMME  
UNDER  
NATIONAL HEALTH MISSION**



**SERVICE PROVIDER**  
Tender No WO-37/2021-2022/69R

No : 1004101

**CYRIX**<sup>®</sup>  
HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility..... <u>THOH</u>	Call Registration Date : .....
Address..... <u>Narakkal</u>	Caller ID : .....
..... <u>Ernakulam</u>	Date of Visit : .....
Ph : .....	Asset No : .....
..... <u>9947240764</u>	EQPT Name : .....
	Manufacture <u>Meditech</u> .. Model : .....
	S. No. <u>NAC.1383201</u> .. Dept. <u>OP</u>

**Service Classification :** Breakdown call  PMS  Calibration  Cust. Training

**Problem Identified :** .....

**Action Taken :** .....

Completed  Date : 17/8/24..... Time : 2:00pm..... Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Sebin Siby</u>	<u>17/8/24</u>	<u>1:00pm</u>	<u>2:00pm</u>

**Customer Remark** Completed  Pending

**Service Engineer Name :** Sebin Siby  
**Signature:** [Signature]  
**Date:** 17/8/24  
**Contact Number:** 9746751738




**Customer Name :** [Signature]  
**Signature:** [Signature]  
**Date:** 17/8/24  
**Contact Number:** 9400191617  
**Designation :** Nanny Officer Gr?  
**Hospital Seal**

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

Sl. No	Particulars	Details
1	Name of District	ERNAKULAM
2	Name of Hospital	THQH NARAKKAL ERNAKULAM
3	Name of Equipment with Make, Model and Serial Number	Equipment: Nebuliser Make: Meditech Model: Handyneb SN :NGC1383201
4	Equipment ID & Barcode	129677 & 0722569
5	Date of purchase / Year of manufacture /Installation Date	20-12-2012
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	16-08-2024 (Toll free)
9	Action taken	Checked and found compressor motor and piston defective. Need to replace these spares for further checking and working condition of the equipment..
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	Rs. 2800/-
14	#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment as BER	Checked and found compressor motor and piston defective. Machine installed on 20-12-2012 and covered up to 11+ years and very old machine. So recommending the equipment for condemnation
17	Name & Signature of CYRIX Authority	Sebin 

\*Not mandatory **#Based on the period of life and value as per the BER guidelines**

\* **Attach Photograph**

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Motor and piston defective reported. change in installation date observed. As per the asset details the installation date of this nebulizer is 01/10/20. But attached stock details in the BER Report it is 20.12.2012. Submitted after 30 days. Quot repair cost / spares enquiry details not attached. Stock details not signed by hospital authority. Not recommended.

**CONSULTANT**  
**Bio-Medical Engineer** *Devathy*  
 National Health Mission (NHM) 17/10/24  
 Signature of JC BM (NHM)  
 Arroykeralam


Date 


*Signature*  
 16/12/24  
 Signature of  
 Superintendent / Medical Officer (i/c)  


**സൂപ്പർവൈസിംഗ് ഓഫീസർ (i/c)**  
**ആര്യകരതം**

Date	Receipt	Issued	Balance	Signature	Remarks
	BF from old register page no 5				
9-10-2009	Received from store Rs 1487/-	1	1		KMSCI Page 4
20-12-12	Received from store Rs 2800/-	1 (unissued)	2		Page 21
01-20	Received from store	1	3		
20-10-22	condemned as per the order no 495/2022/Sept/20 Dt - 20/10/22	1	2		
21-21	Received from store Rs 1200/- (Block Panchayat).	1	3		MSR 9



  
 16/12/22  
**മുഖ്യ മത്സ്യവകുപ്പിൻ്റെ സെക്രട്ടറി**



16/8/24  
 സുഖമുഖ  
 കേരള ആരോഗ്യ വകുപ്പിന് ഞാനറിയിക്കുന്നു