



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

## CYRIX

HEALTHCARE PVT LTD

**No. :**  
205159

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility ..... C.H.C.....</p> <p>Address ..... C.H.C. Sooranadu North Kollam.....</p> <p>Ph : ..... 7510982509.....</p>	<p>Call Registration Date : ..... 29/04/2024.....</p> <p>Caller ID : ..... 108563.....</p> <p>Date of Visit : ..... 30/04/2024.....</p> <p>Asset No. : ..... 0230011.....</p> <p>EQPT Name : ..... UPS.....</p> <p>Manufacture ..... APC..... Model : ..... RC-LK-1A.....</p> <p>S. No. ..... NA..... Dept. ..... Laboratory.....</p>
---	---

**Service Classification :** Breakdown Call  PMS  Calibration  Cust.Training

**Problem Identified :** ..... ups not gettence on.....

**Action Taken :** ..... checked the UPS and found the boards and 3 batteries defective. Engineered spares with OEM.....

Completed  Date : ..... 30/4/24..... Time : ..... 2:00pm..... Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time

**Customer Remark**  Completed  Pending

*for Mr. Bushy*  
*[Signature]*  
**MEDICAL OFFICER**  
In-Charge  
C.H.C. Sooranadu

**Service Engineer Name :** ..... Diswarya Rajk N.....  
**Signature :** ..... [Signature].....  
**Date :** ..... 30/4/24.....  
**Contact Number :** ..... 9645384968.....

**Customer Name :** ..... [Signature] .....  
**Signature :** ..... [Signature] .....  
**Date :** ..... 30/4/24.....  
**Contact Number :** ..... 7510982509.....  
**Designation :** ..... Lab technician.....  
**Hospital Name :** ..... [Signature] .....




**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

<b>Sl. No</b>	<b>Particulars</b>	<b>Details</b>
1	Name of District	Kollam
2	Name of Hospital	CHC SOORANAD NORTH
3	Name of Equipment with Make, Model and Serial Number	UPS Make : APC Model : SRC-1K-IN SI no: NA
4	Equipment ID & Barcode	Barcode : 0230011 / Ticket id : 108563
5	Date of purchase / Year of manufacture /Installation Date	30-01-2020
6	Warranty details (Yes/No)	No Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	29-04-2024
9	Action taken	Checked the machine found that the Driver-control PCB and 3 batteries were defective, Enquired spares with OEM
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not Recommended for Repair
12	Cost of spares (specify parts and cost)	APC 1KVA DRIVER PCB = 4956 APC CONTROL PCB = 2655 Exide 12V/7ah smf battery 3 no's= 2765 Total = 10367 /-


13	Asset Value	Rs. 4500/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	230.37%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix Service report & OEM Quotation attached
16	Reasons for recommending the equipment as BER	Checked the ups found that the Driver&control PCB and 3 batteries were defective. UPS is installed on 30-01-2020 and aged upto 3 year 3 month.Repairing cost of equipment is 230.37%. as per tender clause 5.3.14.1 both criteria met So we are recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	AISWARYA RAJ 

\*Not mandatory #Based on the period of life and value as per the BER guidelines

\* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Verified quotations and found that repairing cost is very high. So recommended for Beyond economic repair


  
30/05/24  
Signature of JC BM (NHM)




Date

Seal



  
Signature of  
Superintendent /  
Medical Officer (i/c)

**MEDICAL OFFICER**  
In charge  
C.H.C Sooranadu



*Handwritten signature*  
Dr. P. S. Srinivasan

MEDICAL OFFICER  
In Charge  
C.H.C Sooranadu

# SEMI AUTOMATED CLINICAL CHEMISTRY ANALYSER

Sl. No.	Date	Model	Unique ID	Calibration Status	Remarks
1.	Due date 29/1/2020  Invoice date 30/1/2020  Date of Time of Supply 30/1/2020.	Agappe.  Mispa Viva  with ups			Serial no: 211930341 92  Rs: 59,78



*Dr. Prakash*  
*Dr. Ph...*

MEDICAL OFFICER  
 In-Charge  
 C.H.C Sooranadu

**AION BACKUPS**

YP/VIII/320 AYLARA, KERALA KOLLAM

Date: 28/05/2024

Contact: 8921950217

Email: [aionbackups12@gmail.com](mailto:aionbackups12@gmail.com)

GSTIN: 32CGUPM9450L1ZN

**1KVA APC ONLINE UPS SERVICE PROPOSAL**

TO,

CYRIX HEALTHCARE PVT. LTD, CHC SOORANAD

Dear Sir/Madam,

We are pleased to submit here with our quote for the same as per below.

SL	DESCRIPTION	QTY	UNIT RATE IN INR	TOTAL	GST	TOTAL PRICE IN INR
1	APC 1KVA DRIVER PCB	1	4200.00	4200.00	18	4956.00
2	APC CONTROL PCB	1	2250.00	2250.00	18	2655.00
3	EXIDE 12V/7AH SMF BATTERY	3	720.00	2160.00	28	2765.00
<b>PAYABLE AMOUNT</b>						<b>10367.00</b>

**Commercial terms and conditions**

1. Payment 80% advance along with PO balance before dispatch, 20% After installation.
2. Delivery period; 2weeks.
3. Tax :GST@18% for spares,28%@Battery
4. Installation; Our scope.

Please arrange to release the PO at our above address

Bank name :HDFC

Account no:50200082743176

IFSC Code:HDFC0004083

We hope you will find the above offer in line with our discussions and await your valuable order please free to call or mail us on above given number if any query.

Thank you, Sincerely,

BHAGATH MOHAN

