

694



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



No : 125338

SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX[®]

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility: <i>GH Neryattinkal</i>	Call Registration Date: <i>22-5-2024</i>
Address: <i>Petta Junction</i>	Caller ID: <i>112492</i>
Ph: <i>9961515246</i>	Date of Visit: <i>26-11-2024</i>
	Asset No: <i>0112969</i>
	EQPT Name: <i>UPS</i>
	Manufacture: <i>Emerson</i> Model: <i>Liebert GXT</i>
	S. No: <i>NA</i> Dept: <i>Lab</i>

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : *Not working*

Action Taken : *checked and found Driver PCB, changes PCB, control PCB and batteries defective need to replace. these spares for further checkup and working condition of the equipment. required spares from vendor.*

Completed Date : Time : Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			



Cyrix Engineer	Date	Start Time	End Time
<i>Ajith George</i>	<i>26-11-24</i>		

Customer Remark: Completed Pending

Service Engineer Name : <i>Ajith George</i>	Customer Name : <i>Superintendent</i>
Signature: <i>[Signature]</i>	Signature: <i>[Signature]</i>
Date: <i>26-11-2024</i>	Date: <i>26-11-2024</i>
Contact Number: <i>759384711</i>	Contact: <i>984609055</i>

NEYYATTINKAL HOSPITAL
SUPERINTENDENT
GENERAL HOSPITAL
NEYYATTINKAL KERALA-69521

PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: GH NEYYATINKARA	Name of District: THIRUVANANTHAPURAM
Name of Equipment: UPS	Equipment ID & Barcode: 112492/0112969
Make: EMERSON	Model: Liebert GXT MT+
Serial Number: NA	Warranty details: NA
Date of purchase/ Year of manufacture/ Installation Date: 25/09/2013	Present status of the equipment: Fully Damaged
Date of breakdown: 22/05/2024 (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: Checked and found Driver PCB, Charger PCB, Control PCB and Batteries defective. Need to replace these spares for further checking and working condition of the equipment. Enquired spares from vendor.	
Recommendations for repair (required service details): Not recommending for repair.	
Cost of spares (specify parts and cost): 1. 3KVA Driver PCB = 7670/- 2. Charger PCB = 5310/- 3. Control PCB = 29568/- 4. 12v 42Ah Batteries (6 Nos) = 29568/- Total = 45498/-	
# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 97.27%	Asset Value: 46774/-
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix Service report and vendor quotation attached	
Reasons for recommending the equipment as BER: Checked and found Driver PCB, Charger PCB, Control PCB and Batteries complaint. The unit was installed on 25/09/2013 and covered up to 11+ years. The repair cost is 97.27%. As per the tender clause 5.3.14.2 both criteria met. So recommended the unit for condemnation.	
<i>Karyep.p.v</i> Name & Signature of CYRIX Authority with date	
Remarks of Junior Consultant (Biomedical) NHM: <i>Repair cost is high as per the estimate received. So as per 5-3.14.2 this can be RBER.</i>	
Recommended for BER (Yes/ No):	 MANJESHA MOHAN M.R. Junior Consultant (Biomedical) National Health Mission Thiruvananthapuram Signature of JC BM (NHM)
Date:	
Date:	 SUPERINTENDING SUPERVISOR GENERAL HOSPITAL NEYYATTINKARA-69512


Seal SUPERINTENDING SUPERVISOR Medical Officer (i/c)
 GENERAL HOSPITAL
 NEYYATTINKARA-69512

STOCK REGISTER

59

From Page No. FULLY AUTOMATED RANDOM To page No. ACCESS CLINICAL CHEMISTRY ANALYSER GM 360.

Date	Previous Stock	Purchase	Total	Sales	Balance	Remarks
25/9/2013		Received from lab. (extra em 360). lab. 1	1		1	
25/9/13						
20/1/2019		Received from lab.	2		2	
	(a)	Fully Automated Biochemistry Analyser (Agappe). Po 390	1		9,82,496/-	
	(b)	Desktop monitor Misra clinix	1 NO.			Serial NO. WT8A001325
	(c)	Desktop Computer Mindray higher end equipments (without monitor).	1 NO.			
	(d)	Computer Table (FAA)	1 NO.			
	(e)	ups Ape ups 600VA. BX 600E IN Schneider electric	1 NO.			
	(f)	Printer hp 1108 (hp 1108)	1 NO.			
	(g)	Compatible online ups 2kVA (1hr) back up	1 NO.			
	(h)	Battery for 2/3 kVA (2V/42A) up	1 NO.			
	(i)	Battery for 2/3 kVA (2V/42A) up	1 NO.			



SUPERINTENDENT
GENERAL HOSPITAL
MYSURU
(BATTERY)

AION BACKUPS

YP/VIII/320.AYLARA, KERALA KOLLAM

Date: 26/11/2024

Contact: 8921950217

GSTIN: 32CGUPM9450L1ZN

EMERSON 3KVA ONLINE UPS SERVICE PROPOSAL

TO,

CYRIX HEALTHCARE PVT. LTD

(Barcode: 0112969 – Ticket ID: 112492)

Dear Sir/Madam,

We are pleased to submit here with our quote for the same as per below.

SL	DESCRIPTION	QTY	UNIT RATE IN INR	TOTAL	GST%	TOTAL PRICE IN INR
1	3KVA DRIVER PCB	1	6500.00	6500.00	18	7670.00
2	CHARGER PCB	1	4500.00	4500.00	18	5310.00
3	CONTROL PCB	1	2500.00	2500.00	18	2950.00
4	AMARON 12V 42Ah BATTERY	6	3850.00	23100.00	28	29568.00
PAYABLE AMOUNT						45498.00

Commercial terms and conditions

1. Payment 80% advance along with PO balance before dispatch, 20% after installation.
2. Delivery period; 2weeks.
3. Tax :GST@18% for spares,28%@Battery
4. Installation; Our scope.

Please arrange to release the PO at our above address

Bank name :HDFC

Account no:50200082743176

IFSC Code:HDFC0004083

We hope you will find the above offer in line with our discussions and await your valuable order please free to call or mail us on above given number if any query.

Thank you, Sincerely,
BHAGATH MOHAN



SUPERINTENDENT
GENERAL HOSPITAL
NEYYATTINKARA-695124



M

REGARDING SERVICE/SPARE QUOTATION FOR UPS-GH NEYYATTINKARA-(112492-0112969)

2 messages

Kasyep P V <kasyepcyrix@gmail.com>
To: aionbackups12@gmail.com
Cc: Afsal Yunuze <zm1.klbemp@cyrix.in>

Mon, 25 Nov 2024 at 1:49 pm

Dear sir,
Please provide the spare/service quotation for UPS which we have sent for service;
Make : Emersion
Barcode : 0112969
Ticket : 112492

Thanks & Regards,

KASYEP PV
District Incharge ,Trivandrum
CYRIX HEALTHCARE Pvt.Ltd
KLBEMP
Ph : +91 7593847134
kasyepcyrix@gmail.com

Bhagath Mohan <aionbackups12@gmail.com>
To: Kasyep P V <kasyepcyrix@gmail.com>

Tue, 26 Nov 2024 at 8:26 am

Thanks & Regards

Bhagath Mohan
PH:-8921950217

[Quoted text hidden]

