



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 1021800

CYRIX[®]

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

<p>Health Facility..... <u>THQ#</u>.....</p> <p>Address..... <u>Cherthala</u>.....</p> <p>..... <u>Alappuzha</u>.....</p> <p>Ph:..... <u>9447309332</u>.....</p>	<p>Call Registration Date : <u>6/9/2024</u>.....</p> <p>Caller ID : <u>123778</u>.....</p> <p>Date of Visit : <u>7/9/2024</u>.....</p> <p>Asset No : <u>0420331</u>.....</p> <p>EQPT Name <u>Pharmacy Refrigerator</u></p> <p>Manufacture <u>cyd scienti</u> Model : <u>FRL-V-300</u></p> <p>S. No. <u>017-B/2013</u> Dept..... <u>LAB</u>.....</p>
--	---

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : Refrigerator cooling issue.....
..... High temperature (19°C).....

Action Taken : checked the machine found that the compressor, fan motor, cooling coil set condenser and thermostat are defective. The internal pipe is broken and gas is leaking. Need to replace these spars for further checking and working condition of the equipment.....

Completed Date : 7/9/2024 Time : 11:00 AM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Nasut</u>	<u>7/9/24</u>	<u>10:00 AM</u>	<u>11:00 AM</u>

Customer Remark Completed Pending

Temperature issue.

Service Engineer Name : Nasut

Signature: Nasut

Date: 7/9/24

Contact Number: 9526744524

Customer Name : Kotha V


Signature: [Signature]

Date: 7-9-24

Contact Number: 9447309332

Designation : SLT

Hospital Seal

PROFORMA RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER) BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)	
Name of Hospital: THQH CHERTHALA	Name of District: ALAPPUZHA
Name of Equipment: PHARMACY REFRIGERATOR	Equipment ID & Barcode: 133778 & 0420331
Make: CRYO SCIENTIFIC SYSTEM PVT LTD	Model: FRC - V - 300
Serial Number: 017-B/2013-14	Warranty details: NO WARRANTY
Date of purchase/ Year of manufacture/ Installation Date: 16/07/2013	Present status of the equipment: FULLY DAMAGED
Date of breakdown: 06/09/2024 (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: NO AMC/CAMC
Action taken: CHECKED THE MACHINE FOUND THAT THE COMPRESSOR, FAN MOTOR, COOLING COIL SET CONDENSOR AND THERMOSTAT ARE DEFECTIVE. THE INTERNAL PIPE IS BROKEN AND GAS IS LEAKING.	
Recommendations for repair (required service details) : NOT RECOMMENDED FOR REPAIR	
Cost of spares (specify parts and cost): NA	
# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA	Asset Value: 124948/-
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not):- Attached Cyrix Service Report	
Reasons for recommending the equipment as BER:- The machine was checked and found to have the following defects: The compressor, cooling coil set, fan motor, condenser and thermostat are defective. Additionally, there is a gas leak and the copper pipe is leaking internally. The equipment was installed on 16/07/2013 and aged up to 11+ years. Wear and tear has rendered the machine in an irreparable condition. So we recommend the equipment for condemnation.	
Name & Signature of CYRIX Authority with date :- NASEEF K N 	
Remarks of Junior Consultant (Biomedical) NHM:-	
Recommended for BER (Yes/ No):	Signature of JC BM (NHM)
Date:	
Date:	Signature of Superintendent / Medical Officer

Handwritten signature
SC-7



**SUPERINTENDENT
MALUK HEAD QUARTERS HOSPITAL
CHERTHALA**

TALUK HEAD QUARTERS HOSPITAL

CHERTHALA

008

NAME : REFRIGERATOR


Date	Received	IN NO	Rate	Receipts	Issues	Balance
3/1/2000	Refrigerator model no 5913600739			1/10		1/10
	Calibrated on 16/03/2010					
	Calibrated by STC					
	Using in Lab				1/10	
2/2/14	Transferred to theatre					
1/7/13	Various Refrigerator model no 541056 DL2 B001407 ESD code THB/Lab/008			1/10		1/10
	(This Refrigerator from Lab. Transferred with Fully Biochemistry analysis IN. NO - 13080053 dt 16/5/13 dt kmscl)					
	KIRLOSKAR BAR CODE NO - (A14) 68 (254) 04 (21) 0093					
2/7/13	Using in Lab.					
16/7/13	Cogn Scientific System Pvt Limited Model no FRC-V-300. Refrigerator with Temperature display			1/10		5/10
Store Page No 10	Serial no - 012/18/013-14 Esd code - THB/Lab/009	IN NO 050/6/13				
	KIRLOSKAR - (A14) 68 (254) 04 (21) 0000					
17/7/13	Using in Lab					
13/7/13	Received from (kmscl suppl) LG Store (maxscl Agency K-4)			1/10		5/10
Store Page No 10	IN. NO B2B/0135/17-18/dt B2B/13/009					
14/7/13	Issued to Lab (Using in Lab)					
3/1/19	Received from STC LG Refrigerator			1/10		2/10

Store Page No 07
Equipment & Furniture



[Handwritten Signature]

SUPERINTENDENT
TALUK HEAD QUARTERS HOSPITAL
CHERTHALA



Handwritten signature and text:
 Chithra.V
 SL7