

699/cen.



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

No : 1000813

# CYRIX<sup>®</sup>

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

## Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility..... <u>PHU</u> .....	Call Registration Date : ..... <u>10/10/24</u> .....
Address..... <u>Amachal</u> .....	Caller ID : ..... <u>139955</u> .....
..... <u>T. vandrum</u> .....	Date of Visit : ..... <u>10/10/24</u> .....
Ph : ..... <u>9446475584</u> .....	Asset No : ..... <u>0140655</u> .....
	EQPT Name : ..... <u>Nebulizer</u> .....
	Manufacture <u>W. L. Co.</u> Model : <u>Alpha</u>
	S. No. <u>NA</u> Dept. <u>OP</u>

Service Classification : Breakdown call  PMS  Calibration  Cust. Training

Problem Identified : ..... not working.....

Action Taken : ..... checked and found that the unit is not providing enough times. Identified that the motor wear and piston defective. Need to replace these spares for further checking and working condition of the equipment......

Completed  Date : ..... 10/10/24..... Time : ..... 12pm..... Spare Required

Spare Replaced  Requested


Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Ranjan</u>	<u>10/10/24</u>	<u>11:30am</u>	<u>12pm</u>

Customer Remark ..... Completed ..... Pending

Service Engineer Name : Ranjan P  
Signature: [Signature]  
Date: 10/10/24  
Contact Number: 9074570652

Customer Name : SANITHA C. Nair  
Signature: [Signature]  
Date: 10/10/2024  
Contact Number: 9446475584  
Designation : Nursing officer  
Hospital Seal Vandrum




Vertical stamp: Family Health Centre (FHC), Amachal, Poonithura, Kochi, 682038. Contact: 9847299500

PROFORMA

RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)  
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: PHC AMACHAL	Name of District: Thiruvananthapuram
Name of Equipment: NEBULISER	Equipment ID & Barcode: 139955 & 0140455
Make: LIFELINE	Model: NA
Serial Number : NA	Warranty details : No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 10/04/2018	Present status of the equipment: Fully Damaged
Date of breakdown: 10/10/2024 (Toll free) (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: Checked and found that the unit is not providing enough fumes. Identified that the motor weak and piston defective. Need to replace these spares for further checking and working condition of the equipment.	
Recommendations for repair (required service details) : Not recommending for repair.	
Cost of spares (specify parts and cost) : NA	
# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA	Asset Value: 1353 /-
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix service report attached.	
Reasons for recommending the equipment as BER: Checked and found that the motor weak and piston defective. The unit was installed on 10/04/2018 and covered up to 6+ years. Quotation not submitted since the spares are not available in the market, so recommending the unit for condemnation.	
<p style="text-align: right;">Kashep pr <i>[Signature]</i> Name &amp; Signature of CYRIX Authority with date</p>	
Remarks of Junior Consultant (Biomedical) NHM: Motor weak and piston defective. ∴ This can be recommended beyond economic repair. Date of submission - 27-12-2024	
Recommended for BER (Yes/ No):	<p style="text-align: right;"><i>[Signature]</i> MANEESHA MOHAN M.R. Junior Consultant (Biomedical) Signal Health Mission Thiruvananthapuram BM (NHM)</p>
Date: 31-12-2024	
Date:	<p style="text-align: right;"><i>[Signature]</i> Signature of Superintendent / Medical Officer (i/c) Medical Officer Family Health Centre (FHC) Kattakada (New), Amachal. Thiruvananthapuram-695572</p>



Invoice no: 17032830022651 / Anril pharmacy, HLL life care h/d, Tum  
 Dated 23.03.2018 / Palliative Care project 2017-18 2123/unt  
 23.3.18 |

23.3.18 To p-nurse | NIL *Dr. Anand R*  
 Invoice no: 302 / Karys medicine depot, Kollam Dated 2/4/2018 / Trish  
 Middle CAC for palliative care 2017-18  
 (ANB, 653/2016, April 2020, 1562.65/unt)  
 8.4.2018 |

20.5.18 To p-c nurse | NIL *Dr. PSR* PCSSR/132  
 Invoice no: 205 / Karys medicine depot, Aluva, Ernakulam  
 Dated 10/04/2018 / SWAAS COPD free supply  
 Vcep 12/2017 NOV. 2022 1022.50/unt  
 10.4.18 |

10.4.18 To Staff nurse | NIL *Dr. PSR* 3.5.2/Pre II  
 Invoice no: 919 / Kerala Surgical Equipment Co / Through  
 Karsel, Tum / free supply, unit rate: 2162, lifeline medical device  
 Dated 30.04.2018  
 30.4.18 |  
 30.4.18 To Staff nurse - | NIL *Dr. PSR* PSR/Pre II

Bill no: C 1556/2019-20 / In-house Drug Bank, SAT, Trivandrum dated  
 06.11.19 / PC / Perfecta (BR-CN116) (JHR: 1765.60/unt)  
 06.11.19 |  
 16.12.19 To Palliative | NIL *Dr. Anand*

Bill no: C-3394/2021-22 / In house drug Bank, SAT, met Trivandrum dated 22/05/21 /  
 Hmc / Lupinis Nebuliser (Pluton Compressor) unit rate: 1355.00  
 22.05.21 2 *Dr. Anand*

Bill no C: 3396/2021-22 / JHDC SAT, Gov met, Trpm dated 22/05/21  
 for oec (Lupinis Nebuliser) unit rate: 1355.00  
 22/05/21 2 *Dr. Anand*



*Dr. Anand R*  
 Family Health Centre (FHC)  
 Kattakudi, Thrissur  
 K. V. Anand  
 9845572

**CYRIX** BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)

TOLL FREE NO: 1800 - 425-7669

BAR CODE - (8004-890615225)  
0140455



# Life-Line® PISTON COMPRESSOR NEBULIZER

Model No. : Alpha

Input : 230V, 50Hz

Current : 0.8A

Mode of operation: 30 Min. ON - 30 Min. OFF

Manufactured by:

LifeLine Medical Devices, Gurgaon. INDIA



CE



**Life-Line®**

PISTON COMPRESSOR NEBULIZER

Model No. Alpha

Input: 230V, 50Hz

Current: 0.8A

Mode of operation: 30 Min. ON - 30 Min. OFF

LifeLine Medical Devices, Gurgaon, INDIA

CE

CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)

TOLL FREE NO: 1800 - 425-7669

BAR CODE - (8004-890615225)

0140455

*Dr. Pravin Kumar*  
Family Health Centre (FHC)  
Kankaroda (New), Amalgarh  
Gurgaon, Haryana

