

**PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

| | |
|--|--|
| Name of Hospital: DH THODUPUZHA | Name of District: IDUKKI |
| Name of Equipment: Elisa Reader | Equipment ID & Barcode: id:144886 barcode: 0610800 5610800 |
| Make: Biorad | Model: iMark |
| Serial Number: Na | Warranty details: No Warranty |
| Date of purchase/ Year of manufacture/ Installation Date: 4/11/2011 | Present status of the equipment: Fully damaged |
| Date of breakdown: 5/11/2024 (toll-free) (Date of registration of complaint through email/ Toll free) | *AMC/ CAMC Period agreed at the time of purchase: No CAMC/AMC |

Action taken: Check and find that the power supply, lamp, 655nm filter, and 450nm filter are defective. Need to be replace this spares for further checking and working condition of the equipment. Enquired spare with OEM


Recommendations for repair (required service details) Not recommending for repair

Cost of spares (specify parts and cost): Na

| | |
|--|----------------------|
| # Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: Na | Asset Value:128700/- |
|--|----------------------|

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)
The Cyrix service report and OEM EOL Letter attached

Reason: for recommending the equipment as BER: Check and find that the power supply, lamp, 655nm filter, and 450nm filter are defective. The equipment was installed on 4/11/2011 and covered up to 13+ years. spares are not available now. OEM declared the end of life for this machine. So, recommend RBER as per tender clause 5.3.14.2


Jithu Abraham

 4/11/2024
 Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM:
 1. Year of working agree with RBER criteria.
 2. OEM declared the end of life for this machine and spare not available. So, recommend RBER.


| | |
|--|---|
| Recommended for BER (Yes/No): <u>Yes</u> | National Health Mission Arogyakeralam Idukki-685509 Signature of JC BM (NHM) |
| Date: 10/12/24 | |

Date: 18-12-2024
 Medical Officer (i/c)

Seal



Signature of Superintendent /
 Superintendent
 District Hospital
 Thodupuzha





BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 1018416

CYRIX[®]
HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrilx.com | Email : bemp.kl@cyrilx.in

| | |
|---------------------------------------|---|
| Health Facility: <i>DH Thodupuzha</i> | Call Registration Date: <i>05-11-2024</i> |
| Address: <i>Thodupuzha</i> | Caller ID: <i>144886</i> |
| <i>Idukki</i> | Date of Visit: <i>06-11-2024</i> |
| Ph: <i>09497279448</i> | Asset No: <i>0610.800</i> |
| | EQPT Name: <i>Elisa Reader</i> |
| | Manufacture: <i>Bi. Rad</i> Model: <i>Imark</i> |
| | S. No: <i>1325.2</i> Dept: <i>Laboratory</i> |

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : *Not working*

Action Taken : *Checked and found lamp 655nm filter and 450nm filter and power supply complaint. Need to replace these spare for checking and working condition of the equipment. handover the equipment to customer.*

Completed Date: *06-11-2024* Time: *2:00 PM* Spare Required

Spare Replaced Requested

| Description | Qty. | Part Number | PR Number |
|--------------------------------|-------|-------------------|-----------------|
| 1. 655 nm filter; 450nm filter | 1 Nos | | |
| 2. lamp | 1 Nos | | |
| 3. power supply | 1 Nos | | |
| Cyrilx Engineer | | Date | Start Time |
| <i>Vishnuprasad TS</i> | | <i>06-11-2024</i> | <i>12:00 PM</i> |
| | | | End Time |
| | | | <i>2:00 PM</i> |

Customer Remark Completed Pending

Service Engineer Name : *Vishnuprasad TS*

Signature: *[Signature]*

Date: *06-11-2024*

Contact Number: *9497279448*

Customer Name : *Shula. J R*

Signature: *[Signature]*

Date: *06/11/2024*

Contact Number: *9497279448*

Designation : *SLT*

Hospital Seal



Superintendent
District Hospital
Thodupuzha



Bio-Rad
Laboratories (India) Private Limited
(A wholly owned subsidiary)

C-1 Ground Floor, Anand Patani, No. 88,
G.M. Chetty Road, T. Nagar, Chennai-600017
Tel: (91) 44 4203494 / 4202494
Fax: (91) 44 4202494
Email: sales.india@bio-rad.com
Tech Support: 1800 103 1224 / 09873177477
www.bio-rad.com



Warehouse
Door No: 116, Devaneri Village Road, Spolavaram Village,
Chennai - 600057, Tel: (91) 044 30011242
CIN No: U32109TN1900PTC078494
GST No: 33AAACB3202A1ZU

Ref #: BR/W/LSG/ATH/022/01362
Date: 17th February 2022

To,
The Director
Cyrix Healthcare Pvt Ltd
30/641B, Petta
Poonithura, Ernakulam
Kerala - 682038

Sub: Authorisation Letter
(Ref: Supply of Spares)

Dear Sir/Madam,

We, **M/s. Bio-Rad Laboratories (India) Pvt. Ltd.**, having Registered Office at Plot No. 1270 Basement, Lal Dora, Village Kapashera, Opp Fun-Food Village, New Delhi-110037, and **Head office** at EMAAR Digital Greens, 9th Floor, Tower A- Sector 61, Golf Course Extension, Gurgaon-122 102, Haryana, India, are the 100% subsidiary of M/s. Bio-Rad Laboratories Inc., 1000 Alfred Noble Drive, Hercules CA 94547 USA.

We hereby confirm that **M/s. Crescent Lab Equipments, Door No. VIII/875, Santosh Lane, Aluva PO, Ernakulam, Kerala- 683 101, India** is our authorized distributor. They are authorized to submit quote, negotiate, accept order, supply material, and accept payment on behalf of us for the Bio-Rad make products under LSG product category.

For any further assistance, you may contact **Mr. Pramod Nair-Account Manager, (LSG)** on 99958 08666, email id: pramod_nair@bio-rad.com, for above enquiry.

Thanks & Best Regards,

Sincerely yours
For Bio-Rad Laboratories India Pvt. Ltd

DocuSigned by:
Rony Banerjee 19-Feb-2022
BCADC16C51D1489...
Rony Banerjee
Divisional Head- LSG India

DocuSigned by:
Amit Gaur 20-Feb-2022
34B51B3866834EF...
Amit Gaur
Accounting Manager- India



jithu abraham <jithuabrahamcyrix@gmail.com>

Fwd: Purchase Order - KL 23 24 103199 -Crescent Lab Equipments-Ernakulam -0110050 -GH THIRUVANANTHAPURAM

1 message

Tue, Dec 3, 2024 at 11:29 AM

Fathima Salam <bempservicekl2@gmail.com>
To: jithu abraham <jithuabrahamcyrix@gmail.com>

Thanks & Regards

Fathima Salam
Sr. Service Coordinator
bempservicekl2@gmail.com
www.cyrixhealthcare.com

Cyrix Healthcare Pvt Ltd
| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India
KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH

----- Forwarded message -----

From: **Fathima Salam** <bempservicekl2@gmail.com>
Date: Thu, Sep 19, 2024 at 3:37 PM
Subject: Fwd: Purchase Order - KL 23 24 103199 -Crescent Lab Equipments-Ernakulam -0110050 -GH THIRUVANANTHAPURAM
To: Jineesh P P <jineeshcyrix@gmail.com>

----- Forwarded message -----

From: **Crescent Lab** <crescentlab07@gmail.com>
Date: Thu, 4 Apr, 2024, 1:22 pm
Subject: Re: Purchase Order - KL 23 24 103199 -Crescent Lab Equipments-Ernakulam -0110050 -GH THIRUVANANTHAPURAM
To: IZRATH <purchasecyrix5@gmail.com>
Cc: Cyrix Purchase Mgr. <purchasemgr@cyrix.in>, Project Manager KLBEMP <pm.bempkl@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, BEMP Service 2 <bempservicekl2@gmail.com>, <anaghacyrix@gmail.com>, Cyrix Accounts Payable <acc.payable@cyrix.in>, Riju Reji <rijurejicyrix@gmail.com>, Logistics Cyrix <logisticscyrix@gmail.com>, Cyrix Logistics <logistics@cyrix.in>, Kerala GRN <keralagr@cyrix.in>, Afsal Yunuze <zm1.klbemp@cyrix.in>, Siju K Mathai-Asst Manager (Accounts) <asst.acmanager3@cyrix.in>, Cyrix Payments <paymentcyrix@gmail.com>, Anjitha T L <purchasecyrix2@gmail.com>, arjun u <arjunucyrix@gmail.com>, Abijith A <abijithacyrix@gmail.com>, <tanimachinnucyrix@gmail.com>


Dear Sir,

Please find the attached letter

With Regards
Sruthy

Crescent Lab Equipments
Door No:19, Opp ESI Dispensary ,
Tele Exchange Road , Aluva -683101
Ernakulam Dist, KERALA

Ph: 9633929586, 9495389586
E-MAIL: info@crescentlab.in; crescentlab07@gmail.com
Web : www.crescentlab.in



12/4/24, 10:15 AM

Gmail - Fwd: Purchase Order - KL 23 24 103199 - Crescent Lab Equipments-Ernakulam -0110050 -GH THIRUVANANTHAPU ..

Pin Code - 682309 -,Kerala, India

DPM Req No: 100784

@ Cyrix Accounts

Payment Term: AGAINST DISPATCH CONFIRMATION

-

-

-

Thanks & Regards

Anjitha T L

Purchase Assistant

PH:- 9947719111

purchasercyrix2@gmail.comwww.cyrixhealthcare.com**Cyrix Healthcare Pvt Ltd**

| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India

KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH cyrix.pdf
116K



Door No 19 | Opp. ESI Dispensary
Mar Gregorios Road | Aluva | Ernakulam 683 101
T: +91 9633929586 | +91 9495389586
E: info@crescentlab.in | crescentlab07@gmail.com
W: www.crescentlab.in

04/04/24

To,

CYRIX HEALTHCARE Pvt. Ltd

30/641B, Petta, Poonithura

Kochi

Dear Sir/Madam,

Here we confirm you that the instrument Biorad imark reader discontinued & the lamp for the instrument is not available.

Thanking You

For CRESCENT LAB EQUIPMENTS



