



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

No : 1019941

# CYRIX<sup>®</sup>

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

<p>Health Facility..... <u>Phc</u>.....</p> <p>Address..... <u>Phc Kodanal</u>..... <u>Gonakulam, Kerala</u>.....</p> <p>Ph:..... <u>9605001262</u>.....</p>	<p>Call Registration Date : ..<u>18.16.2024</u>.....</p> <p>Caller ID : ..<u>117941</u>.....</p> <p>Date of Visit : ..<u>20.16.2024</u>.....</p> <p>Asset No : ..<u>0740209</u>.....</p> <p>EQPT Name : ..<u>Pharmacy Refrigerator</u>.....</p> <p>Manufacture.....<u>Godrej</u>..... Model : ..<u>R.D edge</u>.....</p> <p>S. No.....<u>NA</u>..... Dept.....<u>LAD</u>.....</p>
--	---

Service Classification : Breakdown call  PMS  Calibration  Cust. Training

Problem Identified : ..Not working.....

Action Taken : ..checked the equipment and found inner condenser tube leaked which is not able to replace, compressor, cooling coil are defective and need gas filling. Machine not use in a long period and very old machine......

Completed  Date : ..20.16.2024..... Time : ..12:00pm..... Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Yedhukrishna</u>	<u>20.16.24</u>	<u>11:20am</u>	<u>12:00pm</u>

Customer Remark : ..Completed..... Pending

<p>Service Engineer Name : <u>Yedhukrishna</u></p> <p>Signature: <u>[Signature]</u></p> <p>Date: <u>20.16.2024</u></p> <p>Contact Number: <u>8056147236</u></p>	<p>Customer Name : <u>Pooja Sankaran</u></p> <p>Signature: <u>[Signature]</u></p> <p>Date: <u>20.16.24</u></p> <p>Contact Number: <u>9947062182</u></p> <p>Designation : <u>Nursing officer crI</u></p> <p>Hospital Seal</p>
---	--




Quot

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

PROFORMA


Sl. No	Particulars	Details
1	Name of District	ERNAKULAM
2	Name of Hospital	PHC KODANAD ERNAKULAM
3	Name of Equipment with Make, Model and Serial Number	Equipment : Pharmacy Refrigerator. Make : Godrej Model :RD Edge SN : Na
4	Equipment ID & Barcode	117841 & 0740204
5	Date of purchase / Year of manufacture /Installation Date	Not Available (Customer letter attached)
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	18-06-2024 (Toll free)
9	Action taken	Checked the equipment and found inner condenser tube leaked which is not able to replace, compressor, cooling coil are defective and need gas filling.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	Rs. 26494 /-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached and customer letter attached
16	Reasons for recommending the equipment as BER	Checked the equipment and found inner condenser tube leaked which is not able to replace, compressor, cooling coil are defective and need gas filling. The equipment installed more than 10 years and very old. Machine not use in a long period. Also based on the evaluation of energy efficiency, environmental impact, safety standards and technical obsolescence, it is recommended to condemn. Customer requested to condemn the machine.
17	Name & Signature of CYRIX Authority	YeduKrishna. 

\*Not mandatory **#Based on the period of life and value as per the BER guidelines**

\* **Attach Photograph**

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:  
 Part stock details not available with hospital. Hospital reported as more than 10 year old unit. Compressor and freezer reported damage. Quotation not attached. As per healthcare recommended BER due to the safety standards, technical obsolescence, energy efficiency. BER yearwise criteria met. Submitted after 30 days. Final BER recommendation up to kmSCI.

  
 Signature of JC BM (NHM)  
 Bio-Medical Engineer  
 National Health Mission (NHM)  
 Aranyakeralam Ernakulam

Date \_\_\_\_\_

  
 Seal

  
 Superintendent Medical Officer



അപേക്ഷ

ഈ ഘടനാപരിഷ്കരണ 10 ഘട്ടങ്ങളിൽ  
ചെയ്യുന്നതിനായി  
പുഴുപുറ്റു താഴെ കൊടുത്തിരിക്കുന്ന ചട്ടക്കൂട്ടം  
Stock details (old Stock Register) ലഭ്യമാക്കുന്നതിനായി  
Condemnation ചെയ്യുന്നതിനുമുമ്പ് നടപടികൾ  
സ്വീകരിക്കണമെന്ന് അപേക്ഷിക്കുന്നു.

- Items NO :
- ① Spot light - 1 Barcode - 740226
  - ② Refrigerator - 1 Barcode - 740204
  - ③ Centrifuge - 1 Barcode - 740190
  - ④ Water bath - 1 Barcode - 740209
  - ⑤ Calorie meter - 1 Barcode - 740210.



  
MEDICAL OFFICER INCHARGE  
FHC KODANAD



*Handwritten signature*

**കുമാരികുമാർ ഹാമീഷ് ഇൻചാർജ്**  
**എഫ്.എച്ച്.സി. കോടനാട്**