



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 234268

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility : THOH, Kayamkulam</p> <p>Address : Alappuzha Kerala</p> <p>Ph : 9745338600</p>	<p>Call Registration Date : 11/11/2024</p> <p>Caller ID : 146094</p> <p>Date of Visit : 12/11/2024</p> <p>Asset No. : 0421677</p> <p>EQPT Name : Centrifuge</p> <p>Manufacture : Remi Model : RSC DX</p> <p>S. No. : FALC-10072 Dept. : Lab</p>
---	--

Service Classification : Breakdown Call PMS Calibration Cust. Training

Problem Identified : Not working

Action Taken : We have inspected the machine and found that the electric motor, transformer, speed controller PCB and PCB relay board are defective. Need to replace these spares for further checking and working condition of the equipment.

Completed Date : 12/11/24 Time : 2:00pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
Mridula G	12/11/24	12:30pm	2:00pm

Customer Remark Completed Pending

Service Engineer Name : Mridula G
Signature : *Mridula*
Date : 12/11/2024
Contact Number : 9061199471

Customer Name : RIXIANA A
Signature : *Rixiana*
Date : 12/11/24
Contact Number : 9497514594
Designation : LAB TECHNICIAN
Hospital Seal : [Seal]



@



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 234294

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility : <u>THH, Kayamkulam</u></p> <p>Address : <u>Alappuzha</u> <u>Kerala</u></p> <p>Ph : <u>9745338600</u></p>	<p>Call Registration Date : <u>11/11/2024</u></p> <p>Caller ID : <u>146094</u></p> <p>Date of Visit : <u>12/11/2024</u></p> <p>Asset No. : <u>0421677</u></p> <p>EQPT Name : <u>Centrolige</u></p> <p>Manufacture <u>Remi</u> Model : <u>R&C DX</u></p> <p>S. No. <u>FHLC-10072</u> Dept. <u>Lab</u></p>
---	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not working

Action Taken : We have inspected the machine and found that electric motor, transformer, speed controller PCB and PCB relay board are defective. Replaced spare with the OEM.

Completed Date : 20/11/24 Time : 2:00pm Spare Required


Spare Replaced Requested


Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>M. Abdul G</u>	<u>20/11/24</u>	<u>2:30pm</u>	<u>3:00pm</u>

Customer Remark Completed Pending

<p>Service Engineer Name : <u>M. Abdul G</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>20/11/2024</u></p> <p>Contact Number : <u>9061199471</u></p>	<p>Customer Name : <u>RIXWANA .A</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>20/11/24</u></p> <p>Contact Number : <u>9497517594</u></p> <p>Designation : <u>LAB TECHNICIAN</u></p> <p>Hospital Seal : <u>LAB TECHNICAL</u></p>
--	---



PROFORMA RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER) BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)	
Name of Hospital: THQH KAYAMKULAM	Name of District: ALAPPUZHA
Name of Equipment: CENTRIFUGE MACHINE	Equipment ID & Barcode: 146094 & 0421677
Make: REMI	Model: R8C DX
Serial Number: FHLC-10072	Warranty details: NO WARRANTY
Date of purchase/ Year of manufacture/ Installation Date: 18/07/2014	Present status of the equipment: FULLY DAMAGED
Date of breakdown: 11/11/2024 (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: NO AMC/CAMC
Action taken: WE HAVE INSPECTED THE MACHINE AND FOUND THAT THE ELECTRIC MOTOR, TRANSFORMER, SPEED CONTROLLER PCB AND PCB RELAY BOARD ARE DEFECTIVE. ENQUIRED SPARE WITH THE OEM.	
Recommendations for repair (required service details) NOT RECOMMENDED FOR SERVICE	
Cost of spares (specify parts and cost): ELECTRIC MOTOR = 10920 TRANSFORMER = 3500 SPEED CONTROLLER PCB = 3250 PCB RELAY BOARD = 3000 TOTAL = 20670 (WITHOUT TAX)	
# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 88%	Asset Value: 23430/-
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not) CYRIX SERVICE REPORT AND REFERENCE QUOTATION ATTACHED	
Reasons for recommending the equipment as BER: WE HAVE INSPECTED THE MACHINE AND FOUND THAT THE ELECTRIC MOTOR, TRANSFORMER, SPEED CONTROLLER PCB AND PCB RELAY BOARD ARE DEFECTIVE. THIS MACHINE WAS INSTALLED ON 18/07/2014 AND HAS BEEN IN USE FOR OVER 10 YEARS. THE SPARE PART COSTS ARE EXTREMELY HIGH (88%)AND THE EQUIPMENT MEETS BOTH CRITERIA AS PER THE TENDER CLAUSE 5.3.14.1 SO WE RECOMMENDING THE EQUIPMENT AS BER.	
Name & Signature of CYRIX Authority with date:- MRIDULA G <i>Mridula</i> 20/11/2024	
Remarks of Junior Consultant (Biomedical) NHM: <i>Recommended for BER</i>	
Recommended for BER (Yes/ No): <i>YES</i>	 JYOTHISH THOMAS Consultant (Biomedical Engineering) NHM ALAPPUZHA
Date: <i>21.12.24</i>	
Date:	Signature of Superintendent / Medical Officer <i>[Signature]</i>



2

21

2024

taluk health centre



Centrifuge

No	Date	Receipt	Issue	Balance
1.	20/3/12	Received from Slore REMI ✓ AAC	1	1
2.	18/7/14	Received from Slore REMI - R8CDX ✓	1	2
3.	12/1/16	Received from Slore KEMI ✓ <small>Condensation certificate obtained from OME website from the condensation certificate is attached herewith</small>	1	3
4.	2/12/15	Received from Slore ELTEK Received from Scientific Enterprises P.B No 1951 ph 0487 - 2306954 Invoice No 1994/15-16 dt 2/12/15 through KMSL as per Order No 575/2015/21A/30/10/15 TC → 650 D + TC 652 Equip. Code 21A 73 Rs 23,893/-	1	4
5	1/6/2022	Received from Slore Centrifuge-16 Well Model Purchased through KMSCH Total Amt Rs. 4,3,542.00 Warranty 01/6/22 - 30/6/2024 No installation 01/6/2022 Purchase order no: 50/119/2022-2023/210 UNIVERSAL AGENCIES Ph no 0487 2440 319 8943341569 24-27-03	1	1

Issued to Blood Storage Unit
2/12/15
Recommended for condensation certificate obtained from OME website
certificate is attached herewith



Superintendent
Govt. Taluk Hospital,
Kanyakulam

KI/RE/305/23-24

Dated :31.01.2024

To,
Cyrix Healthcare Pvt Ltd
Kochi

Sub: Spare Quotation FOR R8C DX CENTRIFUGE

With reference to the above, we are giving our estimate as per the request.

Sl.No	Description	Qty	Rate	Amount
1	R8C DX ELECTRIC MOTOR	1 No	10,920.00	10,920.00
2	R8C DX TRANSFORMER BIG	1 No	3,500.00	3,500.00
3	R8C DX SPEED CONTROLLER PCB	1 No	3,250.00	3,250.00
4	R8C DX PCB RELAY BOARD	1 No	3,000.00	3,000.00
			Taxable Value	20670.00
			Gst 18%	3720.00
			G.Total:	24390.00

Rupees Twenty Four Thousand Three Hundred And Ninety Only


Kindly confirm the above estimate .

- 1. Payment -100 % Advance**
- 2. Mention GST No: in the purchase orders.**
- 3. Delivery will be made in 4 to 5 weeks from the date of order acceptance by the company service Department**

Thanking you,

Yours faithfully,

For Kochin Inlab Equipments India Pvt Ltd



Authorized Signatory



Abhilash A <abhilashacyrix@gmail.com>

spare Quotation

1 message

Service@Beston Industries <service@bestonindustries.com>
To: Abhilash A <abhilashacyrix@gmail.com>

Wed, Jan 31, 2024 at 3:45 PM

Dear Sir,

Please find the attached Spare Estimate.

Regards
sarath

 **305.pdf**
464K