



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 133308

CYRIX[®]
HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility..... <u>CHC</u>	Call Registration Date : <u>11/11/2024</u>
Address..... <u>Muhamma</u>	Caller ID : <u>145989</u>
..... <u>Alappuzha</u>	Date of Visit : <u>12/11/2024</u>
Ph : <u>9497634806</u>	Asset No : <u>0430322</u>
	EQPT Name : <u>Pharmacy Refrigerator</u>
	Manufacture <u>Whirlpool</u> Model : <u>14287</u>
	S. No. <u>INC1341342</u> Dept. <u>Pharmacy</u>

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : Refrigerator cooling issue.....

Action Taken : checked the machine and found that the compressor, condensers, OLP relay, Freon Doors, Beading are defective and has is leaking and need to replace these spare for further checking and working condition of the equipment

Completed Date : 12/11/24..... Time : 12:00 pm..... Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer <u>Nasut.KIN</u>	Date <u>12/11/24</u>	Start Time <u>10:00 am</u>	End Time <u>12:00 pm</u>
------------------------------------	-------------------------	-------------------------------	-----------------------------

Customer Remark Completed Pending


Not working.

Service Engineer Name : Nasut.KIN

Signature: Nasut

Date: 12/11/24

Contact Number: 9526744524



Customer Name : Babikal P.S

Signature: Babikal

Date: 12/11/24

Contact Number: 9497634806

Designation : Pharmacist

Hospital Seal



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No WO-37/2021-2022/698

No : 133350

CYRIX[®]

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website :www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility..... <u>C.H.C</u> Address..... <u>Mubamma</u> <u>Alappuzha</u> Ph:..... <u>9497634806</u>	Call Registration Date :..... <u>11/11/2024</u> Caller ID :..... <u>145989</u> Date of Visit :..... <u>12/11/24</u> Asset No :..... <u>0430322</u> EQPT Name : <u>Pharmacy Refrigerator</u> Manufacture..... <u>Whirlpool</u> Model : <u>19257</u> S. No. <u>INL134134</u> Dept. <u>Pharmacy</u>
--	--

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : Refrigerator cooling issue

Action Taken : checked the machine and found that the compressor, condenser, OLP relay, freezer door, reading are defective and gas is leaking. Enquired spare with the vendor.

Completed Date : 28/11/24 Time : 12:00pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrux Engineer <u>Nasut. Kin</u>	Date <u>28/11/24</u>	Start Time <u>10:00am</u>	End Time <u>12:00pm</u>
-------------------------------------	-------------------------	------------------------------	----------------------------

Customer Remark Not working Completed Pending

Service Engineer Name : <u>Nasut. Kin</u> Signature: <u>Nasut. Kin</u> Date: <u>28/11/24</u> Contact Number: <u>9526744529</u>	Customer Name : Signature: Date: Contact Number: Designation : Hospital Seal
---	---



Jayanthi
 മുഖ്യമന്ത്രിയുടെ ഓഫീസർ
 സാമൂഹികാരോഗ്യ ക്ഷേത്രം
 മൂവാറ്റുപുഴ

PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: CHC MUHAMMA	Name of District: ALAPPUZHA
Name of Equipment: PHARMACY REFRIGERATOR	Equipment ID & Barcode: 145989 & 0430322
Make: WHIRLPOOL	Model: 14287
Serial Number: INC13413422	Warranty details: NO WARRANTY
Date of purchase/ Year of manufacture/ Installation Date: 25/02/2013	Present status of the equipment: FULLY DAMAGED
Date of breakdown: 11/11/2024 (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: NO AMC/CAMC
Action taken: CHECKED THE MACHINE AND FOUND THAT THE COMPRESSOR, CONDENSER, OLP RELAY FREEZER DOOR, BEADING ARE DEFECTIVE AND GAS IS LEAKING. ENQUIRED SPARE WITH THE VENDOR.	
Recommendations for repair (required service details) NOT RECOMMENDED FOR REPAIR	
Cost of spares (specify parts and cost): COMPRESSOR = 4000 GAS CHARGING = 3500 CONDENSER = 1000 OLP RELAY = 900 FREEZER, DOOR = 1600 BEADING = 1200 TOTAL = 12200	
# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 61%	Asset Value: 19990/-
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not) CYRIX SERVICE REPORT AND QUOTATION ATTACHED	
Reasons for recommending the equipment as BER: WE HAVE INSPECTED THE MACHINE AND FOUND THAT THE COMPRESSOR, CONDENSER, OLP RELAY, FREEZER DOOR, BEADING ARE DEFECTIVE AND GAS IS LEAKING. THIS MACHINE WAS INSTALLED ON 25/02/2013 AND HAS BEEN IN USE FOR OVER 11 YEARS. THE SPARE PART COSTS ARE EXTREMELY HIGH (61%), AND THE EQUIPMENT MEETS BOTH CRITERIA AS PERTHE TENDER CLAUSE 5.3.14.1. THEREFORE, WE RECOMMEND THE EQUIPMENT AS BER.	
Name & Signature of CYRIX Authority with date NASEEF K N <i>Nax</i> 25/11/2024	
Remarks of Junior Consultant (Biomedical) NHM: <i>Recommended for BER</i>	
Recommended for BER (Yes/ No): <i>Yes</i>	Signature of JC BM (NHM) <i>Jyothish Thomas</i> Biomedical Engineering NHM ALAPPUZHA
Date: <i>31.12.24</i>	
Date:	Signature of Superintendent / Medical Officer <i>Sunitha</i> മെഡിക്കൽ ഓഫീസർ സാമൂഹികാരോഗ്യ കേന്ദ്രം മുഹമ്മ



18/8/83 DMOH Alpy on 14-9-1983

20-9-2006 Amount on 2P

2 G' e Stabilizer & 2-particles

Bill no: 11637 Rs. 11615/-

20-9-06 Received On me in Stabilizer 1

~~20-9-06 Received On me in Stabilizer 1~~

2

25.2.2013

25.2.2013 Received on 2P. Whist/post

Issued to phone

Amulya

Issued to pay.

Jayanthi

മലപ്പുറം ജില്ലാ ഭരണകമ്മീഷൻ
മലപ്പുറം ജില്ലാ ഭരണകമ്മീഷൻ
2006





Jagathi
 സാമൂഹികാശയ കേന്ദ്രം
 മുഹമ്മദ്
Jasir

