



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER

Tender No. WO-37/2021-2022/698

No 1013404

CYRIX[®]
HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

| | |
|---|---|
| Health Facility..... <u>THAH KAVAMKULAM</u> | Call Registration Date : <u>08/11/24</u> |
| Address..... | Caller ID : <u>145638</u> |
| <u>Dist: Alappuzha</u> | Date of Visit : <u>09/11/24</u> |
| Ph: <u>9947661940</u> | Asset No : <u>0421572</u> |
| | EQPT Name : <u>Phototherapy</u> |
| | Manufacture..... <u>Neotech</u> Model: <u>4000 CFL</u> |
| | S. No..... <u>1.11.203E+11</u> Dept. <u>Labour Room</u> |

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : Not getting on

Action Taken : checked the machine and found that the CFL, STARTER, ELECTRONIC CHoke and MAIN BOARD are defective, machine is Rusted and PCB Track is Burned condition need to replace these spare for further checking and working condition of the equipment

Completed Date : 09/11/24 Time : 10:30 AM Spare Required

Spare Replaced Requested

| Description | Qty. | Part Number | PR Number |
|-------------|------|-------------|-----------|
| 1. | | | |
| 2. | | | |
| 3. | | | |

| | | | |
|-------------------------------------|-------------------------|-------------------------------|-----------------------------|
| Cyrix Engineer <u>Adithyan R</u> | Date <u>09/11/24</u> | Start Time <u>10:00 AM</u> | End Time <u>10:30 AM</u> |
|-------------------------------------|-------------------------|-------------------------------|-----------------------------|

Customer Remark Completed Pending

| | |
|---|-----------------------------------|
| Service Engineer Name : <u>Adithyan R</u> | Customer Name : <u>Ani Joseph</u> |
| Signature: <u>Adi</u> | Signature: <u>Ani Joseph</u> |
| Date: <u>09/11/24</u> | Date: <u>09/11/24</u> |
| Contact Number: <u>6282537422</u> | Contact Number: <u>9947661940</u> |
| | Designation: <u>NO</u> |
| | Hospital Seal |





BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No: 1013416

CYRIX[®]

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

| | |
|---|--|
| Health Facility..... <u>THQH KAYANKULAM</u> Address..... <u>DIST: Alappuzha</u> Ph:..... <u>9947661940</u> | Call Registration Date :..... <u>08/11/24</u> Caller ID :..... <u>145638</u> Date of Visit :..... <u>09/11/24</u> Asset No :..... <u>0421572</u> EQPT Name :..... <u>Phototherapy</u> Manufacture..... <u>Neotech</u> Model: <u>L1000CFL</u> S.No..... <u>1.11203E11</u> Dept..... <u>Labours Room</u> |
|---|--|

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified :..... not getting on

Action Taken :..... checked the machine and found that the CFL, STARTER, ELECTRONIC CHoke AND MAIN BOARD are defective, machine PS RUPTD and PCB track is Burned condition and pose possible

Completed Date :..... 16/11/24..... Time :..... 12:00 PM..... Spare Required

Spare Replaced Requested

| Description | Qty. | Part Number | PR Number |
|-------------|------|-------------|-----------|
| 1. | | | |
| 2. | | | |
| 3. | | | |

| Cyrix Engineer | Date | Start Time | End Time |
|--------------------|-----------------|-----------------|-----------------|
| <u>Aadithyan R</u> | <u>16/11/24</u> | <u>11:30 AM</u> | <u>12:00 PM</u> |




Customer Remark Completed Pending

Service Engineer Name : Aadithyan R
 Signature: Aadithyan R
 Date: 16/11/24
 Contact Number: 9282537422

Customer Name : St. Joseph's
 Signature: [Signature]
 Date: 16/11/24
 Contact Number: 9947661940
 Designation : NO
 Hospital Seal



PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

| | |
|---|---|
| Name of Hospital: THQH KAYAMKULAM | Name of District: ALAPPUZHA |
| Name of Equipment: PHOTOTHERAPY | Equipment ID & Barcode: 145638 & 0421572 |
| Make: NEOTECH | Model: 4000 CFL |
| Serial Number: 1.11203E+11 | Warranty details: NO WARRNTY |
| Date of purchase/ Year of manufacture/ Installation Date: 08/03/2012 | Present status of the equipment: FULLY DAMAGED |
| Date of breakdown: 08/11/2024 (Date of registration of complaint through email/ Toll free) | *AMC/ CAMC Period agreed at the time of purchase: NO AMC/CAMC |
| Action taken: CHECKED THE MACHINE AND FOUND THAT THE CFL, STARTER, ELECTRONIC CHOKE AND MAIN BOARD ARE DEFECTIVE. MACHINE IS RUSTD AND PCB TRACK IS BURNED CONDITION. | |
| Recommendations for repair (required service details) : NOT RECOMMENDED FOR REPAIR | |
| Cost of spares (specify parts and cost): NA | |
| # Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA | Asset Value: 25657/- |
| Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): CYRIX SERVICE REPORT ATTACHED | |
| Reasons for recommending the equipment as BER: CHECKED THE MACHINE AND FOUND THAT THE CFL, STARTER, ELECTRONIC CHOKE AND MAIN BOARD ARE DEFECTIVE. THIS MACHINE WAS INSTALLED ON 08/03/2012 AND AGED UP TO 12+ YEARS. THE MACHINE IS RUSTD, PCB TRACK IS BURNED AND IN IRREPARABLE CONDITION. SO WE RECOMMENDING THE EQUIPMENT FOR CONDEMNATION. | |
| Name & Signature of CYRIX Authority with date: ADITHYAN R <i>Adhi</i> 16/11/24 | |
| Remarks of Junior Consultant (Biomedical) NHM: <i>Recommended for BER</i> | |
| Recommended for BER (Yes/ No): <i>YES</i> |  JYOTHISH THOMAS Consultant (Biomedical Engineering) NHM ALAPPUZHA Signature of JC BM (NHM) |
| Date: <i>21/12/24</i> | |
| Date: _____ |  Seal  Signature of Superintendent Medical Officer Superintendent Govt. Taluk Hospital Kayamkulam |

Baby Reinstallation Unit

106

(Photo-therapy unit)

24-3-2011: Rd from stock

8-3-2012: Rd from stock
(Transfer to OT)

26-6-2013 Rd
From page no. 35

Phototherapy unit 5
CO₂, etc.
Warmer 1

Transf to CH ward

9/8/18
13/11/18

Infant radiant warmer
Received from Store (Phototherapy)
(SB 4A 14) → page no store book

Warmer 5
Transf to CH ward - Phototherapy 1

MSEU Infant warmer - 2
new LR " - 1 5
old LR " - 1
MSEU " - 1


phototherapy - 1 Warmer -
MSEU rd - 1 phototherapy -

Side lamp - 1

Baby warmer - 5
Phototherapy py - 2
Side lamp - 3

[Handwritten signature]

Sri
Govt. Taluk Hospital
Kayamkulam



[Handwritten signature]

Bal...

| | |
|---|---|
| 1 | 1 |
| 1 | 1 |
| 3 | 3 |
| 1 | 4 |
| 1 | 6 |
| 1 | 7 |



[Handwritten signature]

Superintendent
Govt. Taluk Hospital
Kayamkulam



