



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 131870

CYRIX®

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility..... <u>PHC</u> Address..... <u>PHC vilakkudi</u> <u>Kallam</u> Ph:..... <u>9544570720</u>	Call Registration Date : <u>19-11-24</u> Caller ID : <u>147504</u> Date of Visit : <u>20-11-24</u> Asset No : <u>0241396</u> EQPT Name : <u>BP Apparatu</u> Manufacture..... <u>olex</u> Model:..... <u>NA</u> S.No..... <u>802491</u> Dept..... <u>OP</u>
---	--

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : Not getting on

Action Taken : Checked and found problem with display
 main board and pump. Need to replace this
 span for further checking and working condition
 of the equipment.

Completed Date : 20-11-24 Time : 11:00AM Spare Required

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Sudhraj. S</u>	<u>20-11-24</u>	<u>10:00AM</u>	<u>11:00AM</u>

Customer Remark Completed Pending



Sampath
Medical Officer-in-Charge
Wakayam Health Centre
Wakayam, Kottayam-P.O.

Service Engineer Name : Sudhraj. S
 Signature: Suy
 Date: 20-11-24
 Contact Number: 9447621489




Jayaram
Nursing Officer
9544570720
Jayaram S

PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: PHC VILAKUDY	Name of District: KOLLAM
Name of Equipment: BP APPARATUS	Equipment ID & Barcode: 147604 & 0241396
Make: OLEX	Model: VM44
Serial Number: W18040802491	Warranty details: NO WARRANTY
Date of purchase/ Year of manufacture/ Installation Date: 12-05-2020	Present status of the equipment: FULLY DAMAGED
Date of breakdown: 19-11-2024 (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: NO AMC/ CAMC
Action taken: Checked the bp apparatus found that not switching on, found problem with main board, Display & pump need to be replaced these spares for further checking & working condition of the equipment	
Recommendations for repair (required service details) Not Recommending for repair	
Cost of spares (specify parts and cost): NA	
# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA	Asset Value: 2200
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not) Cyrix service report	
Reasons for recommending the equipment as BER: Checked the bp apparatus found that not switching on, found problem with main board, Display & pump need to be replace these spare for further checking & working condition of the equipment. Equipment installed on 12-05-2020 aged up to 4 years 6 month. Quotation not submitted since spare parts are not available on market. So we are recommending the equipment for condemnation.	
SUDHIRAJ	 13/12/24 Name & Signature of CYRIX Authority with date
Remarks of Junior Consultant (Biomedical) NHM: Beyond economic repair Fully damaged. Recommended for	
Recommended for BER (Yes/ No):	 19/12/24 Signature of JOC BM (NHM) KOLLAM
Date: 19/12/24	
Date: Medical Officer (JOC)	Signature of Superintendent /

Medical Officer in Charge
 Family Health Centre
 Vilakkudy, Kunnicothu-P.O.




STOCK REGISTER


20

Name digital B.P Apparatus

Date	Received from or Issued to	Inv. No. or Issue	Rate		Receipts	Issues	Balance	Remarks
			Rs.	Ps.				
15/3/23	One digital BP apparatus physically found.							
16/3/23	Received from Koorunja Members Kunnikodu as per no - A-43559 dt-15/3/23 digital BP apparatus 1870 Rs - 1 + Alphone multi cell charger 1					3 3 (charger)		
15/11/23	Created one to substock <u>substock</u> volume II page No - 31 on 15/11/23.							




 Medical Officer-in-Charge
 Family Health Centre
 Vilakkudy, Kunnikodu-P.O.


 Medical Officer-in-Charge
 Family Health Centre
 Vilakkudy, Kunnikodu-P.O.



