



17/12/24
Jithu.

PROFORMA RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER) BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)	
Name of Hospital: DH Idukki	Name of District: IDUKKI
Name of Equipment: Suction Apparatus	Equipment ID & Barcode: id:149646 barcode:0610380
Make: MEDEC	Model: Na
Serial Number: Na	Warranty details: No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 4/11/2020	Present status of the equipment: Fully damaged
Date of breakdown: 29-11-2024 (toll-free) (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: No CAMC/AMC
Action taken: Check and find that the suction motor and suction jar were defective. These spares need to be replaced for further checking and the working condition of the equipment. Enquired spare with OEM	
Recommendations for repair (required service details) Not recommending for repair	
Cost of spares (specify parts and cost): Suction jar:3600/- Suction motor:4650/- Total:8250/-	
# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 142%	Asset Value:5800/-
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not) The Cyrix service report and OEM quotation are attached	
Reasons for recommending the equipment as BER: Check and find that the suction motor and suction jar were defective. The equipment was installed on 4/11/2020 and covered up to 4 + years. The repair cost of the equipment is 142%, and both criteria are met. So recommending the equipment for condemnation as per the tender clause 5.3.14.1.	
Jithu Abraham  Name & Signature of CYRIX Authority with date 17/12/24	
Remarks of Junior Consultant (Biomedical) NHM: * Year of working and % of spare cost agree with RBER criteria * A mismatch found in installation date with respect to asset master and stock register.	
Recommended for BER (Yes/ No):	Signature of JC BM (NHM)
Date: 17/12/2024	
Date: Medical Officer (i/c)	Signature of Superintendent /  SUPERINTENDENT DIST. HOSPITAL IDUKKI





**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 147744

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility <u>DH Welucki</u>	Call Registration Date : <u>29.11.24</u>
Address <u>Cheruthoni</u>	Caller ID : <u>149646</u>
<u>Welucki</u>	Date of Visit : <u>30.11.24</u>
Ph : <u>9445927974</u>	Asset No. : <u>0610380</u>
	EQPT Name : <u>Suction Apparatus</u>
	Manufacture <u>Meeke</u> Model : <u>QA</u>
	S. No. <u>QA</u> Dept. <u>ENT OP</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not working

Action Taken : Check the Machine found that it's Motor and Suction fan are defective. These spare need to be replace for further checking and the working condition of the equipment require spare with DEM.

Completed Date : 30.11.24 Time : 2:00pm Spare Required


Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Neerali</u>	<u>30.11.24</u>	<u>1:30pm</u>	<u>2:00pm</u>

Customer Remark _____ Completed Pending

Service Engineer Name : <u>Neeralikrishna ID</u>	Customer Name : _____
Signature : <u>[Signature]</u>	Signature : _____
Date : <u>30.11.24</u>	Date : <u>30.11.2024</u>
Contact Number : <u>9847798098</u>	Contact Number : <u>9845927974</u>
	Designation : <u>Asst Engg No.1.</u>
	Hospital Seal : _____





9744473043, 9744473243

cloud99biotech@gmail.com

1 KMC/XIII-1333
Room No: 22,
CSI Golden Jubilee Complex
Baker Hill
Shastri Road
Kottayam

CLOUD99 BIOTECH

GSTIN: 32AMWPT5633Q1ZD
State: 32 - Kerala

MSME NO.: UDYAM-KL-07-0030714

Estimate

Estimate For:

KASHYAP

1ST FLOOR 30/641B CYRIX HEALTHCARE PVT LTD
KOCHI DHANUSHKODI ROAD POONITHURA

Contact No.: 7593247134

GSTIN Number: 32AAFCC2499H2ZM

State: 32-Kerala

Estimate No.: C99B-QTN-28

Date: 08/07/2024

Place of Supply: 32-Kerala

#	Item name	HSN/ SAC	Quantity	Unit	Price/ Unit	GST	Amount
1	SUCTION JAR	90189099	1	Nos	₹ 3,214.29	₹ 385.71 (12%)	₹ 3,600.00
2	SUCTION MOTOR (MOTOR FOR HOSPITAL SUCTION)	9018	1	-	₹ 3,940.68	₹ 709.32 (18%)	₹ 4,650.00
Total			2			₹ 1,095.03	₹ 8,250.00



Pay To:

Bank Name : UNION BANK OF INDIA, KOTTAYAM

Bank Account No. : 361201010035646

Bank IFSC code : UBIN0536121

Account holder's name : CLOUD99 BIOTECH



Sub Total	₹ 7,154.97
SGST@6%	₹ 192.86
CGST@6%	₹ 192.86
SGST@9%	₹ 354.66
CGST@9%	₹ 354.66
Total	₹ 8,250.00
You Saved	₹ 0.01

Estimate Amount In Words

Eight Thousand Two Hundred Fifty Rupees only

Terms And Conditions

Thanks for doing business with us!

For: CLOUD99 BIOTECH



Authorized Signatory



SUPERINTENDEN,
DIST. HOSPITAL
IDUKKI





PERFORMED BY
 NEXT PMS DUE 05/11/25
 PMS DONE ON 05/11/25
 PREVENTIVE MAINTENANCE SERVICE
 BEMP SERVICE PROVIDER
 CYRIX HEALTH CARE PVT.LTD

RBER **CYRIX™**
 HEALTH CARE PVT.LTD
 "Commitment to Healthcare"

Ticket No.:
 RBER approved on:
 Condemn on:
 Remark:

(Handwritten signature)
 SUPERINTENDENT,
 DIST. HOSPITAL,
 IDUKKI



S/No	Name of equipment	Barcode No	Company Name	Installation Date	Unit Price	Unit No	Warranty	Remarks
11	Telescope / FNT Operating microscope	0610376	LABORERD. Van Savi, Madhav Savio Villavakkal Chennai	11/01/2015	5,45,000	1	11/01/2015 (3 years)	model - Prima star → NPRD → OR → OR
				Total	5,72,250.00			
12	Suction Apparatus	0610379 0610391	Instakric Kerala Sirogrop Equipments m.p. stand	11/11/2020	9500/-	2		
		0610379 0610380	Meda C					



~~SUPERINTENDENT
DIST. HOSPITAL
IDUKKI~~