





**PROFORMA  
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)  
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

<b>Name of Hospital:</b> GH THRISSUR	<b>Name of District:</b> Thrissur
<b>Name of Equipment:</b> X ray machine	<b>Equipment ID &amp; Barcode:</b> #150802- 0810164
<b>Make:</b> ALLENGERS	<b>Model:</b> OPEL 2.5 SBM
<b>Serial Number:</b> 2K110302091-X/HF	<b>Warranty details:</b> No Warranty
<b>Date of purchase/Year of manufacture/Installation Date:</b> In Ticket master installation date is :NA Actual installation date as per stock book is - 19-04-2011	<b>Present status of the equipment:</b> Fully damaged
<b>Date of breakdown:</b> 05-12-2024 <b>Date of registration of complaint through email/Toll free):</b> Toll free	<b>*AMC/CAMC Period agreed at the time of purchase:</b> No AMC/CAMC
<b>Action taken:</b> checked the machine and found Main board and xray-tube are defective. Need to replace these spares for further checking working condition of equipment.	
<b>Recommendations for repair (required service details):</b> Not Recommending	
<b>Cost of spares (specify parts and cost):</b> NA	
<b>#Percentage value of the cost of spares with respect to Cost of Purchase/Asset Value:</b> NA	<b>Asset Value:</b> 186000/-
<b>Abstract of Service Report provided by the OEM/Authorized Service Provider/CYRIX (Attached or Not):</b> Cyrix report and oem EOL attached	
<b>Reasons for recommending the equipment as BER:</b> Upon inspection, it was found that the mainboard and xray tube of the machine are faulty. As per the stock book (attached for reference), the installation date is recorded as 19-04-2011, making the equipment 12 years and 7 months old. The OEM (Allengers) has confirmed that the X-ray machine Opel 2.5 is an obsolete model, and spare parts are no longer available. As per tender clause 5.3.14.2, we recommend the equipment for condemnation.	
<b>BIJO T JOY</b>	 <b>BIJO T JOY</b> Name & Signature of CYRIX Authority with date
<b>Remarks of Junior Consultant (Biomedical) NHM:</b> <p align="center"><i>Obsolete model.</i></p>	
<b>Recommended for BER (Yes/No):</b> Yes	<b>Jr. Consultant (Biomedical)</b> NHM, Thrissur
<b>Date:</b>	<b>Signature of JCBM (NHM)</b>  13/12/24
<b>Date:</b>	 <b>Signature of Superintendent / Medical Officer (i/c)</b>  GENERAL HOSPITAL (Dy. DHS) THRISSUR - PIN: 680 001

\*Not mandatory Based on the period of file and value as per the BER guidelines \*Attach Photograph



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

No : 140034

## CYRIX<sup>®</sup>

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

<p>Health Facility: <u>U.H. THRISSUR</u></p> <p>Address: <u>THRISSUR</u></p> <p>Ph: <u>9388574985</u></p>	<p>Call Registration Date: <u>5/12/2024</u></p> <p>Caller ID: <u>150802</u></p> <p>Date of Visit: <u>6/12/2024</u></p> <p>Asset No: <u>081016A</u></p> <p>EQPT Name: <u>X-ray machine</u></p> <p>Manufacture: <u>Allen engers</u> Model: <u>OPEL 2.5 SBM</u></p> <p>S. No. <u>2.K.11.0302091</u> Dept. <u>Radiology</u></p>
---	---

Service Classification : Breakdown call  PMS  Calibration  Cust. Training

Problem Identified : X-Ray machine is not working.

Action Taken : checked the machine and found main-board and x-ray tube are defective need to replace these spares for further checking working condition of equipment

Completed  Date : 6/12/2024 Time : 12:00pm Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrux Engineer <u>ABSIL TONY</u>	Date <u>6/12/2024</u>	Start Time <u>11:00am</u>	End Time <u>12:pm</u>
-------------------------------------	--------------------------	------------------------------	--------------------------

Customer Remark Completed  Pending

Service Engineer Name : <u>ABSIL TONY</u> Signature: Date: <u>6/12/2024</u> Contact Number: <u>9745328470</u>	Customer Name : _____ Signature: Date: _____ Designation : <u>SUPERINTENDENT (Dy. DHS)</u> Hospital Seal <u>GENERAL HOSPITAL</u> <u>THRISSUR - PIN: 680 061</u>
--	--



Afsal Y <afsalcyrix@gmail.com>

**OPEL 2.5 PORTABLE X RAY REGARDING.**

7 messages

**Project Manager KLBEMP** <pm.bempkl@cyrix.in> Thu, Jul 28, 2022 at 3:28 PM  
To: RENJITH <ker.service@allengers.net>  
Cc: Afsal Y <afsalcyrix@gmail.com>, "sivancyrix@gmail.com" <sivancyrix@gmail.com>, "bmetrivandrum1@gmail.com" <bmetrivandrum1@gmail.com>

Dear Renjith.

We have an open complaint in GH trivandrum Portable X ray Opel 2.5.

Kindly sent the confirmation stating that Model is Discontinued.

Thanks & Regards

**Sarath K B**  
**Project Manager-Bemp Kerala Project**  
**Cyrix Healthcare Pvt. Ltd.**

30/641 B, Petta Junction | Poonithura  
Kochi | Kerala-682038

**KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH || MAHARASHTRA || RAJASTHAN ||  
KUWAIT || UAE**

Mob.:- 9744555622,9605603452

Email:- pm.bempkl@cyrix.in | Website:- www.cyrixhealthcare.com

"Genius in Healthcare"

**RENJITH** <ker.service@allengers.net> Thu, Jul 28, 2022 at 3:51 PM  
To: Project Manager KLBEMP <pm.bempkl@cyrix.in>  
Cc: Afsal Y <afsalcyrix@gmail.com>, sivancyrix@gmail.com, bmetrivandrum1@gmail.com

Dear Sarath,


**We hereby confirm that Opel 2.5 is discontinued and hence be considered as obsolete model.**

Thanks and regards

**RENJITH RAMESH C** | Executive, Office Operations | Cochin Branch  
Allengers Medical Systems Ltd  
5-B, 5<sup>th</sup> Floor, Hedge House (Pearl Opulence), Edapally-Palarivattom Road, Ernakulam – 682 025  
Tel (484) 4035626 | mobile +91 9946660218 | Toll free 1800 266 8800

[www.allengers.com](http://www.allengers.com) | [ker.service@allengers.net](mailto:ker.service@allengers.net)

**Confidentiality Note:** This e-mail and any attachments are confidential and may be protected by legal privilege. If you are not the intended recipient, be aware that any disclosure, copying, distribution or use of this e-mail or any attachment is prohibited. If you have received this e-mail in error, Please notify us immediately by returning it to the sender and delete this copy from your system. Thank you for your cooperation.

 Please consider the environment before printing this e-mail!

[Quoted text hidden]

---

**Project Manager KLBEMP** <pm.bempkl@cyrix.in>  
To: Afsal Y <afsalcyrix@gmail.com>

Tue, Oct 11, 2022 at 11:55 AM

FYI

Thanks & Regards

**Sarath K B**

**Project Manager-Bemp Kerala Project**

**Cyrix Healthcare Pvt. Ltd.**

30/641 B, Petta Junction | Poonithura

Kochi | Kerala-682038

**KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH || MAHARASHTRA || RAJASTHAN ||**

**KUWAIT || UAE**

Mob.:- 9744555622,9605603452

Email:- pm.bempkl@cyrix.in | Website:- www.cyrixhealthcare.com

*"Genius in Healthcare"*

---

**From:** RENJITH <ker.service@allengers.net>

**Sent:** Thursday, July 28, 2022 3:51 PM

**To:** Project Manager KLBEMP <pm.bempkl@cyrix.in>

**Cc:** 'Afsal Y' <afsalcyrix@gmail.com>; sivancyrix@gmail.com <sivancyrix@gmail.com>;

bmetrivandrum1@gmail.com <bmetrivandrum1@gmail.com>

**Subject:** RE: OPEL 2.5 PORTABLE X RAY REGARDING.

Dear Sarath,

**We hereby confirm that Opel 2.5 is discontinued and hence be considered as obsolete model.**

Thanks and regards

**RENJITH RAMESH C** | Executive, Office Operations | Cochin Branch

Allengers Medical Systems Ltd

5-B, 5<sup>th</sup> Floor, Hedge House (Pearl Opulence), Edapally-Palarivattom Road, Ernakulam – 682 025

Tel (484) 4035626 | mobile +91 9946660218 | Toll free 1800 266 8800

[www.allengers.com](http://www.allengers.com) | [ker.service@allengers.net](mailto:ker.service@allengers.net)

**Confidentiality Note:** This e-mail and any attachments are confidential and may be protected by legal privilege. If you are not the intended recipient, be aware that any disclosure, copying, distribution or use of this e-mail or any attachment is prohibited. If you have received this e-mail in error, Please notify us immediately by returning it to the sender and delete this copy from your system. Thank you for your cooperation.

 Please consider the environment before printing this e-mail

---

**From:** Project Manager KLBEMP [mailto:pm.bempkl@cyrix.in]  
**Sent:** Thursday, July 28, 2022 3:28 PM  
**To:** RENJITH  
**Cc:** Afsal Y; sivancyrix@gmail.com; bmetrivandrum1@gmail.com  
**Subject:** OPEL 2.5 PORTABLE X RAY REGARDING.

Dear Renjith.

[Quoted text hidden]

---

**Afsal Y** <afsalcyrix@gmail.com>  
To: Sajin Simon <sajinncyrix@gmail.com>

Tue, Oct 11, 2022 at 11:58 AM

FYI

Regards,  
AFSAL YUNUZE  
Zonal Incharge (South Region)  
CYRIX HEALTHCARE Pvt Ltd  
Kollam- 691001  
Ph:+919895061437

----- Forwarded message -----

From: **RENJITH** <ker.service@allengers.net>  
Date: Thu, Jul 28, 2022 at 3:51 PM  
Subject: RE: OPEL 2.5 PORTABLE X RAY REGARDING.  
To: Project Manager KLBEMP <pm.bempkl@cyrix.in>

[Quoted text hidden]

[Quoted text hidden]

---

**Afsal Y** <afsalcyrix@gmail.com>  
To: jithu abraham <jithuabrahamcyrix@gmail.com>

Wed, Oct 12, 2022 at 9:13 AM

FYI  
EOL Letter for Allengers Opel 2.5

Regards,  
AFSAL YUNUZE  
Zonal Incharge (South Region)  
CYRIX HEALTHCARE Pvt Ltd  
Kollam- 691001  
Ph:+919895061437

----- Forwarded message -----

From: **RENJITH** <ker.service@allengers.net>  
Date: Thu, Jul 28, 2022 at 3:51 PM  
Subject: RE: OPEL 2.5 PORTABLE X RAY REGARDING.  
To: Project Manager KLBEMP <pm.bempkl@cyrix.in>  
Cc: Afsal Y <afsalcyrix@gmail.com>, <sivancyrix@gmail.com>, <bmetrivandrum1@gmail.com>

[Quoted text hidden]

**Afsal Y** <afsalcyrix@gmail.com>  
To: aneeshcyrix@gmail.com, jithu abraham <jithuabrahamcyrix@gmail.com>

Thu, Oct 27, 2022 at 5:48 PM

FYI,  
REGARDING ALLENGERS OPEL 2.5

Regards,  
AFSAL YUNUZE  
Zonal Incharge (South Region)  
CYRIX HEALTHCARE Pvt Ltd  
Kollam- 691001  
Ph:+919895061437

----- Forwarded message -----

From: **RENJITH** <ker.service@allengers.net>  
Date: Thu, Jul 28, 2022 at 3:51 PM  
Subject: RE: OPEL 2.5 PORTABLE X RAY REGARDING.  
To: Project Manager KLBEMP <pm.bempkl@cyrix.in>  
Cc: Afsal Y <afsalcyrix@gmail.com>, <sivancyrix@gmail.com>, <bmetrivandrum1@gmail.com>

[Quoted text hidden]

**Afsal Y** <afsalcyrix@gmail.com>  
To: Abhilash A <abhilashacyrix@gmail.com>

Tue, Nov 22, 2022 at 9:49 AM

OPEL 2.5 EOL LETTER

Regards,  
AFSAL YUNUZE  
Zonal Incharge (South Region)  
CYRIX HEALTHCARE Pvt Ltd  
Kollam- 691001  
Ph:+919895061437

----- Forwarded message -----

From: **RENJITH** <ker.service@allengers.net>  
Date: Thu, Jul 28, 2022 at 3:51 PM  
Subject: RE: OPEL 2.5 PORTABLE X RAY REGARDING.  
To: Project Manager KLBEMP <pm.bempkl@cyrix.in>  
Cc: Afsal Y <afsalcyrix@gmail.com>, <sivancyrix@gmail.com>, <bmetrivandrum1@gmail.com>




[Quoted text hidden]

# X Ray Machine

5

SNO	ITEM	Qty	DATE OF RECEIPT	ISSUE FROM	REMARKS
1	300mA GE X-ray Machine DA 300, SER NO. FW27-09-28 CONDITIONATION DONE Bg - 4316/16/CHISP	1	9-3-2006		Old machine [faded text]
2	500mA Allengers X-ray Machine 525 FC / HTS. NO. 2K1015073A Rate - 538549/-	1	8-2-2011	Store	
3	Portable 60mA Allengers Rate :- 1,95300/-	1	19-4-2011	"	[faded text] [signature]





*[Handwritten signature]*  
26/12/24