

**PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: PHC KARALAM	Name of District: Thrissur
Name of Equipment: B P Apparatus	Equipment ID & Barcode: #152454- 0841546
Make: Med Vision	Model: NA
Serial Number: NA	Warranty details: No Warranty
Date of purchase/Year of manufacture/ Installation Date: Ticket master installation date Is- NA Actual installation date As per Stock book is : 29-01-2018	Present status of the equipment: Fully damaged
Date of breakdown: 13-12-2024 Date of registration of complaint through email/Toll free): Toll free	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC

Action taken: checked the machine and found display board and cuff are defective. Need to replace these spares for further checking working condition of equipment.


Recommendations for repair (required service details): Not Recommending

Cost of spares (specify parts and cost): NA

#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA	Asset Value: 2200/-
---	---------------------

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix report attached.


Reasons for recommending the equipment as BER:
Upon checking the equipment, it was found that the display board and cuff are defective. The equipment was installed on 29-01-2018 and is now 6 years and 10 months old, with an asset cost of ₹2200. A quotation has not been submitted as the necessary spares are unavailable in the market. Therefore, we recommend the equipment for condemnation.


BIJO T JOY
 Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM:
Spares are not available.

Recommended for BER (Yes/ No): <i>Yes</i>	Signature of JC BM (NHM)
Date:	

Date: *27/12/2024* Seal Signature of Superintendent / Medical Officer (in charge)


 Medical Officer (in charge)
 Family Health Centre, Karalam
 Thrissur - 680 711

*Not mandatory #Based on the period of life and value as per the BER guidelines

Date	No. Date of Voucher or Invoice	From whom received or to whom issued	Receipt	Issued Balance	Remarks
		Received from State			
27/1/18	315				
	28/12/17				
29/1/18		BP - APPARATUS LCD	2	2	Rs. 4209.82
		0841546			





RBER
CYRIX™
"Caring in Healthcare"

Ticket No.: 152454
RBER approved on:
Condemn on: 4/1/25
Remark: MISPLACED (RBER) RESERVE



CYRIX MEDICAL EQUIPMENT MAINTENANCE PROGRAMME GROUP
TOLL FREE NO. 1800 - 425-7668
BAR CODE - (8004-890615225) 0841546

CYRIX MEDICAL EQUIPMENT MAINTENANCE PROGRAMME GROUP
TOLL FREE NO. 1800 - 425-7668
BAR CODE - (8004-890615225) 0841546



MERCURY-FREE
SPHYGMOMANOMETER
LED DISPLAY

CE
0197

MODEL: 1018

