


**PROFORMA  
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)  
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: WOMEN & CHILDREN HOSPITAL KOZHIKODE	Name of District: KOZHIKODE
Name of Equipment: BP APPARATUS	Equipment ID & Barcode: 153877 & 1111214
Make: ELKO	Model: EL-510
Serial Number: NA	Warranty details: No Warranty
Date of purchase/Year of manufacture/ Installation Date: 17/02/2017	Present status of the equipment: Fully damaged
Date of breakdown: 21/12/2024 Date of registration of complaint through email/Toll free): Toll free	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: Checked and found display and main board defective. Need to replace these spares for further checking of working condition of the equipment.	
Recommendations for repair (required service details): Not recommending for repair	
Cost of spares (specify parts and cost): N/A	

#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: N/A	Asset Value: 2200 /-
--	----------------------

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix service report attached


Reasons for recommending the equipment as BER: Checked and found display and main board defective. Need to replace these spares for further checking of working condition of the equipment. Equipment installed on 17/02/2017 (stock book attached for reference) Equipment out lived 7 Years 10 Month, with an asset cost of rupees 2200. Quotation not submitted since spares are not available in the market So we recommend the equipment for condemnation,

  
**JINEESH PP**  
 Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM:

*Checked & found that display and main board case damaged. Need to replace the spares. So recommended for condemnation.*

Recommended for BER (Yes/ No): Yes	Consultant Bio Medical Engineer National Health Mission Kozhikode
Date: 11/1/2025	Signature of JC BM (NHM)

Date: 21/1/2025	 Seal Signature of Superintendent / Medical Officer (i/c) Govt. Women & Children Hospital Kozhikode
-----------------	---

\*Not mandatory

#Based on the period of life and value as per the BER & CYRIX Guidelines

Attach Photograph



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

## CYRIX

HEALTHCARE PVT LTD

**No. :** 200306

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>WOMEN &amp; CHILDREN</u></p> <p>Address <u>HOSPITAL KOZHICODE</u></p> <p>Ph : <u>9074961566</u></p>	<p>Call Registration Date : <u>21/12/2024</u></p> <p>Caller ID : <u>153877</u></p> <p>Date of Visit : <u>23/12/2024</u></p> <p>Asset No. : <u>1111214</u></p> <p>EQPT Name : <u>BPA Paratus</u></p> <p>Manufacture <u>ELKO</u> Model : <u>EL-510</u></p> <p>S. No. <u>NA</u> Dept. <u>CASUALTY</u></p>
---	--

**Service Classification :** Breakdown Call  PMS  Calibration  Cust.Training

**Problem Identified :** → NOT switching on

**Action Taken :** checked and found that display and main board defective so need to replace these spares for further checking of working condition of the equipment

Completed  Date : 23/12/24 Time : 10:30 AM Spare Required


Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Jineesh PP</u>	<u>23/12/24</u>	<u>10 AM</u>	<u>10:30 AM</u>

**Customer Remark** Completed  Pending

<p>Service Engineer Name : <u>Jineesh PP</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>23/12/2024</u></p> <p>Contact Number : <u>8943527662</u></p>	<p>Customer Name : <u>KAMARUNNISA OP</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>23/12/24</u></p> <p>Contact Number : <u>9074961566</u></p> <p>Designation : <u>NO GRD</u></p> <p>Hospital Seal : <u>[Seal]</u></p>
--	--




**SUPERINTENDENT**  
Women & Children Hospital  
KOZHICODE



BP. APPARATUS [ ELKO ]

DATE	Received from store	Received	Issued	Balance
17/2/17	Received from store. 2			2
	ELKO 125 (1764) Tooth Instrument pack Register (Page no 50)		2	Nil
11/1/21				
5/5/2021	Rd from store. (Digital) 3 nos. Keerthi surgicals, (RAK-268)			<del>3</del> 320002
5/5/2021	Rd from store (Digital) 04 BKF surgicals. (S-Cure)			<del>07</del> 32004455

*Sindhu*  
H/N



*nyal*  
SUPERINTENDENT  
South Women & Children Hospital  
Kozhikode



*Sindha*  
H/N



*yan*  
2/11/2022  
SUPERINTENDENT  
Govt Women & Children Hospital  
Kozhikode