

**PROFORMA**  
**RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)**  
**BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: THQH NADAPURAM	Name of District: KOZHIKODE
Name of Equipment: NEBULIZER	Equipment ID & Barcode: 152431 & 1121407
Make: LIFE CARE	Model: CNB69022
Serial Number: LC	Warranty details: No Warranty
Date of purchase/Year of manufacture/ Installation Date: In ticket master installation date is 08/08/2018, Actual installation date As per stock book is 02/10/2018	Present status of the equipment: Fully damaged
Date of breakdown: 13/12/2024 Date of registration of complaint through email/Toll free): Toll free	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: Checked equipment and found compressor motor with piston defective. Need to replace these spares for further checking working condition of the equipment.	
Recommendations for repair (required service details): Not recommending for repair	
Cost of spares (specify parts and cost): N/A	

#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: N/A	Asset Value: 1353 /-
--	----------------------



Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix service report attached

Reasons for recommending the equipment as BER: Checked and found compressor motor with piston defective. Need to replace these spares for further checking of working condition of the equipment. In ticket master installation date is 08/08/2018, as per stock book installation date is 02/10/2018 (stock book attached for reference) Equipment out lived 6 Years 2 Month with an asset cost of rupees 1353. Quotation not submitted since spares are not available in the market. So we recommend the equipment for condemnation.

JINEESH PP *Jineesh*  
 Name & Signature of CYRIX Authority with date 17/12/24

Remarks of Junior Consultant (Biomedical) NHM:  
*Checked & found that compressor motor with piston defective. Machine is installed on 2/10/2018. spares are not available in the market. So recommended for condemnation.*

Recommended for BER (Yes/ No): Yes	Consultant Bio Medical Engineer
Date: 17/12/2024	National Health Mission Kozhikode Signature of JC BM (NHM)

	 Signature of Superintendent / Medical Officer (i/c)	* Attach Photograph
--	--	---------------------

#Based on the period of life and value as per the BER guidelines








# STOCK BOOK OF


Name of article NEBULIZER

Date	Invoice No.	From whom received or to whom issued	Receipt	Issued	Balance	Batch No Exp. Date Mfg. Date	Price per Unit/MFG	Remarks
1/6/17		Brought Forward Stock Book VOL I			1	12/6/07		
10/6/17		Page 68 Shifted to Condemned as per order No. 1015/15 dated 11/23/15 Unserviceable Reg. N/H						
10/5/18		Neubulizer (Sponsored Item)		1	1	Exp MS-OCT-2015		
2/10/18		Neubulizer (Sponsored Item) Heavily		1	2	MS-3/2017		
		put in use		2				
23/03/2020		Received From the Store	1		3			
5/6/2020		Received from the store (HM)	1		4		1500/-	
		Issued for use		2				
1/4/2022		TR. in form Covid ward (Received from store on 23/8/2020)			05	(Covered word Stock Book Page No 07.)		
		Condemned as per order No C/1365/22 dated 11/2/22				10/5/19 (1) 21/0/18 (1)		
1-9-22		Pr. out from the stock book Vol. I p. No. 8		1		2 Nos.		
29/6	23	Received from store	1	1		3 Nos. 3 (L)		



For  
  
**SUPERINTENDENT**  
 TALUK HOSPITAL, NADAPURAM



**Compressor Nebulizer**  Lifecare<sup>®</sup>

Model No : CNB69022







Input : AC~220V/50HZ

Power : 180VA

Serial No.: LC

Mode of operation-continuous

No AP/APG IPXO


**CE**       **AV**


**204L211**





**CYRIX** BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)

**TOLL FREE NO: 1800 - 425-7669**

**BAR CODE - (8004-890615225) 1121407**





**AV**    

Power : VA081

Serial No.: LC

Mode of operation-continuous

No AP/APG IPXO

