

**PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: DH PALAKKAD	Name of District: PALAKKAD
Name of Equipment: SYRINGE PUMP	Equipment ID & Barcode:#150522-0912926
Make: AKAS MEDICAL SYSTEMS	Model: TERUMAX RX
Serial Number:RZE102222881	Warranty details: No Warranty
Date of purchase/Year of manufacture/ Installation Date: In ticket master installation date = NIL, As per stock book equipment installed on 16/01/2019 stock book attached for reference	Present status of the equipment: Fully damaged
Date of breakdown:04/12/2024 Date of registration of complaint through email/Toll free):Toll free	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: Checked the syringe pump found that display board, mother board and battery faulty. Need to replace display board, mother board and battery for further checking working condition of machine.	

Recommendations for repair (required service details):No recommendations

Cost of spares (specify parts and cost):

Display board = 4500/-
Mother board = 5750/-
Battery = 1950/-
12% GST = 1464/-
TOTAL = 13664/-

#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value:74.8%

Asset Value:18250/-

Abstract of Service Report provided by the OEM/Authorized Service Provider/CYRIX(Attached or Not):Cyrix Service report and Quotation attached


Reasons for recommending the equipment as BER: Checked the syringe pump found that display board, mother board and battery faulty. Need to replace display board, mother board and battery for further checking working condition of machine..As per stock Equipment installed on 16/01/2019 aged up to 5 years 10 months. Repairing cost is 74.8% both criteria for RBER met. As per tender clause 5.3.14.1 we recommending for condemnation.

AKHILESH P  31/12/24
Name & Signature of CYRIX Authority with date


Remarks of Junior Consultant (Biomedical)NHM: syringe pump installed on 16/01/2019 aged up to 5 years 10 months. Quotation attached. Repairing cost is 74.8%.


Recommended for BER (Yes/ No): Yes .

Date: 31/12/24.

Consultant (Biomedical Engg)
NHM (Arogyakeralam), Palakkad 
Signature of JCBM(NHM)

Date: 31/12/24




Signature of Superintendent/Medical Officer(i/c)

Deputy Director of Health Services
& Superintendent,
District Hospital, Palakkad

*Not mandatory

Based on the period of life and value as per the BER guidelines



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 134333

CYRIX[®]

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility..... <u>OH palakkad</u> Address..... Ph:.....	Call Registration Date : <u>4/12/2024</u> Caller ID : <u>150522</u> Date of Visit : <u>04.12.2024</u> Asset No : <u>0912926</u> EQPT Name : <u>Syringe Pump</u> Manufacture... <u>A.K.A.S</u> Model : <u>Termax RX</u> S. No <u>RZE10222881</u> Dept... <u>Cath ICU</u>
---	---

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : Not working

Action Taken : checked the syringe pump found that display board, mother board and battery is faulty. need to replace display board, mother board and battery for further checking.

Completed Date : 4/12/2024 Time : 10:00 am Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Dileep</u>	<u>04/12/24</u>	<u>10:00 am</u>	<u>10:30 am</u>

Customer Remark Completed Pending

Service Engineer Name : Dileep A.D
 Signature: [Signature]
 Date: 4/12/2024
 Contact Number: 9041975199



Customer Name : [Signature]
 Signature: [Signature]
 Date: 04/12/24
 Contact Number:
 Designation : Deputy Director of Health Services & Superintendent, District Hospital, Palakkad
 Hospital Seal

Spare Quotation

Ref: 063/12/2024
Dec 27, 2024

To,
Cyrix Health care,
Palakkad.

Dear Sir/Madam,

Subject: Estimate of AKAS Syringe & Infusion Pump

S.No	Model	Machine Sr.No	Spare parts	Qty	Total
1	Akas syringe pump	RZE1052222881	Display board	1	4500
			Mother board	1	5750
			Battery	1	1950
Total					12200
Gst 12%					1464
Grand Total					13664

Rupees (Thirteen Thousand Six Hundred and Sixty Four) only/-

Cheque/DD to be drawn in favour of 'AKAS Medical Equipment 50% advance payment and 50 % against delivery terms, 6 months replacement . Kindly approve the estimate and send us the purchase order, so as to do the needful.

For further clarification and service support, please contact us over 09884079116 or mail us at yuva.service@akasinfusions.com We value your relationship with us.

Thanking and assuring you of our best attention and services at all times.

Bank Details

**A/c Name. AKAS MEDICAL
EQUIPMENT**

Account No.00172320000479

IFSC CODE. HDFC0000017

Bank Name. HDFC BANK LTD

Branch. Anna Nagar chennai-600040

Ours sincerely
For AKAS Medical Equipment,
Yuvashree,

yuvashree





REQUIRED QUOTATION FOR SYRINGE PUMP 0912926 # 150522 @ DH PALAKKAD , PALAKKAD

5 messages

Akhilesh P <akhileshcyrix@gmail.com>
To: support@akasinfusions.com, support@akasmmedical.com
Cc: Blesson Jose <Zm2.klbemp@cyrix.in>

Tue, 17 Dec, 2024 at 1:24 pm

Dear sir

Please share the price and stock availability of

1. Control board
 2. Battery
 3. Power Supply board
- For SYRINGE PUMP at DH PALAKKAD

Deatils are furnished below

Hospital Name : DH PALAKKAD

Equipment Name : SYRINGE PUMP

Make : AKAS

Model : THERUMAX Rx

SN : RZE1052222881

Thanks & Regards

AKHILESH P
DISTRICT INCHARGE - PALAKKAD
Bemp-Kerala
Mobile : +91 7593847154
Email : akhileshcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD
30/641B, Petta, Punithura
Ernakulam- 682038
www.cyrixhealthcare.com

Mail Delivery Subsystem <mailer-daemon@googlemail.com>
To: akhileshcyrix@gmail.com

Tue, 17 Dec, 2024 at 1:24 pm



Message not delivered

Your message couldn't be delivered to **support@akasmmedical.com** because the remote server is misconfigured. See technical details below for more information.

The response from the remote server was:

554 5.7.1 : Relay access denied

[Quoted text hidden]

Akhilesh P <akhileshcyrix@gmail.com>
To: support@akasinfusions.com

Wed, 25 Dec, 2024 at 11:10 pm

Dear sir
Please update the status.

Thanks & Regards

AKHILESH P
DISTRICT INCHARGE - PALAKKAD
Bemp-Kerala
Mobile : +91 7593847154
Email : akhileshcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD
30/641B, Petta, Punithura
Ernakulam- 682038
www.cyrixhealthcare.com
[Quoted text hidden]

Akhilesh P <akhileshcyrix@gmail.com>
To: yuva.service@akasinfusions.com <yuva.service@akasinfusions.com>

Fri, 27 Dec, 2024 at 12:55 pm

Thanks & Regards

AKHILESH P
DISTRICT INCHARGE - PALAKKAD
Bemp-Kerala
Mobile : +91 7593847154
Email : akhileshcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD
30/641B, Petta, Punithura
Ernakulam- 682038
www.cyrixhealthcare.com

----- Forwarded message -----

From: **Akhilesh P** <akhileshcyrix@gmail.com>
Date: Tue, 17 Dec, 2024, 1:24 pm
Subject: REQUIRED QUOTATION FOR SYRINGE PUMP 0912926 # 150522 @ DH PALAKKAD , PALAKKAD
To: <support@akasinfusions.com>, <support@akasmedical.com>
Cc: Blesson Jose <Zm2.klbemp@cyrix.in>

[Quoted text hidden]

Yuvashree P <yuva.service@akasinfusions.com>
To: Akhilesh P <akhileshcyrix@gmail.com>
Cc: Muthukumar K <muthukumar@akasinfusions.com>, Muthumani N <muthumani@akasinfusions.com>, Nithyaprabha R <yuvashree@akasinfusions.com>, Aldrine Alphonse <aldrine@akasinfusions.com>

Fri, 27 Dec, 2024 at 1:17 pm

Dear Sir/Madam

Greetings from Akas Medical Equipment,

With reference to the discussion we had with you, we are sending you the Spare Quotation .

Please find the enclosed file for your kind reference.

Please feel free to call us back for any further clarification.

[Quoted text hidden]

—

Regards & Thanks

Yuvashree.P

AKAS MEDICAL EQUIPMENT

Service Team Leader

Contact Number : 9381179116

E-mail Id: yuva.service@akasinfusions.com


SYRINGE INFUSION PUMP

169

		Qd	id	Q/S	SD	Pr...
16-1-19	Revised from store (Rs 63000/-)	1		1		
18/1/19	Issued 0912926		1	nil	1	2/1/19
15/3/19	In out to call. b. Stock entered in stock Book F. Page no 179.		1	nil	nil	2/4/19
	Received Date SN-D (09/1/19)					

✓ M.A.S. 14

Deputy Director of Health Services
& Superintendent
District Hospital, Palakkad



CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)

TOLL FREE NO: 1800 - 425-7669

BAR CODE - (8004-890615225) 0912926



Handwritten signature

Deputy Director of Health Services & Superintendent, District Hospital, Palakkad



AKAS MEDICAL EQUIPMENT

Head Office: 25, PALAKKAD BAZAAR ROAD, MOONATTUR, TIRUCHIDAPPALI 620003, TAMILNADU, INDIA
 Office: No. 7/401, FEVA COLONY, ATHIPET, AMBATTUR, CHENNAI - 600058, TAMILNADU, INDIA
 Website: akasmedical.com

PRODUCT Name	Syringe Pump	ECI REF	
MODEL No.	TERUMAX RX	SN	R7E1052227881

Operating Voltage/Power: 220V ± 10%, AC 50 Hz, 7VA/VA
 Battery: 12V 600mAh

IPX4

MADE IN INDIA

For Support Call: 1919884079176
 Email: support@akasmedical.com

Check syringe size & Brand detection before start operating

1. Equipment should be used

Operating & Storage Condition



WARRANTY VOID IF SEAL