

**PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: PHC AYALUR	Name of District: PALAKKAD
Name of Equipment: WEIGHING MACHINE	Equipment ID & Barcode: #150251-0942396
Make: UNIGRIP	Model: UNIFICARE
Serial Number: NA	Warranty details: No Warranty
Date of purchase/Year of manufacture/ Installation Date: In ticket master Installation date – NIL. Actual installation date As per stock book is 31/08/2018	Present status of the equipment: Fully damaged
Date of breakdown: 03/12/2024 Date of registration of complaint through email/Toll free: Toll free	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: Checked the weighing machine found not switching on. Problem with display and load cell. Need to be replaced these spares for further checking working condition of Equipment.	


Recommendations for repair (required service details): No recommendations

Cost of spares (specify parts and cost): NA

#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA	Asset Value: 850/-
---	--------------------



Abstract of Service Report provided by the OEM/Authorized Service Provider/CYRIX (Attached or Not): Cyrix Service report attached.

Reasons for recommending the equipment as BER: Checked the weighing machine found not switching on. Problem with display and load cell, Need to be replaced these spares for further checking working condition of machine. In ticket master equipment installation date is NIL As per the stock book Equipment installed on 31/08/2018 aged up to 6 years 3 months with an asset cost of rupees 850. Quotation not submitted since spares are not available from market. So we recommended the equipment for condemnation.

AKHILESH P 
Name & Signature of CYRIX Authority with date 18/12/24

Remarks of Junior Consultant (Biomedical) NHM:
Equipment installed on 31/08/2018 aged up to 6 years and 3 months. Quotation not submitted. Asset Value is Rs. 850/-

Recommended for BER (Yes/ No): Yes.
Date: 20/12/24.


Signature of JCBM (NHM) Consultant (Biomedical Engg) NHM (Arogya Keralam), Palakkad

Signature of Superintendent/Medical Officer (i/c)



Date: 23/12/24

*Not mandatory # Based on the period of life and value as per the BER guidelines ASSISTANT ATTACH PHOTOGRAPH
FAMILY HEALTH CENTRE AYALUR



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 1022043

CYRIX[®]

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website :www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility.....PHC.....Ayalar..... Address.....palakkad (ctt)..... Ph:.....08590424583.....	Call Registration Date :03-12-24..... Caller ID :150251..... Date of Visit :04-12-24..... Asset No :0942396..... EQPT Name :Weighing machine..... Manufacture.....Vignip..... Model : Unitcare..... S. No.....na..... Dept.....op.....
--	--

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified :Equipment not working.....

Action Taken :Checked the weighing machine. Found not switching on. problem with display and load cell. Need to be replace these spares for further checking.....

Completed Date : 04-12-24 Time : 11:45 Am Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
Naveena-K.S	04-12-24	11:45 Am	11:45 Am

Customer Remark Completed Pending


Service Engineer Name : Naveena-k.s
 Signature:
 Date: 04-12-24
 Contact Number: 904871408

Customer Name : Mymoon-w
 Signature:
 Date: 11/21/24
 Contact Number: 9995675599
 Designation :
 Hospital Seal





Dr. S. S. Rao
ASSISTANT SURGEON
FAMILY HEALTH CENTRE
AYALUR



CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE
HEALTHCARE PVT LTD PROGRAMME (BEMP)

TOLL FREE NO: 1800 - 425-7669



KMSCL

BAR CODE - (8004-890615225)

0942396

