

**PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: PHC KANNAMBRA	Name of District: PALAKKAD
Name of Equipment: WEIGHING MACHINE	Equipment ID & Barcode: #152050- 0944978
Make: SAMSO	Model: NA
Serial Number: NA	Warranty details: No Warranty
Date of purchase/Year of manufacture/ Installation Date: 13/11/2019 stock book attached for reference.	Present status of the equipment: Fully damaged
Date of breakdown: 11/12/2024 Date of registration of complaint through email/Toll free: Toll free	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: Checked the weighing machine found not switching on. Problem with display and load cell. Need to be replaced these spares for further checking working condition of equipment.	

Recommendations for repair (required service details): No recommendations

Cost of spares (specify parts and cost): NA


#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA	Asset Value: 6000/- Actual book value of the equipment is 2590/- , stock book attached for reference.
---	--

Abstract of Service Report provided by the OEM/Authorized Service Provider/CYRIX (Attached or Not): Cyrix Service report attached.

Reasons for recommending the equipment as BER: Checked the weighing machine found not switching on. Problem with display and load cell, Need to be replaced these spares for further checking working condition of equipment. Equipment installed on 13/11/2019 aged up to 5 years. Quotation not submitted since spares are not available from market. So we recommended the equipment for condemnation.

AKHILESH P  18/12/24
Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM:
Weighing Machine installed on 13/11/2019 aged up to 5 years. Quotation not submitted. Asset value is 6000/-

Recommended for BER (Yes/ No): Yes.	 Consultant (Biomedical Engg) Palakkad NHM (Ayazur) CBM (NHM)
Date:	

Date: 24/12/24


 Seal
 Signature of Superintendent/Medical Officer (i/c)

*Not mandatory #Based on the period of life and value as per the BER guidelines * Attach Photograph
 Primary Health Centre, Kannambra, Palakkad-06



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 1022038

CYRIX[®]

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility..... <u>P.H.C. kannambra</u> Address..... <u>palakkad</u> Ph:..... <u>06238695966</u>	Call Registration Date : <u>11-12-2024</u> Caller ID : <u>152050</u> Date of Visit : <u>12-12-2024</u> Asset No : <u>0944978</u> EQPT Name : <u>weighing machine</u> Manufacture..... <u>Samso</u> Model : <u>na</u> S. No..... <u>na</u> Dept..... <u>Sub centre</u>
---	---

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : weighing machine not working.....

Action Taken : Checked the weighing machine & found not switching on. problem with display and load cell. Need to be replace these spares for further checking......

Completed Date : 12-12-24 Time : 11:30 AM..... Spare Required


Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			


Cyrux Engineer	Date	Start Time	End Time
<u>Naveena K.S</u>	<u>12-12-24</u>	<u>11:30 AM</u>	<u>11:30 AM</u>

Customer Remark Completed Pending

Service Engineer Name : <u>Naveena K.S</u> Signature: <u>[Signature]</u> Date: <u>12-12-24</u> Contact Number: <u>9048771408</u>	Customer Name : <u>Sheeba Abdul Azeez</u> Signature: <u>[Signature]</u> Date: <u>12/12</u> Contact Number: <u>9567 394 259</u> Designation : <u>Nsgottu a.I.</u> Hospital Seal
---	---




Medical Officer
Kannambra Health Centre
Palakkad-678655



SARISO





Fuc Kannambra



CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE
TECHNOLOGIES PRIVATE LTD
 PROGRAMME (BEMP)

TOLL FREE NO: 1800 - 425-7669

**BAR CODE - (8004-890615225)
 0944978**


[Signature]
 Medical Officer
 Primary Health Centre
 Kannambra, Palakkad-673558