



**PROFORMA  
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)  
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM  
(BEMP)**

Name of Hospital: CHC KUZHALMANNAM		Name of District: PALAKKAD	
Name of Equipment: BP APPARATUS		Equipment ID & Barcode: #151714 -0930126	
Make: DIAMOMD		Model: NA	
Serial Number: NA		Warranty details: No Warranty	
Date of purchase/Year of manufacture/ Installation Date: In ticket master installation date = NIL, As per stock book equipment installed on 15/11/2019 stock book attached for reference		Present status of the equipment: Fully damaged	
Date of breakdown: 10/12/2024 Date of registration of complaint through email/Toll free): Toll free		*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC	
Action taken: Checked the BP Apparatus found that mercury spilled and glass tube broken. Need to replace glass tube and mercury for further checking.			
Recommendations for repair (required service details): No recommendations			
Cost of spares (specify parts and cost): Mercury 3ml = 991/- Glass tube = 112/- TOTAL = 1103/-			
#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 83.3%		Asset Value: 1323/-	
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix Service report and Quotation attached			
Reasons for recommending the equipment as BER: Checked the BP Apparatus found that mercury spilled and glass tube broken. Need to replace glass tube and mercury for further checking. As per stock Equipment installed on 15/11/2019 aged up to 5 years 1 months. Repairing cost is 83.3% both criteria for RBER met. As per tender clause 5.3.14.1 we recommending for condemnation.			
		AKHILESH P  Name & Signature of CYRIX Authority with date 31/12/24	
Remarks of Junior Consultant (Biomedical) NHM: Bp apparatus installed on 15/11/2019 aged up to 5 years 1 month. Quotation attached. Repairing cost is 83.3% both criteria for RBER met.			
Recommended for BER (Yes/ No): <i>yes.</i>		Consultant (Biomedical Engg) NHM (Arogyakeralam), Palakkad Signature of JC BM (NHM) 	
Date: <i>31/12/24.</i>			
Date:		Superintendent Community Health Centre Kuzhalmannam - 678 702	



31-12-24  
**Superintendent**  
**Community Health Centre**  
**Kuzhalmannam - 678 702**



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

No : 134113

## CYRIX®

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

<p>Health Facility.....<u>E.H.C. Kuzhalmannam</u></p> <p>Address.....<u>palakkad (dt)</u></p> <p>Ph:.....<u>09562505109</u></p>	<p>Call Registration Date : .....<u>10-12-24</u></p> <p>Caller ID : .....<u>151714</u></p> <p>Date of Visit : .....<u>11-12-24</u></p> <p>Asset No : .....<u>0930126</u></p> <p>EQPT Name : .....<u>Bp Apparatus</u></p> <p>Manufacture.....<u>Diamond</u> Model : .....<u>na</u></p> <p>S.No.....<u>na</u> Dept.....<u>Casualty</u></p>
---	--

Service Classification : Breakdown call  PMS  Calibration  Cust. Training

Problem Identified : .....Mercury Spillage.....

Action Taken : .....Checked the Bp Apparatus and found that mercury spilled and glass tube broken. Need to replace glass tube and mercury for further checking......

Completed  Date : .....11-12-24 Time : .....11:30 AM Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrux Engineer	Date	Start Time	End Time
<u>Naveena . k . s</u>	<u>11-12-24</u>	<u>11:10 AM</u>	<u>11:30 AM</u>

Customer Remark   **Superintendent** Pending

Service Engineer Name : Naveena

Signature: [Signature]

Date: 11-12-2024

Contact Number: 9048771408

Community Health Centre  
 Palakkad - 678 702  
 Kuzhalmannam  
 Date: 11/12/2024  
 Contact Number: 9495707602  
 Designation : SARO  
 Hospital Seal



# AR MEDICARE

ROOM NO: 865, 2<sup>ND</sup> FLOOR  
AL-RAHA ARCADE, THAZHE CHOVVA  
KANNUR-670018  
EMAIL: [armedicare07@gmail.com](mailto:armedicare07@gmail.com)  
MOB: 9446057350, 7994654108,



## QUOTATION

REF: AR/QTN-113/24-25

Date: 17-08-2024

To,  
Cyrix Health Care Private Limited  
30/641B, Vandipetta Rd, near SBI Bank, Poonithura, Maradu  
Ernakulam, - 682038  
Kerala 682038

No.	Description	Qty	Unit Price(INR)	TAX	Total(INR)
1.	BP APPARATUS ACCESSORIES				
1.	MERCURY 3ML	1	835.00	151.2	991.20
2.	GLASS TUBE	1	100.00	12.00	112.00
				<b>Grand Total</b>	<b>1103.20 INR</b>
				<b>Round off</b>	<b>- 0.20 INR</b>
				<b>Net Total</b>	<b>1103 INR</b>

**INR One thousand one hundred three Only**

### Terms and Conditions for sale :

- Validity:** Six months from the date of quotation.
- Payment:** Within 30 days from the date of supply.
- Delivery:** Within 7 days after receipt of the signed purchase order.
- If any change in Govt. statutory charges like Customs duty, Excise duty, GST etc. will be applicable at the time of invoicing.
- Bank Details:

Account holder's name : AR MEDICARE  
Bank Name : State Bank Of India, Thazhechowva  
Account No. : 39648455024  
IFSC code : SBIN0016861

For AR MEDICARE,  
KANNUR




*[Signature]*  
Superintendent  
Community Health Centre  
Kuzhalmannam-678 702

580

B.P. APPARATUS (TRUCKER, ELECT)

Date	From whom received or to whom issued	Receipt	Issued	Balance	Part
15/11/19	Purchased from Kerala Community Pharmacy - DH Compound, Palakkad IN.No 52466 dt. 15/11/19 No. 12/2018 Exc. 11/23 MFR XCEL Value Rs 8210/-	3 Nos	3 Nos	NIL	✓
15/11/19	Issued to HS - 0930126		3 Nos	NIL	✓
	DIGITAL C. ADAPTER & BATTERY				
10/1/2020	Purchased from EXCEL ENTERPRISES Chakkamthara Pkd invoice no: EE 1150 date: 10/1/2020 Value: 89712.00	36 Nos	36 Nos		✓
10/1/2020	Issued to HS		36 Nos	NIL	✓



Superintendent  
Community Health Centre  
Kuzhalmannam - 678 702

**CYRIX** BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)  
TOLL FREE NO: 1800 - 425-7669



BAR CODE - (8004-890615225)  
0930126

**Diamond**<sup>®</sup>

*Regular*

B.P. APPARATUS

112 MAY.17 07407

IS: 3390



CM/L-0196043



*Zeen*  
+ Superintendent  
Community Health Centre  
Kuzhalmannam-678 702