



PROFORMA RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER) BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)	
Name of Hospital: PHC PULIKKAL	Name of District: MALAPPURAM
Name of Equipment: ECG MACHINE	Equipment ID & Barcode: #148665 - 1040447
Make: ECG MAC	Model: EM - 301
Serial Number: EM3012018030298	Warranty details: No Warranty
Date of purchase/Year of manufacture/ Installation Date: Installation date is 30/04/2018	Present status of the equipment: Fully damaged
Date of breakdown: 23/11/2024 Date of registration of complaint through email/Toll free): Toll free	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: Checked the machine and found Arm board & Printer head defective. Need to replace these spares for further checking.	
Recommendations for repair (required service details): No Recommendation	
Cost of spares (specify parts and cost): Printer head: 14200/- Arm board : 9700/- TOTAL: 23900/-	
#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 79%	Asset Value: 29900/-
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix service report attached & Oem Quotation attached.	
Reasons for recommending the equipment as BER: Checked the machine and found Printer head & Arm board defective. Installation date is 30/04/2018. Aged up to 6 years & 7 month, repairing cost is 79% of asset value, both criteria for RBER met. As per the tender clause 5.3.14.1 We recommending the equipment the equipment for condemnation.	
GOKUL K. I # 7/12/24 Name & Signature of CYRIX Authority with date	
Remarks of Junior Consultant (Biomedical) NHM: <i>Verified with the damage of printer head &amp; Arm board. Beyond economic repairs.</i>	
Recommended for BER (Yes/ No):	 Jr. Consultant B. # 7/12/24 NHM Malappuram Signature of JOC
Date: 7/12/24	
Date: _____ Officer (i/c)	 MEDICAL OFFICER PHC PULIKKAL - 673 637 Signature of PHC Medical Officer

\*Not mandatory

##Based on the period of life and value as per the RBER guidelines

\* Attach Photograph



**BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAMME  
UNDER  
NATIONAL HEALTH MISSION**



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

No : 1019251

**CYRIX**<sup>®</sup>  
HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala  
Ph : 98472 99500 Website :www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility..... <u>P H C</u>	Call Registration Date : ... <u>23-11-2024</u>
Address..... <u>PULIKKAL</u>	Caller ID : ... <u>148665</u>
..... <u>(MALAPPURAM)</u>	Date of Visit : ... <u>25-11-2024</u>
Ph: ... <u>9847422694</u>	Asset No : ... <u>1040447</u>
	EQPT Name : ... <u>E C G machine</u>
	Manufacture. <u>E C G mac</u> Model : <u>EM-301</u>
	S. No <u>EM3012018030298</u> Dept. <u>E C G Room</u>

Service Classification : Breakdown call  PMS  Calibration  Cust. Training

Problem Identified : ... E. c g machine not on .....

Action Taken : ... checked the machine and found printer head & arm board defective. Need to replace these spares for further checking. .....

Completed  Date : 25-11-2024 Time : 2:00pm Spare Required


Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Noufal M</u>	<u>25-11-24</u>	<u>11:50am</u>	<u>2:00pm</u>

Customer Remark:  Completed  Pending

Service Engineer Name	Customer Name :
<u>Noufal M</u>	<u>PULIKKAL</u>
Signature: <u>Noufal</u>	Signature: <u>[Signature]</u>
Date: <u>25-11-2024</u>	Date: <u>25-11-2024</u>
Contact Number: <u>7593847177</u>	Contact Number: <u>9845386513</u>
	Designation : <u>Nursing officer</u>
	Hospital Seal



[Signature]  
MEDICAL OFFICER  
PRIMARY HEALTH CENTRE



Gokul K T 1 Dec

Dear sir, There is a complaint reported from PHC PULIKKAL - MALAPURAM,

2



PADMANA... 3 days ago



to me, Blesson

Dear Gokul,

Prices for the required items are as given below

● Arm board – INR 9700 plus Tax

● Printer head – INR 14200 Plus Tax

**S Padmanaban**  
Deputy General Manager (ICU Business)

**TRIVITRON**  
HEALTHCARE  
enabling your serenity

M : +91 87545 80897  
Triviron Healthcare D-117, TTC Industrial  
Area, Nerul,  
Navi Mumbai 400 706, India



**PRICE ENQUIRY 1040447 ECG MACHINE @ PHC PULIKKAL -MALAPURAM (Ticket id 148665)**

4 messages

**Gokul K T** <gokulcyrix@gmail.com>  
To: padmanaban.s@kiranxray.com  
Cc: Blesson Jose <Zm2.klbemp@cyrix.in>

Sun, 1 Dec, 2024 at 10:34 pm

Dear sir,

There is a complaint reported from PHC PULIKKAL - MALAPURAM, Regarding the complaint with ECG MACHINE ECG mac , EM 301 , serial Number - EM3012018030297 barcode number (1040447). Required price quotation for following spares, kindly share the estimate.

- Arm board
- Printer head

Thanks & Regards

GOKUL KT  
CYRIX HEALTHACARE  
DISTRICT INCHARGE  
MALAPPURAM  
9809314280

**Gokul K T** <gokulcyrix@gmail.com>  
To: padmanaban.s@kiranxray.com  
Cc: Blesson Jose <Zm2.klbemp@cyrix.in>

Tue, 3 Dec, 2024 at 7:05 am

Reminder  
[Quoted text hidden]

**Gokul K T** <gokulcyrix@gmail.com>  
To: padmanaban.s@kiranxray.com  
Cc: Blesson Jose <Zm2.klbemp@cyrix.in>

Tue, 3 Dec, 2024 at 10:27 pm

Reminder  
[Quoted text hidden]

**PADMANABAN S** <padmanaban.s@kiranxray.com>  
To: Gokul K T <gokulcyrix@gmail.com>  
Cc: Blesson Jose <Zm2.klbemp@cyrix.in>

Wed, 4 Dec, 2024 at 7:42 am

Dear Gokul,

Prices for the required items are as given below

- Arm board – INR 9700 plus Tax
- Printer head – INR 14200 Plus Tax

# STOCK REGISTER

1



Name of Article : ECG MACHINE THREE - CHANNEL

Date	Received from or issued to	Inv. No or V. No	Amount		Receipts	Issues	Balance	Remarks
			Rs.	Ps.				
30/4/18	Rd. from TAMIRAM Healthcare Pvt Ltd. No. 260/1 - D, DOORNOOR - alle bypass road, Doornoor alle Chennai, Tamil Nadu (Rd. from ICRSU Thiruvananthapuram) as per Invoice No. 301002416 dtd 30/03/2018	-						
	Barcode: 1040447							
	Item code - CLS-70042, Cap code 243 Hex 90181100 mfg - Shenzhen GRAMAC 165T @ 12% Medical electronics Co Ltd.					1210.	1210.	3
	Model: EM-301, Project Name - DHS 2017-18 Developing EM3012018030298 -> Serial No.							Super speciality facilities.
	Accessories supplied							
	1. Patient cable				2 spf			
	2. clip electrodes				8			
	3. Chest electrode with silicon rubber bulb.				12			
	4. Recording paper				1 roll			
	5. ECG Jelly				1 bt			
	6. Power cable				1			
	7. Ground cable				1			
	8. User manual				1			
11/2/18	Issued to sln, all the above items.				NK			
					Received	Bobby: 03		
						Street Name.		



**MEDICAL OFFICER**  
**PRIMARY HEALTH CENTRE**  
**PULIKKAL - 673 697**

**MEDICAL OFFICER**  
**PRIMARY HEALTH CENTRE**  
**PULIKKAL - 673 697**



MEDICAL OFFICER  
PRIMARY HEALTH CENTRE  
PULIKKAL - 673 631