


PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: PHC OLLUKKARA	Name of District: Thrissur
Name of Equipment: Weighing Machine	Equipment ID & Barcode: #150871- 0822918
Make: SAMSO	Model: NA
Serial Number: NA	Warranty details: No Warranty
Date of purchase/Year of manufacture/ Installation Date: In Ticket master installation date is- NA As per Stock book installation date is - 21-11-2019	Present status of the equipment: Fully damaged
Date of breakdown: 05-12-2024 Date of registration of complaint through email/Toll free): Toll free	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: checked the machine and found display and sensor are defective. Need to replace these spares for further checking working condition of machine.	
Recommendations for repair (required service details): Not Recommending	
Cost of spares (specify parts and cost): NA	


#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA	Asset Value: 1600/-
---	---------------------

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix report attached.

Reasons for recommending the equipment as BER: Checked the equipment and found that the display and sensor are defective. The equipment was installed on 21-11-2019, and is over 5 years old with an asset cost of ₹1600. A quotation has not been submitted as the necessary spares are not available in the market. Therefore, we recommend the equipment for condemnation


BIJO T JOY
 Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM:
weighing machine's spares are not available.

Recommended for BER (Yes/ No): <i>Yes</i>	 Signature of Jr. Consultant (Biomedical) NHM (NHM)
Date:	

Date:	Seal	 Signature of Superintendent / Medical Officer (i/c)
-------	------	--

*Not mandatory //Based on the period of life and value as per the BER guidelines *Attach Photograph



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 138046

CYRIX[®]

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website :www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility..... <u>Phc Ollukkara</u> Address..... <u>Thrissur</u> Ph:.....	Call Registration Date : <u>5/12/24</u> Caller ID : <u>150871</u> Date of Visit : <u>6/12/24</u> Asset No : <u>0822918</u> EQPT Name : <u>weighing machine</u> Manufacture..... <u>s.o.mso</u> Model : <u>nil</u> S. No..... <u>nil</u> Dept..... <u>Public Health</u>
---	--

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : weighing machine not working - Display issue.....

Action Taken : checked the equipment and found that the display and sensor are defective need to replace these spares for further checking working condition of the machine......

Completed Date : 6/12/24..... Time : Spare Required


Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrux Engineer	Date	Start Time	End Time

Customer Remark Completed Pending Medical Officer in Charge Primary Health Centre

Service Engineer Name : <u>Suhes.p.Rajan</u> Signature: <u>[Signature]</u> Date: <u>6/12/24</u> Contact Number: <u>7593847135</u>	Customer Name: <u>[Signature]</u> Signature: <u>[Signature]</u> Date: <u>6/12/24</u> Contact Number: <u>805932733</u> Designation : <u>STAFF NURSE</u> Hospital Seal
--	---





KARUNYA MEDICINE DEPOT

SWARAJ ROUND EAST, TRISSUR

PH: 0487-2975601 EMAIL : kcpdepottr@gmail.com

State : Kerala

State Code : 32

DL No1: KL-TSR-139473, 1

DL No2: KL-TSR-139474, 1

[DUPLICATE]

INVOICE

GSTIN : 32AADCK4029M1ZK

Name : TSR-Ollurkara-PHC
 Cust.Code : Ollur
 Address : Medical Officer I/c
 Primary Health Centre
 Ollurkara


*Kit 1
Maincentre.*

Invoice No : 5965
 Date : 21/11/2019
 Time : 01:48 PM
 Prepared By : SINDHU K B

Phone No :
 GSTIN :
 DL No1 :
 DL No2 :

Page 2 of 2

SI#	MFR	Product Name	KCP S.P	Exp.Dt	MRP	Batch	Unit	Qty	Rate	CGST %	CGST Amt	SGST %	SGST Amt	Amount
17	GE	SALT TESTING KIT FOR IODINE (70 TEST/KIT) - I CHECK - GLOBAL BIOSCIENCES	21.80	Oct/2020	28.00	IC-655	NOS	1.0	21.80	6.00	1.31	6.00	1.31	21.80
18	GE	STADIOMETER (EASYCARE / KRIVES) - LIFECARE	467.80	Dec/2024	1250.00	1122	NOS	1.0	467.80	9.00	42.10	9.00	42.10	467.80
19	GE	STETHOSCOPE (SUPER DELUXE - LIFECARE /SE)	308.04	Dec/2028	900.00	111	NOS	1.0	308.04	6.00	18.48	6.00	18.48	308.04
20	GE	THERMOMETER DIGITAL (LIFELINE - LIFECARE)	67.76	Dec/2028	235.00	1107	NOS	1.0	67.76	9.00	6.10	9.00	6.10	67.76
21	GE	URINE STRIP (GLUCOSE & ALBUMIN, 100 TEST/KIT - MISSION) ACON - LIFECARE	102.68	May/2021	434.00	URS9050037	NOS	10.0	102.68	6.00	61.61	6.00	61.61	1026.80
22	GE	WEIGHING MACHINE (DIGITAL PAEDIATRIC - PE) SAMSO - DR. BELLI RAM & SONS	4239.44	Dec/2030	6000.00	111	NOS	1.0	4239.44	9.00	381.55	9.00	381.55	4239.44
23	GE	WEIGHING MACHINE (DIGITAL ADULT - SCALE CAPACITY 150 KGS - PE) SAMSO - DR. BELLI RAM & SONS	1608.04	Jan/2028	2590.00	113	NOS	1.0	1608.04	9.00	144.72	9.00	144.72	1608.04




Am
 Medical Officer in-charge
 Primary Health Centre
 OLLURKARA

CGST: 1419.28	SGST: 1419.28	CESS: 189.17	Total: 21945.10
			Handling Charge: 1324.22
			Handling Charge Gst18% : 238.36
			Round off: 0.32
MRP Amount: 47381.00			Net Amount: 23508.00

Rupees: Twenty Three Thousand Five Hundred and Eight

Terms & Conditions

- Sales returns only within 10 days from the date of billing
- Sales return -10AM to 4PM Only
- No Sales return on Sunday
- Products to be kept in refrigerator will not be taken back after sales.



AUTHORIZED SIGNATORY

CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)



TOLL FREE NO: 1800 - 425-7669

BAR CODE - (8004-890615225)
0822918

