


**PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: PHC VALAPATTANAM	Name of District: KANNUR
Name of Equipment: WEIGHING MACHINE	Equipment ID & Barcode: #153717 - 1340378
Make: HEALCURE	Model: N/A
Serial Number: NA	Warranty details: No Warranty
Date of purchase/Year of manufacture/ Installation Date: In ticket master installation date is 20-07-2018, As per the stock book installation date is 10-01-2018 (Stock book attached for reference)	Present status of the equipment: Fully damaged
Date of breakdown: 20-12-2024 Date of registration of complaint through email/Toll free): Toll free	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: Checked the weighing machine found error err showing. Found Problem with display and mainboard. Need to be replace these spares for further checking of machine.	
Recommendations for repair (required service details): No recommendations	
Cost of spares (specify parts and cost): NA	


#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA	Asset Value: 850/-
---	--------------------

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix Service report attached.

Reasons for recommending the equipment as BER: Checked the weighing machine found error err showing. Found Problem with display and mainboard. Need to be replace these spares for further checking of machine. In ticket master installation date is 20-07-2018, As per the stock book installation date is 10-01-2018 (Stock book attached for reference) aged up to 6 years and 11 months. Quotation not submitted since spare are not available from market. So we are recommended the equipment for condemnation.

Aswis 
31/12/24
Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM: checked and identified that machine not working, due to mainboard issue. Aged up to 6 years and above. Quotation not submitted. Since the spare not available. So recommended for BER as per BER Protocol.

Recommended for BER (Yes/ No): Yes.	CONSULTANT NATIONAL HEALTH COMMISSION KANNUR DISTRICT	 31/12/24	Signature of JC BM (NHM)
-------------------------------------	---	--	--------------------------

Date: 30/12/24	Seal	Signature of Superintendent Medical Officer (i/c)
BINDU. N.M. N/A 31/12/2024		



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 126702

CYRIX[®]

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

<p>Health Facility: <u>PHC VALAPATTANAM</u></p> <p>Address: <u>VALAPATTANAM</u> <u>KANNUR</u></p> <p>Ph: <u>9895478862</u></p>	<p>Call Registration Date: <u>20/12/24</u></p> <p>Caller ID: <u>153717</u></p> <p>Date of Visit: <u>21/12/24</u></p> <p>Asset No: <u>1340378</u></p> <p>EQPT Name: <u>weighing Machine</u></p> <p>Manufacture: <u>Healure</u> Model: <u>na</u></p> <p>S. No. <u>na</u> Dept. <u>OP</u></p>
--	--

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : Showing error err

Action Taken : checked the weighing machine found error err showing. Found problems with display and mainboard. Need to replace these spares for further checking of machine.

Completed Date : 21/12/24 Time : 12:10 PM Spare Required


Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer Aswin Date 21/12/24 Start Time 11:40 AM End Time 12:10 PM

Customer Remark Completed Pending

Service Engineer Name : Aswin
Signature: [Signature]
Date: 21/12/24
Contact Number: 8801012199



Customer Name : SINDU N.M.
Signature: [Signature]
Date: 21-12-2024
Contact Number: [Number]
Designation : [Text]
Hospital Seal: [Text]

WEIGHING MACHINE


Date		Qty	Cond	Balance	Sd/-	Sd/-
1/12/11	Tr. in from old book vol 3 page no: 63			2.	✓	
14/6/13	Auctioned & condemned as per order no 63/2013 dated 14/6/13 MO PHC Valapattanam		2	NIL	[Signature]	[Signature]
10/1/18	Received from the store Digital Adult scale capacity 150kgs (Khalcuure - Mahaveer enterprises) Batch number 24361 ₹ 976.57/- each Expiry date Dec 2022 Aardhan kmscl.	2		2	[Signature] HN Jayasree	
1/6/18	Tr. out to new stock book page number 69 vol 3 [Signature] HN Jayasree					



മുഖ്യമന്ത്രിയുടെ
കുടുംബാരോഗ്യകേന്ദ്രം
വല്ലാട്ടണം
[Signature]
12/12/2022



Max 150kg d=100g



Handwritten signature and date '12/12/2024' above a printed name 'മെഡിക്കൽ ഓഫീസർ' and title 'കുടുംബാരോഗ്യകേന്ദ്രം വളപ്പട്ടണം'.