

**PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: THQH PAYANGADI	Name of District: KANNUR
Name of Equipment: NEBULIZER	Equipment ID & Barcode: #150869 - 1320753
Make: Easy care	Model: EC-7020
Serial Number: 07261	Warranty details: No Warranty
Date of purchase/Year of manufacture/ Installation Date: In ticket master Installation date is 25-02-2020, As per stock installation date is 21-04-2022(stock book attached for reference)	Present status of the equipment: Fully damaged
Date of breakdown: 05-12-2024 Date of registration of complaint through email/Toll free): Toll free	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: Checked the nebulizer found motor stuck. Problem with motor and piston, need to be replace these spares for further checking of machine.	
Recommendations for repair (required service details): No recommendations	
Cost of spares (specify parts and cost): NA	
#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA	Asset Value: 1353/-
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix Service report attached.	
Reasons for recommending the equipment as BER: Checked the nebulizer found motor stuck. Problem with motor and piston, need to be replace these spares for further checking of machine. In ticket master installation date is 25-02-2020, As per the stock book Equipment installed on 21-04-2022 aged up to 2 years 8 months (stock book attached for reference). Quotation not submitted since spare are not available from market. So we are recommended the equipment for condemnation	
<p>Vishnu P <i>Vishnu</i> Name & Signature of CYRIX Authority with date</p>	
Remarks of Junior Consultant (Biomedical) NHM: checked and identified that machine not working due to motor, Piston need to replace. Aged up to 2 years and 8 months. Quotation not submitted. The spare cost is economically not feasible. so recommended for BER as per BER Protocol.	
Recommended for BER (Yes/ No): Yes.	<p><i>Rathna</i> 13/12/24 CONSULTANT BIOMEDICAL NATIONAL HEALTH MISSION Signature of JCBM (NHM)?</p>
Date: 13/12/2024.	
Date: 20/12/2024	<p><i>Om</i> Signature of Superintendent Medical Officer (i/c)</p>



*Not mandatory

//Based on the period of life and value as per the BER guidelines. Attach Photograph



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 128332

CYRIX[®]
HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility: <u>Taluk Hospital</u>	Call Registration Date : <u>05.11.2024</u>
Address: <u>Payangadi</u>	Caller ID : <u>150869</u>
<u>Kannur</u>	Date of Visit : <u>06.12.24</u>
Ph: <u>9946483944</u>	Asset No : <u>1320753</u>
	EQPT Name : <u>Nebulizer</u>
	Manufacture: <u>Easy care</u> Model: <u>EC7020</u>
	S. No. <u>07261</u> Dept. <u>OP</u>

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : Motor not working

Action Taken : checked the machine and found motor stuck. Problem with motor and piston, need to replace these spares for further checking of machine

Completed Date : 06.12.24 Time : 11:00 AM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Vishnu P</u>	<u>06/12/24</u>	<u>10:00 AM</u>	<u>11:00 AM</u>

Customer Remark: Completed Hospital Stamp: GOVT TALUK HOSPITAL PAYANGADI KANNUR Status: Pending

Service Engineer Name : Vishnu P
Signature: [Signature]
Date: 06/12/24
Contact Number: 9072011476

Hospital Seal: [Seal]
Signature: [Signature]
Date: 6/12/24
Contact Number: 9946483944

Nebulizer


67

Date	Description	Received	Issue	Balance	Signature	
					Hen	Mo
15/3/11	Register Tr. in from old stock	16.10.09-1				
11-6-12	Received from store - Ultrasonic (Rs. 5046) put in use - wood	3.2.10-1 TALO		2 Nos.	OK	b.
20.2.13	Received from store - Ultrasonic (Rs. 5045.95 x 2) Model. TSD 50004			3 No.	Pin	
27.6.2011	Received from store (misc) (HMC) (Rs. 1740 x 1)	2 Nos 1 unit 2 Nos		5 Nos.	Pin	
15-1-2016	Received from store (Nebulizer - BPL) put in use - OP, wood	2		7	Pin	
3/10/2017	Received from store (pulmo aid) (Rs. 5357 x 2)	2		9	Pin	
25/9/2019	Received from store (Nedemast - Glenmark) (KMSCL) (small) put in use OP + wood	1		12	Pin	
22/2/2019	Condemned as per order No. 93/2019 dtl 22/2/2019 (suppl) T.H. PZDI is disposed in auction		6	6		
25/02/20	Received from store (LP) Rs 2850/- (Merepan Nebulizer) put in use Pool 801334/11			173	Pin	
25/2/20	Received from store (M.A Fund) (Donated items) Bunal 801324/10 801334/09 (Life care)	2		92	Pin	
11/11/2021	Tr. out to unservicable book p. no 118 16/2016-27/6/2014-1					
21/4/22	Received from store Rs: 1550 (BATCH No. 90192090 NA) Exp: 31/12/25 (Easy Care)	2		6 8		
11/4/22	Tr out to unservicable book order no 386/2022 Hospital, Parbhayanganj, order no. 815/2022 dtl	2				



Assistant Surgeon
TALUK HOSP
PARBHAYANGANJ

12/12/2022 Suppl
Taluk Hospital Parbhayanganj - 2
At. Balan 8/



CYRIX MEDICAL EQUIPMENT MAINTENANCE
 PROGRAMME ENTRY
 TOLL FREE NO: 1800 - 425-7669
 BAR CODE - (8004-890615225)
 1320753


CYRIX MEDICAL EQUIPMENT MAINTENANCE
 PROGRAMME ENTRY
 TOLL FREE NO: 1800 - 425-7669
 BAR CODE - (8004-890615225)
 1320753

Compressor Nebulizer
 Model No: ECT020
 Input: AC 220V/50Hz 0.7A
 Lot: 11119
 S/N: 07201
 Mode of operation: continuous
 No AP/APG IPX0

Compressor Nebulizer
 Model No: ECT020
 Input: AC 220V/50Hz 0.7A
 Lot: 11119
 S/N: 07201
 Mode of operation: continuous
 No AP/APG IPX0

CYRIX
 CALON :
 ID :
 PRO :
 SN :

CYRIX
 CALON :
 ID :
 PRO :
 SN :



17
 17

Om
 ASSISTANT SURGEON
 TALUK HOSPITAL
 PAZHAYANGUDI

GOVT. TALUK HOSPITAL
 PAZHAYANGUDI
 PIN: 670 303