


**PROFORMA  
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)  
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: DH KANNUR	Name of District: KANNUR
Name of Equipment: Cautery	Equipment ID & Barcode: # 153016- 1310920
Make: SHALYA	Model: SHALYA MX
Serial Number: M16B04B	Warranty details: No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 23-05-2016	Present status of the equipment: Fully damaged
Date of breakdown: 17-12-2024 Date of registration of complaint through email/Toll free): Toll free	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: Checked the equipment and found that cut and coagulation not working. Found main board, power supply board, key pad unit defective. Need to replace these spares for further checking of the equipment. Enquired oem for quotation	
Recommendations for repair (required service details): No recommendations	


Cost of spares (specify parts and cost): N/A	
#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: N/A	Asset Value: 209500 /-


Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix service Report and Oem Letter attached

Reasons for recommending the equipment as BER: Checked the equipment and found that cut and coagulation not working. Found main board, power supply board, key pad unit defective. Need to replace these spares for further checking of the equipment. Enquired oem for quotation. In ticket master installation date is 23-05-2016 aged up to 8 years and 6 months. Oem informed that this machine already passed more than 8 years. So instead of purchasing main board of unit, requesting to purchase new cautery machine. Because the cost of main board is almost equal to the price of new cautery machine. After the end of working life of cautery machine leads to deterioration of the internal electronic spare parts which may create difficulty during the surgical procedure. So As per tender clause 5.3.14.2 we recommending the equipment for condemnation.


  
 30/12/2024  
 Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM: checked and identified that machine not working due to main board issue, power supply board and . Aged up to 8 years and above. OEM issue the EOL of the same machine. So recommended for BER as per BER Protocol.

Recommended for BER (Yes/ No): <b>yes</b>	 30/12/24 Signature of JC BM (NHM)
Date: 30/12/24	

Bindu. S  
 Senior nursing officer  
  
 Date: \_\_\_\_\_ Seal \_\_\_\_\_ Signature of Superintendent / Medical Officer (i/c) \_\_\_\_\_

\*Not mandatory Based on the period of life and value as per the BER guidelines \* Attach Photograph





Date: 27<sup>th</sup> Dec 2024

**To,**  
**District Hospital Kannur**  
At Post, Ayikkara Government Hospital Rd  
Burnacherry, Kannur, Kerala 670017  
Kind Atten. – District Hospital (Divisional Manager)  
North Zone- kannur, Kerala.

Dear Sir,

**SUB: Condemnation Letter**

With reference towards above, we would like to inform you that international shelf life of SHALYA unit is 7 years & extended life is till 10 years, same depends on the usage.

It has been observed that, the below mentioned SHALYA Cautery's, already passed more than 8 years. So, instead of purchasing main board of unit, request to purchase the new Cautery Machine. Because the cost of Main Board is almost equal to the price of New Cautery Machine.

Sr. No.	Model	Sr. No. of the unit	Purchased Year
1	SHALYA MX	M16B04B	2016

After the end of working life of Cautery Machine leads to deterioration of the internal electronic spare parts, which may create difficulty during surgical procedure.

We request to Authority / Department to condemn this unit and purchase a new unit.

Thanking you,

**For Xcellance Medical Technologies Pvt Ltd.**


Authorized Signatory

**Xcellance Medical Technologies Pvt. Ltd.**

W-239, Rabale MIDC, TTC Industrial Area, Rabale, Navi Mumbai - 400 701, INDIA

+91-22-2760 2261 / 2 / 3 +91-22-2760 2263 [www.shalya.in](http://www.shalya.in) [info@shalya.in](mailto:info@shalya.in)



Sarang KM <sarangcyrix@gmail.com>

---

## Required Quotation 1310920 #153016 DIATHERMY @ DH KANNUR

2 messages

---

Sarang KM <sarangcyrix@gmail.com>  
To: Praseetha Ants <praseetha.ants@gmail.com>  
Cc: BLESSEN JOSE <zm2.klbemp@cyrix.in>

Wed, Dec 25, 2024 at 11:25 PM

Dear Sir

Please share price and spare availability of MAIN BOARD, POWER SUPPLY UNIT, KEY PAD UNIT

Barcode- 1310920  
Make: SHALYA  
Model: SHALYA MX  
SN: M16B04B

--  
**Thanks & Regards**

**Sarang.k.m**  
**Sr.Biomedical Engineer**  
**Bemp-Kerala**  
**Mobile : +91 7593847150**  
**Email : sarangcyrix@gmail.com**

**CYRIX HEALTHCARE (P) LTD**  
**30/641B | Petta | Punithura | Kochi | Kerala- 682038**  
**www.cyrixhealthcare.com**

---

Praseetha Ants <praseetha.ants@gmail.com>

Fri, Dec 27, 2024 at 5:10 PM

To: Sarang KM <sarangcyrix@gmail.com>  
Cc: BLESSEN JOSE <zm2.klbemp@cyrix.in>, Ravi Kumar <ravi.antslifecare@gmail.com>, Manulal H <manulal.ants@gmail.com>

Dear Sarang,

**Greetings from ANTS Lifecare Private Limited.!!!**

As per your requirement, please find the condemnation letter attached for your reference.

Thanks & Regards

*Praseetha K*  
*Asst. Manager - Business Development,*  
*Mob: 9567860999*



HOUSE OF ANTS, 55/3052,

TNRWA - 16, Tagore Nagar,

Ponneth Temple Road, Kadavanthra,

Ernakulam, Kerala - 682020

Tel : 0484 - 2323045/ 3512580

[Quoted text hidden]



**Condemnation letter for ANTS Lifecare Private Limited 27.12.pdf**

205K


STOCK REGISTER OF MEDICINES

Name of Drug: *Electric Country*

Date	From whom Received or to whom issued	Receipt	Issue	Balance	Batch No.	Exp. Date	Price	Remarks
1/9/15	BF from vol IV Stock							
	register page no. 119			3				✓
i →	Rd on 17/8/2001						53,200/-	
ii →	where abants not available							
iii →	Rd on 23/5/16						21975/-	
	Shalys Mr Electro Surgical		month					
	kmscc Supply		unit					
	3 years warranty up to		21/5/19					
	IN no. SR Vol III page no. 40							
					Service No - 0482325:492			
					9895052846			
29/5/20	II Country In and to (Page 22)		1	2				✓
	Unserviceable repaired							
	condemned as per order no. 93-83-5/2021 dated						15722	
14/6/21	In. on from Eye Clinic							⊗
BINDU.S	Stock register of equipments							
Senior nursing officer	Vol. V Page no. 107	1		3			1,32,720	
<i>Bindu S</i>	<i>Shalys</i>							
<i>31/10/22</i>	<i>(kmscc Supply)</i>							
5/11	from the store	i		4	(LP/20/21/24)		94,964.00	
	my machine including 11 accessories							
	del. Country 3000 Euro each							
	Serial No. 2122060-3 Installation date 4/1/25							
	company from 4/1/25 to 3/1/22							



→ In and to Page No. 101



Bindu . S  
Senior almsy officer  
Bindu.S.

30/11/24

SUPERINTENDENT  
DISTRICT HOSPITAL  
KANNUR

