

PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: PHC CHERUKUNNU	Name of District: KANNUR
Name of Equipment: Sterlizer (In ticket master equipment name is sterilizer , actual equipment is Autoclave)	Equipment ID & Barcode: #145976-1341629
Make: Shiva	Model: N/A
Serial Number: N/A	Warranty details: No Warranty
Date of purchase/Year of manufacture/ Installation Date: In Ticket master installation date is 02-05-2020, as per stock book installation date is 24-03-2017(stock book copy attached for reference)	Present status of the equipment: Fully damaged

Date of breakdown: 11-11-2024 Date of registration of complaint through email/Toll free): Toll free	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
--	---

Action taken: Checked the equipment found leakage and coil not heating, found that sterilizer coil and cord wire, gasket, locking knob, door and outer chamber defective. Need to replace these spare for further checking of the autoclave.

Recommendations for repair (required service details): Not recommending for service

Cost of spares (specify parts and cost): ELEMENT FOR AUTOCLAVE- 680/-
 AUTOCLAVE KNOB-450/-
 AUTOCLAVE GASKET-980/-
 OUTER CHAMBER-5000/-
 CODE WIRE-650/-
 AUTOCLAVE DOOR-4000/-
TOTAL-11760/-

#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 78.4%	Asset Value: 15000/-
--	----------------------



Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix service Report and vendor quotation attached

Reasons for recommending the equipment as BER: Checked the equipment found leakage and coil not heating, found that sterilizer coil and cord wire, gasket, locking knob, door and outer chamber defective. Need to replace these spare for further checking of the autoclave. Enquired vendor for quotation. In Ticket master installation date is 02-05-2020, as per stock book installation date is 24-03-2017(stock book copy attached for reference) aged up to 7 years 9 months. Spare Cost is 78.4%. Both criteria RBER met. As per tender clause 5.3.14.1 we recommending the equipment for condemnation.

Vishnu P Vishnu
 Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM: Sterilizer not working due to the coil, gasket, locking knobs, chambers are defective. Leakage also found. Age upto 7 years and above. As per quotation spare cost is about 78.4%. So recommended for BER.

Recommended for BER (Yes/ No): Yes	Signature of JC.BM (NHM) <i>Jeethu</i> 13/12/24
Date: 13/12/2024.	

Date: 16/12/2024		Signature of Superintendent / Medical Officer  MEDICAL OFFICER (JC) FAMILY HEALTH CENTRE CHERUKUNNU THARA-670301
------------------	---	---

*Not mandatory

#Based on the period of life and value as per the BER guidelines



Sreeraj Surgicals

GST No: 32ABYFS3674H1ZL
DI No : 13/023/20B/2011
13/024/21B/2011
Dt.: 29-01-2011

Aiswarya Commercial Complex, Nr. Ashirvad Hospital, Thavakkara, Kannur - 670 002
Ph: 0497 2706307. Mob: 9349846307, 9387529848
Email: sreerajsurgicals@gmail.com, rajeshsreerajsurgical@gmail.com
www.sreerajsurgical.com Follow us on  www.facebook/ sreerajsurgicals

Date:

QUOTATION

DATE: 13-12-2024

TO
CYRIX HEALTHCARE (P) LTD

SIR,

THANK YOU FOR YOUR ENQUIRY TO THE RATE OF THE FOLLOWING ITEMS.

SL.NO	ITEMS	AMOUNT
1	ELEMENT FOR AUTO CLAVE	680.00
2	AUTOCLAVE KNOB	450.00
3	AUTOCLAVE GASKET	980.00
4	OUTER CHAMBER	5000.00
5	CODE WIRE	650.00
6	AUTOCLAVE DOOR	4000.00
TOTAL		11760.00 .

WE EXPECT YOUR VALUABLE ORDERS AND KIND CO-OPERATION WITH US
GST INCLUDED



THANKING YOU

Authorised Dealer:



Suppliers of Surgical Instruments, Hospital Furnitures, X-ray Materials, Lab Equipments & Disposables



Autoclave quotation required #1341629, 145976 @PHC CHERUKUNNU

7 messages

avinash t <avinashtcyrix@gmail.com>
To: sreerajsurgicalsg@gmail.com

Fri, 15 Nov, 2024 at 8:01 am

Dear sir,
Please provide a quotation of below mentioned parts of autoclave
1. Element for autoclave
2. Autoclave knob
3. Autoclave gasket
4. outer chamber
5. code wire
6. autoclave door

Thanks & Regards

Avinash T
District charge, kannur
bemp kerala
mobile: +91 6364064948
E mail: avinashtcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD
30/641B | Petta | Punithura | Kochi | Kerala- 682038
www.cyrixhealthcare.com

avinash t <avinashtcyrix@gmail.com>
To: sreerajsurgicalsg@gmail.com, Sarang KM <sarangcyrix@gmail.com>

Fri, 13 Dec, 2024 at 10:25 am

@Sarang KM
[Quoted text hidden]

avinash t <avinashtcyrix@gmail.com>
To: sreerajsurgicalsg@gmail.com, Sarang KM <sarangcyrix@gmail.com>

Fri, 13 Dec, 2024 at 10:25 am

#reminder
[Quoted text hidden]

sreerajsurgicalsg <sreerajsurgicalsg@gmail.com>
To: avinash t <avinashtcyrix@gmail.com>

Fri, 13 Dec, 2024 at 10:42 am

Sent from Mail for Windows
[Quoted text hidden]

avinash t <avinashtcyrix@gmail.com>
Draft to: Sarang KM <sarangcyrix@gmail.com>

Fri, 13 Dec, 2024 at 10:53 am

—— Forwarded message ——
From: avinash t <avinashtcyrix@gmail.com>
Date: Fri, 15 Nov, 2024, 8:01 am
Subject: Autoclave quotation required #1341629, 145976 @PHC CHERUKUNNU
To: <sreerajsurgicalsg@gmail.com>

Dear sir,
Please provide a quotation of below mentioned parts of autoclave
1. Element for autoclave
2. Autoclave knob
3. Autoclave gasket
4. outer chamber
5. code wire
6. autoclave door

Thanks & Regards

Avinash T
District charge, kannur
bemp kerala
mobile: +91 6364064948
E mail: avinashtcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD
30/641B | Petta | Punithura | Kochi | Kerala- 682038
www.cyrixhealthcare.com

avinash t <avinashtcyrix@gmail.com>
Draft to: sreerajsurgicals@gmail.com, Sarang KM <sarangcyrix@gmail.com>

Fri, 13 Dec, 2024 at 10:54 am

[@Sarang KM](#)

On Fri, 15 Nov, 2024, 8:01 am avinash t, <avinashtcyrix@gmail.com> wrote:

Dear sir,
Please provide a quotation of below mentioned parts of autoclave
1. Element for autoclave
2. Autoclave knob
3. Autoclave gasket
4. outer chamber
5. code wire
6. autoclave door

Thanks & Regards

Avinash T
District charge, kannur
bemp kerala
mobile: +91 6364064948
E mail: avinashtcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD
30/641B | Petta | Punithura | Kochi | Kerala- 682038
www.cyrixhealthcare.com

sreerajsurgicals <sreerajsurgicals@gmail.com>
To: avinash t <avinashtcyrix@gmail.com>

Fri, 13 Dec, 2024 at 11:19 am

Sent from Mail for Windows

From: avinash t
Sent: 13 December 2024 10:25 AM
To: sreerajsurgicals@gmail.com; Sarang KM
[Quoted text hidden]

[Quoted text hidden]


STOCK REGISTER

Name of Article
वस्तु का नाम

Autoclave

Rs- 3299

Date तिथि	PARTICULARS विवरण	Bill No. बीजक नं०	Receipt प्राप्ति	Issue देय	Balance बाकी	Remarks टिप्पणी
1.1.04	BIF from Page 103 Vol 17 (Received from UNISKAR 295 Pk IN No 551 dt. 5.8.79)	SR Page (5)			Nil	W. Rs 3299/-
25/10/08	Received from Mrs. P. Anwar Surgical, Parvathur as per In. No. 2979 dt. 24/10/08 of -			1 No	1 No	NR fund
29/10/08	Issued to S/N - Received - Room	SR No. 5		1 No	ML	Rs 5100.00 L. No.
24/3/2017	BIF from Page 14 page Vol I page no. 32 (Rs. 6476.19/-)	V13116690 24/3/2017	1		1	2
	Issued to H/N			1	Nil	



MEDICAL OFFICER
FAMILY HEALTH CENTRE
CHERUKUNNU THARA-670 301

Medical Engineering & Services
Dharmapuri - 21 Ph: 0417-2420367

QMS CALIBRATION CENTRE

CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)

TOLL FREE NO: 1800 - 425-7669

BAR CODE - (8004-890615225) 1341629



Sm
MEDICAL OFFICER
FAMILY HEALTH CENTRE
CHERUKUNNU THARA-670 301

FAMILY HEALTH CENTRE
KANNUR DT.
Date.....
CHERUKUNNU THARA-670301