

PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

| | |
|--|---|
| Name of Hospital: THQH KOTHUPARAMBA | Name of District: KANNUR |
| Name of Equipment: Laboratory Microscope | Equipment ID & Barcode: # 147593- 1321512 |
| Make: LABOMED | Model: LX400 EFL |
| Serial Number: 157775 | Warranty details: No Warranty |
| Date of purchase/Year of manufacture/ Installation Date: In ticket master installation date is 03-02-2016,As per the stock book installation date is 29-01-2016 (Stock book attached) | Present status of the equipment: Fully damaged |
| Date of breakdown: 19-11-2024 Date of registration of complaint through email/Toll free): Toll free | *AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC |
| Action taken: Checked the equipment and found stage and condenser defective, objectives not clear. Need to replace these spares for further checking of equipment. | |


Recommendations for repair (required service details): No recommendations

| | | |
|--|--------------------------|--|
| Cost of spares (specify parts and cost): | CONDENSER - 2124/- | |
| | OBJECTIVE 100X - 10502/- | |
| | OBJECTIVE 40X - 8142/- | |
| | STAGE FULL SET - 7670/- | |
| | TOTAL - 28438/- | |


| | |
|---|----------------------|
| #Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 66.95% | Asset Value: 42471/- |
|---|----------------------|

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix service Report and Oem quotation attached

Reasons for recommending the equipment as BER: Checked the equipment and found stage and condenser defective, objectives not clear. Need to replace these spares for further checking of equipment. Enquired oem for quotation. In ticket master installation date is 03-02-2016,As per the stock book installation date is 29-01-2016 (Stock book attached) aged up to 8 years and 10 months. Repairing Cost is 66.95%. Both criteria for RBER met. As per tender clause 5.3.14.1 we recommending the equipment for condemnation.



 21/12/2024
 Prathyush T. Bhaskar
 Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM: checked and identified that stage and condenser of microscope defective and objectives not clear. Aged upto 8 years and above. its repairing cost is 66.95% as per quotation. The BER criteria met - so as per BER protocol recommended for BER.

| | |
|---|--|
| Recommended for BER (Yes/ No): yes |  21/12/2024 Signature of JC BM (NHM) |
| Date: 21 / 12 / 2024. | |


Date: _____

*Not mandatory



Seal

Signature of Superintendent / Medical Officer (I/c)



SUPERINTENDENT
GOVT. TALUK HEAD QUARTERS HOSPITAL
KUTHUPARAMBA, PIN - 670 622

Based on the period of life and value as per the BER guidelines. Attach Photograph



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 126642

CYRIX[®]

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

| | |
|---|--|
| Health Facility..... <u>THRIKUNJAPARAMBA.....</u> | Call Registration Date : ... <u>19/11/2024</u> |
| Address..... <u>KOOTHUPARAMBA.....</u> | Caller ID : <u>147593</u> |
| <u>KANNUR</u> | Date of Visit : <u>20/11/2024</u> |
| Ph : <u>9946557278</u> | Asset No : <u>1321512</u> |
| | EQPT Name : <u>Laboratory Microscope</u> |
| | Manufacture..... <u>Labomed</u> Model : <u>LX400SEL</u> |
| | S. No..... <u>1577775</u> Dept. <u>Blood storage Unit.</u> |

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified :Stage and condenser defective......

Action Taken :Checked the equipment and found stage and condenser defective. Objectives not clear. Need to replace these spares for further checking of equipment......

Completed Date : 20/11/2024 Time : 11:30 am Spare Required

Spare Replaced Requested

| Description | Qty. | Part Number | PR Number |
|-------------|------|-------------|-----------|
| 1. | | | |
| 2. | | | |
| 3. | | | |

| Cyrix Engineer | Date | Start Time | End Time |
|-------------------------------|-------------------|----------------|----------------|
| <u>Prathyush T. Bhaskaran</u> | <u>20/11/2024</u> | <u>10:50am</u> | <u>11:30am</u> |

Customer Remark Completed Pending

Service Engineer Name : Prathyush T. Bhaskaran
 Signature: [Signature]
 Date: 20/11/2024
 Contact Number: 9947028643

Customer Name : Rajeshwar
 Signature: [Signature]
 Date: 20/11/24
 Contact Number: 98557276
 Designation: Technical



GOVT. TALUK HEADQUARTERS HOSPITAL KOTTAYAM, PIN - 670 643

UNIVERSAL AGENCIES

ISO 9001 : 2008

Vision for excellence



PALAKKAL ANGADI, THRISSUR - 680 001, KERALA
 0487 - 2427038, 2427031, Mob : 93870 211122, 8943341569
 Email : uniage@gmail.com, uniage.sales@gmail.com

CYRIX HEALTHCARE (P) LTD
 30/641B
 PETTA
 PUNITHURA
 KOCHI
 KERALA- 682038

Our Quotation No: UA-Q. 417/2024

Dt. 20.12.2024

Ref: Service Estimate

Sir,

We hereby quote our lowest possible rates for the following items for your kind consideration and early favorable orders.

| NO | ITEM | UNIT RATE | GST | TOTAL |
|----|----------------------|-----------|-----|----------|
| 1. | STAGE FULL SET | 6900.00 | 18% | 7670.00 |
| | OBJECTIVE LENS 100X, | 8900.00 | 18% | 10502.00 |
| | OBJECTIVE LENS 40X) | 6900.00 | 18% | 8142.00 |
| | CONDENSER | 1800.00 | 18% | 2124.00 |

Terms and Conditions

Free delivery at your stores.

(Further shifting charges if any, shall be borne by the concerned department.)

Rates are inclusive of GST @ 18%.

(Note: Civil / Electrical / Water Inlet & Outlet works of any nature will be of your scope)

Rates will remain firm for acceptance till 31.03.2025

PAYMENT : 100% ADVANCE

Account Name: M/s. Universal Agencies

Our bank details:

STATE BANK OF INDIA

SME BRANCH PARAMAKKAVU THRISSUR

A/c No.00000067213868605

IFSCODE SBIN063880

CONTACT DETAILS:

Purchase/Inquiry: 8943112277 / 9387021122

Order status/Sales/Service:

8943341568/8943341569/0487-2427031

Accounts/Payment: 8943112288 /0487-2427031

Others: Mr. Prasannan (Manager) 9447137031

Yours Faithfully,
 For Universal Agencies,

(Signature)
 Manager.



SUPERINTENDENT
GOVT. TALUK HEADQUARTERS
HOSPITAL



Sarang KM <sarangcyrix@gmail.com>

**QUOTATION REQUIRED FOR MICROSCOPE #1321512,141939 @THQH
KOOOTHUPARAMBA**

8 messages

Sarang KM <sarangcyrix@gmail.com>
To: UNIVERSAL AGENCIES <uniage.sales@gmail.com>
Cc: Anjitha T L <anjithacyrix@gmail.com>, BLESSEN JOSE <zm2.klbemp@cyrix.in>

Fri, Dec 13, 2024 at 10:40 PM

Dear sir,
Please provide the stock availability and quotation of below mentioned part of
Microscope (make-LABOMED, model-LX 400 ELF)
1. STAGE FULL SET
2. OBJECTIVE LENS(100X, 40X)
3.CONDENSER

—
Thanks & Regards

Sarang.k.m
Sr.Biomedical Engineer
Bemp-Kerala
Mobile : +91 7593847150
Email : sarangcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD
30/641B | Petta | Punithura | Kochi | Kerala- 682038
www.cyrixhealthcare.com

Sarang KM <sarangcyrix@gmail.com>
To: UNIVERSAL AGENCIES <uniage.sales@gmail.com>
Cc: Anjitha T L <anjithacyrix@gmail.com>, BLESSEN JOSE <zm2.klbemp@cyrix.in>

Tue, Dec 17, 2024 at 12:02 PM

Dear sir,

REMINDER , Kindly share the quotation
[Quoted text hidden]

Sarang KM <sarangcyrix@gmail.com>
To: UNIVERSAL AGENCIES <uniage.sales@gmail.com>
Cc: Anjitha T L <anjithacyrix@gmail.com>, BLESSEN JOSE <zm2.klbemp@cyrix.in>

Wed, Dec 18, 2024 at 2:06 PM

Dear sir,

REMINDER , Kindly share the quotation
[Quoted text hidden]

Anjitha T L <purchasecyrix2@gmail.com>
To: Sarang KM <sarangcyrix@gmail.com>
Cc: Anjitha T L <anjithacyrix@gmail.com>, BLESSEN JOSE <zm2.klbemp@cyrix.in>, UNIVERSAL AGENCIES <uniage.sales@gmail.com>

Wed, Dec 18, 2024 at 4:42 PM

Dear Sarang ,

Please share image for below mentioned spare

Microscope (make-LABOMED, model-LX 400 ELF)

1. STAGE FULL SET

[Quoted text hidden]

—

Thanks & Regards

Anjitha T L

Sr. Purchase Executive

PH:- 9947719111

purchasecyrix2@gmail.com

www.cyrixhealthcare.com

Cyrix Healthcare Pvt Ltd

| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India

KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH

Sarang KM <sarangcyrix@gmail.com>

Thu, Dec 19, 2024 at 12:49 PM

To: Anjitha T L <purchasecyrix2@gmail.com>

Cc: Anjitha T L <anjithacyrix@gmail.com>, BLESSEN JOSE <zm2.klbemp@cyrix.in>, UNIVERSAL AGENCIES <uniage.sales@gmail.com>

Dear madam,

Kindly find the attached images

[Quoted text hidden]

3 attachments



WhatsApp Image 2024-12-18 at 4.56.08 PM.jpeg
676K



WhatsApp Image 2024-12-18 at 4.56.06 PM.jpeg
575K



WhatsApp Image 2024-12-18 at 4.56.07 PM.jpeg
550K

UNIVERSAL AGENCIES <uniage.sales@gmail.com>

Thu, Dec 19, 2024 at 5:26 PM

To: Sarang KM <sarangcyrix@gmail.com>

Cc: Anjitha T L <purchasecyrix2@gmail.com>, Anjitha T L <anjithacyrix@gmail.com>, BLESSEN JOSE <zm2.klbemp@cyrix.in>

Dear Sir

tomorrow will send the quotation

[Quoted text hidden]

-

Warm Regards,

M/S. Universal Agencies

(ISO 9001 : 2008 Certified)

H.O - "Palakkal Angadi, Thrissur -680001 Kerala - India"

Ph No: 0487-2440319/2427031

Cel : 8943341568

GST : 32AAAFU5662Q1ZQ

Email : uniage@gmail.com

uniage.sales@gmail.com

Branch Office: " 2nd Floor, Pride of Aristo,

C.S.I Cemetery Road, Mission Quarters

East Fort, Thrissur- 680005"

Tel: 0487-2427038

Sarang KM <sarangcyrix@gmail.com>

Fri, Dec 20, 2024 at 3:15 PM

To: UNIVERSAL AGENCIES <uniage.sales@gmail.com>

Cc: Anjitha T L <purchasecyrix2@gmail.com>, Anjitha T L <anjithacyrix@gmail.com>, BLESSEN JOSE <zm2.klbemp@cyrix.in>

Dear sir,

Kindly share the quotation

[Quoted text hidden]

UNIVERSAL AGENCIES <uniage.sales@gmail.com>

Fri, Dec 20, 2024 at 3:28 PM


To: Sarang KM <sarangcyrix@gmail.com>

Cc: Anjitha T L <purchasecyrix2@gmail.com>, Anjitha T L <anjithacyrix@gmail.com>, BLESSEN JOSE <zm2.klbemp@cyrix.in>

Dear Sir

Quotation attached

[Quoted text hidden]

 **CYRIX 400EFL.pdf**
162K

3

MICROSCOPE Binoocular (Lebomed)



B/p vol-2 Page 43

| Date | Particulars | Rpt | Issue | Balance | Sig |
|---------------------|--|-----|-------|---------|-----|
| 1-12-2011 | B/p for Page 43 of vol. 2 | - | - | 2 | ✓ |
| 29-01-16 | using for lab purpose since 6/4/15 (RNTCP) | 1 | | | ✓ |
| 29-01-16 | " " since 10/10/15 (General) | ! | | | ✓ |
| | Received from store Rs no 90 vol 1110! | | | 03. | |
| | Model - LX-400 Type-1 | | | | |
| | with LED illumination! | | | | |
| | Warranty period - 29/1/16 to 29/1/19 | | | | |
| 29/1/16 | Issued for Lab purpose. | | 01 | 02 | |
| R. N. S. S. S. | for Lab purpose | 01 | | | ✓ |

Serial No. 944704472



SUPERINTENDENT
GOVT. TALUK HEADQUARTERS
HOSPITAL
KUTHUPARAMBA, PIN - 670 647



TALUK HEADQUARTERS
GOVT. TALUK HEADQUARTERS
HOSPITAL
KUTUPARAMBA
KARNATAKA

SUPERINTENDENT
GOVT. TALUK HEADQUARTERS
HOSPITAL
KUTUPARAMBA, PIN - 670 643