

**PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: PHC EZHOME	Name of District: KANNUR
Name of Equipment: CENTRIFUGE	Equipment ID & Barcode: #152893- 1342580
Make: Rotek Instruments	Model: RMC 6
Serial Number: N/A	Warranty details: No Warranty
Date of purchase/Year of manufacture/ Installation Date: In ticket master installation date is 18-05-2018	Present status of the equipment: Fully damaged
Date of breakdown: 16-12-2024	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Date of registration of complaint through email/Toll free): Toll free	
Action taken: Checked the equipment and found that centrifuge motor not working, found problem with motor and armature. Need to replace these spares for further checking of the equipment. Enquired oem for quotation	
Recommendations for repair (required service details): No recommendations	
Cost of spares (specify parts and cost):	Centrifuge Motor - 3500/- Centrifuge Armature - 900/- GST 18% -792 TOTAL - 5192/-


#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 81.1%	Asset Value: 6400 /-
--	----------------------

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix service Report and Oem quotation attached

Reasons for recommending the equipment as BER: Checked the equipment and found that centrifuge motor not working, found problem with motor and armature. Need to replace these spares for further checking of the equipment. Enquired oem for quotation. In ticket master installation date is 18-05-2018 aged up to 6 years and 7 months. Repairing Cost is 81.1%. Both criteria for RBER met. As per tender clause 5.3.14.1 we recommending the equipment for condemnation.

Vishnu.p 
Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM: checked and identified that machine not working due to motor and armature defective. Aged up to 6 years & above. The quotation cost is 81.1%. So service cost economically not feasible. So recommended for BER as per BER Protocol

Recommended for BER (Yes/ No): Yes.	 NATIONAL HEALTH MISSION KANNUR-670002 Signature of JC BM (NHM)
Date: 04/01/25	

Date:	 മെഡിക്കൽ ഓഫീസർ പി.എച്ച്.സി. മൂന്നാം പിടില എഴുതാം, കണ്ണൂർ - 670 334 Signature of Superintendent/ Medical Officer (i/c)
-------	---

*Not mandatory # Based on the period of life and value as per the BER guidelines * Attach Photograph

Required Spare Quotation -1342580- 152893-Centrifuge-PHC EZHOME- KANNUR

5 messages

Sarang KM <sarangcyrix@gmail.com>
To: mail@rotekinstruments.com, info@rotekinstruments.com
Cc: BLESSEN JOSE <zm2.klbemp@cyrix.in>

Tue, Dec 17, 2024 at 1:25 PM

Dear Sir

Please share price and spare availability of centrifuge motor , armature

Centrifuge
Make : Rotek
Model: NA
SN: 17331

Here I am attaching photos for your reference

—
Thanks & Regards

Sarang.k.m
Sr.Biomedical Engineer
Bemp-Kerala
Mobile : +91 7593847150
Email : sarangcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD
30/641B | Petta | Punithura | Kochi | Kerala- 682038
www.cyrixhealthcare.com



1342580_asset_1654613160.jpg
327K

Sarang KM <sarangcyrix@gmail.com>
To: mail@rotekinstruments.com, info@rotekinstruments.com
Cc: BLESSEN JOSE <zm2.klbemp@cyrix.in>, Purchase Cyrix <purchasecyrix2@gmail.com>

Wed, Dec 18, 2024 at 11:59 PM

Dear sir,

REMINDER
[Quoted text hidden]

Sarang KM <sarangcyrix@gmail.com>
To: mail@rotekinstruments.com, info@rotekinstruments.com
Cc: BLESSEN JOSE <zm2.klbemp@cyrix.in>, Purchase Cyrix <purchasecyrix2@gmail.com>, abdulla2cyrix@gmail.com

Sat, Dec 21, 2024 at 1:17 PM

Dear sir,

REMINDER , Kindy share the quotation

[Quoted text hidden]

Rotek Instruments <mail@rotekinstruments.com>
To: Sarang KM <sarangcyrix@gmail.com>

Sat, Dec 21, 2024 at 1:43 PM

SIR,

As per your enquiry , pls find below the rate of spares :

Medico Armature	Rs. 900/-
Medico Motor	Rs.3500

	Rs.4400/-
GST@ 18%	Rs. 792/-

Total	Rs.5192/-
	=====

Thanks & Regards



ROTEK INSTRUMENTS

B&C Industries, West Vengola P.O.,Perumbavoor- 683556, Ernakulum Dist, Kerala,

PH: 8593070000/8593000606 / 9446139563 web: www.rotekinstruments.com



[Quoted text hidden]

Rotek Instruments <mail@rotekinstruments.com>
To: Sarang KM <sarangcyrix@gmail.com>

Sat, Dec 21, 2024 at 2:35 PM

Cc: info@rotekinstruments.com, BLESSEN JOSE <zm2.klbemp@cyrix.in>

Quotation

sir
pfa

Thanks & Regards




ROTEK INSTRUMENTS

B&C Industries, West Vengola P.O.,Perumbavoor- 683556, Ernakulum Dist, Kerala,

PH: 8593070000/8593000606 / 9446139563 web: www.rotekinstruments.com



On Tue, Dec 17, 2024 at 1:25 PM Sarang KM <sarangcyrix@gmail.com> wrote:
[Quoted text hidden]

 **Qtn- cyrix.pdf**
151K



8

മെഡിക്കൽ സാധനങ്ങൾ
പരിപാലനം, പരിഷ്കരണം - 879 834
8004-890615225

മെഡിക്കൽ സാധനങ്ങൾ
പരിപാലനം, പരിഷ്കരണം
8004-890615225