


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

Sl. No	Particulars	Details
1	Name of District	THRISSUR
2	Name of Hospital	PHC ELANAD
3	Name of Equipment with Make, Model and Serial Number	OXYGEN CONCENTRATOR Make: NA Model: NA SN: NA
4	Equipment ID/ Barcode	0845025 - #91810
5	Date of purchase Year of manufacture/Installation Date	Ticket master installation date - NA As per stock book installation date- 27-10-2019
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	Toll free 20-01-2024
9	Action taken	Checked the machine and found that compressor motor, sieve bed and main board are defective, need to replace these spares for further checking
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	NA



13	Asset Value	26500/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix Service report Attached
16	Reasons for recommending the equipment as BER	Upon inspecting the machine, it was discovered that the compressor motor, sieve bed, and main board are faulty. These spare parts need replacement to assess the equipment's operational status. As per hospital stock the equipment received on 20-01-2024. Unfortunately, the supplier information is unavailable from end user as the equipment was donated without manufacturing details. Due to the lack of spare part information, we advise condemning the equipment as a special case (5.3.14.2).
17	Name & Signature of CYRIX Authority	BIJO T JOY 

\*Not mandatory #Based on the period of life and value as per the BER guidelines


\* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

*Supplier information not available for getting spares*

  
  
 Signature of JC BM (NHM)

Date: \_\_\_\_\_



  
**MEDICAL OFFICER**  
 Superintendent  
**PRIMA HEALTH CENTRE**  
 Trissur  
 Elanad.



# Oxygen Concentrator

27/10/19 10684  
24/10/19

Recd from KCP Aluva 1

1

1600/-

28/10/19

Issued to SN

1

Nil

0067562

12/10/19

Pro. SWE SR-77

66



