

**PROFORMA  
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)  
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: GH THRISSUR	Name of District: Thrissur
Name of Equipment: Digital Weighing Machine	Equipment ID & Barcode: #150032- 0854769
Make: Sknol	Model: Sknol 7245SK
Serial Number:NA	Warranty details: No Warranty
Date of purchase/Year of manufacture/ Installation Date: In Ticket master installation date is- 28-04-2023 As per Stock book installation date is - 28-04-2020	Present status of the equipment: Fully damaged
Date of breakdown:02-12-2024 Date of registration of complaint through email/Toll free): Toll free	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC

Action taken: checked the machine and found display and sensor are defective. Need to replace these spares for further checking working condition of machine.


Recommendations for repair (required service details): Not Recommending

Cost of spares (specify parts and cost): NA

#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA	Asset Value: 1935/-
---	---------------------


Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix report attached.

Reasons for recommending the equipment as BER: Checked the equipment and found that the display and sensor are defective. As per stock book the equipment was installed on 28-04-2020, and is over 4 years and 7 months old with an asset cost of ₹1935. A quotation has not been submitted as the necessary spares are not available in the market. Therefore, we recommend the equipment for condemnation

  
**BIJO T JOY**  
 Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM:


*Spares all not available*

  
 Dr. Consultant (Biomedical)  
 NHM, Thrissur

Recommended for BER (Yes/ No): <i>Yes</i>	Signature of JC BM (NHM)
Date:	<i>27/12/24</i>

Date:	Seal	Signature of Superintendent, Medical Officer (i/c)
-------	------	--



\*Not mandatory      #Based on the period of life and value as per RBER guidelines      \*Attach Photograph

  
 GENERAL HOSPITAL  
 THRISSUR - PIN: 680 001











WEIGHING MACHINE

135

28/04/20


RES

LS/00075/20-21

085,769

1 - 20

Rs - 1935/-



*[Handwritten signature]*