





**PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: PHC EDAVILANGU	Name of District: Thrissur
Name of Equipment: B P Apparatus	Equipment ID & Barcode: #151424- 0844383
Make: Shenzhen Goldway Industrial Inc.	Model: BPM-103
Serial Number: NA	Warranty details: No Warranty
Date of purchase/Year of manufacture/Installation Date: Ticket master installation date is- NA As per Stock book installation date is - 09-01-2020	Present status of the equipment: Fully damaged
Date of breakdown: 09-12-2024 Date of registration of complaint through email/Toll free: Toll free	*AMC/CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: checked the machine and found display and motor are defective. Need to replace these spares for further checking working condition of equipment.	
Recommendations for repair (required service details): Not Recommending	
Cost of spares (specify parts and cost): NA	
#Percentage value of the cost of spares with respect to Cost of Purchase/Asset Value: NA	Asset Value: 2200/-
Abstract of Service Report provided by the OEM/Authorized Service Provider/CYRIX (Attached or Not): Cyrix report attached.	
Reasons for recommending the equipment as BER: Checked the equipment and found that the display and motor are defective. The equipment was installed on 09-01-2020 and is 3 years and 11 months old. A quotation has not been submitted as the necessary spares are not available in the market. Therefore, we recommend the equipment for condemnation..	
BIJO T JOY	 Name & Signature of CYRIX Authority with date
Remarks of Junior Consultant (Biomedical) NHM: <i>Space not available</i>	
Recommended for BER (Yes/No): ✓	 Signature of JCBM (NHM)
Date: 18/12/24	 Signature of Superintendent/Medical Officer (i/c)



*Not mandatory Based on the period of life and value as per the BER guidelines

*Attending Medical Officer in charge

Family Health Centre
Edavilangu



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 1020699

CYRIX[®]

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility..... <i>PHC Edavilangu</i> Address..... <i>Thrissur</i> Ph:..... <i>9447833869</i>	Call Registration Date :..... <i>9/12/2024</i> Caller ID :..... <i>151424</i> Date of Visit :..... <i>10/12/2024</i> Asset No :..... <i>0844383</i> EQPT Name :..... <i>BP Apparatus</i> Manufacture..... <i>shenzhen</i> Model : <i>BPM-103</i> S. No..... <i>NA</i> Dept..... <i>observation room</i>
--	---

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : ..*Equipment not working*.....
.....
.....

Action Taken : ..*checked the machine and found display and motor are defective need to replace these spares for further checking working condition of equipment*.....
.....
.....

Completed Date : ..*18/12/2024*.. Time : ..*12:30PM*.. Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<i>Ramdas</i>	<i>18/12/2024</i>	<i>12:00 PM</i>	<i>12:30PM</i>

Customer Remark Completed Pending

Service Engineer Name : <i>Ramdas</i> Signature: <i>[Signature]</i> Date: <i>18/12/2024</i> Contact Number: <i>9745520163</i>	 Customer Name : <i>Dr Dewi L</i> Signature: <i>[Signature]</i> Date: <i>18/12/24</i> Contact Number: <i>9447299500</i> Designation : <i>Medical Officer in charge</i> Hospital Seal: <i>Family Health Centre Edavilangu</i>
--	--

DIGITAL BP APPARATUS (Common)

2019-2024

		Received	Issue	Balance
9/1/20	Received from the store (Common) ^{Digital BP Apparatus & adaptor}	1		1
10/1/20	Issued to OP ⁰⁹⁴⁴³⁸³		1	
17/8/2019	(Late entry) Received from the store Memory free LED Display Model 1015 (Blue)	2		2+1
	Issued to OP		2	
14/5/2020	NCD - From (store) MicroLife BP apparatus	1		1+3
15	Issued to OP		1	(N)
24/2/23	Received from store [OXY-MED] WITH ADAPTOR BP DVO1 / BP 136A	2		4+2
25/2/23	Issued to OP		2	(N)
26/10/23	Received from the store - E adapts	1		7
26/10/23	Received from the store - E adapts	1		7
27/10/23	Issued to OP		1	(B)
22/3/24	Received from the store - oxy med - (LP) Used in pre-check OP.	2		9
			2	
02/07/24	Transfer out to unserviceable articles Register Page No. 6. Balance		2	7



