

PROFORMA

RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)

BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

| | |
|--|---|
| Name of Hospital: UHC Poothikkad | Name of District: Wayanad |
| Name of Equipment: Digital BP Apparatus | Equipment ID & Barcode: #147993-1250619 |
| Make: Beurer | Model: BM27 |
| Serial Number: 651.86 | Warranty details: No warranty |
| Date of purchase/Year of manufacture/ Installation Date: In ticket Master installation date is 23-06-2023 | Present status of the equipment: Fully damaged |
| Date of breakdown: 20-11-2024 Date of registration of complaint through email/ Toll free: Toll free | *AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC |
| Action Taken: Checked the machine and found that Machine is not switching on, found main board and display are defective. Need to replace the main board and display for further checking. | |
| Recommendations for repair (required service details): No recommendation for repair | |
| Cost of spares (specify parts and cost): N/A | |
| % Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: N/A | Asset Value: 1700/- |
| Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix report Attached | |
| Reasons for recommending the equipment as BER: Checked the machine and found that Machine is not switching on, found main board and display are defective. In ticket master installation date is 23-06-2023 and aged up to 1 years 5 months. Quotation not submitted since spares are not available from the market. So we recommend the equipment for condemnation. | |
| Name & Signature of CYRIX Authority with date: NiyasHadi A P 04/12/2024 | |
| Remarks of Junior Consultant (Biomedical) NHM: Main board + display board etc are complaint. Recommended condemnation | |
| Recommended for BER (Yes/ No): Yes | Signature of JC BM (NHM) |
| Date: 05/12/24 | |
| Date: | Seal Signature of Superintendent Medical Officer (JC) |

*Not mandatory

#Based on the period of life and value as per the B&K guidelines

*Affiliation: 673 592



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 125402

CYRIX[®]

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

| | |
|--|---|
| Health Facility..... UHWC..... Address..... UHWC Poothikad..... Wayanad..... kerala..... Ph: 8.138861349..... | Call Registration Date : 20/11/2024..... Caller ID : 147983..... Date of Visit : 22/11/2024..... Asset No : 1250619..... EQPT Name : Digital BP Apparatus..... Manufacture: Beurex..... Model : BM 27..... S. No..... NA..... Dept..... OP..... |
|--|---|

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : Not getting ON.....

Action Taken : Checked the equipment and found that the machine is not getting ON. Found the main board and display are defective. need to replace the main board and display for further checking.

Completed Date : 22/11/2024.. Time : 2:00 PM..... Spare Required

Spare Replaced Requested

| Description | Qty. | Part Number | PR Number |
|-------------|------|-------------|-----------|
| 1. | | | |
| 2. | | | |
| 3. | | | |

| Cyrix Engineer | Date | Start Time | End Time |
|----------------|------------|------------|----------|
| Neeraj D.R | 22/11/2024 | 1:00 PM | 2:00 PM |

Customer Remark Completed Pending

Service Engineer Name : Neeraj D.R
Signature:
Date: 22/11/2024
Contact Number: 8086938072

Customer Name :
Signature:
Date: 22/11/2024
Contact Number: 673592
Designation:
Hospital Name:

DIGITAL BP APPARATUS

| | | | |
|-----------|-------------------------------|---|--------------------------------|
| 1805 | Received from Kerala Surgical | | |
| 23-Jun-23 | Equipments 6. | 1 | 1700/- |
| 3-7-23 | Issued to nurse | 1 | Received <i>[Signature]</i> |


For the Dr Deekshith

[Signature]



മെഡിക്കൽ ഓഫീസർ
 ആർബൻ ഹെൽത്ത് &
 വെൽഫെയർ ഡിവിഷൻ
 പുതിയ - 673 592



For the Dr. Deekshith


മെഡിക്കൽ ഓഫീസർ
 അർബൻ ഹെൽത്ത് &
 വെൽനെസ്സ് സെന്റർ
 പുതിക്കാട് - 673 592

