

**PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: UPHC OTTAPALAM	Name of District: PALAKKAD
Name of Equipment: CENTRIFUGE	Equipment ID & Barcode: #149227-0943361
Make: REMI	Model: C854/8
Serial Number: HHLC-12635	Warranty details: No Warranty
Date of purchase/Year of manufacture/ Installation Date: In ticket master installation date 30/06/2015	Present status of the equipment: Fully damaged
Date of breakdown: 27/11/2024 Date of registration of complaint through email/Toll free: Toll free	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: Checked the centrifuge found centrifuge not get rotating and also sample cup broken. Problem with electric motor, speed regulator knob, brass ring and plastic pp tube damaged. need to be replace these spares for further checking working condition of machine.	

Recommendations for repair (required service details): No recommendations

Cost of spares (specify parts and cost):
 Electric motor = 3313/-
 Knob for speed regulator = 531/-
 Plastic PP tube = 566/-
 Brass ring 50ml = 2454/-
Total = 6865/-

#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 71.13%	Asset Value: 9651/-
---	---------------------

Abstract of Service Report provided by the OEM/Authorized Service Provider/CYRIX (Attached or Not): Cyrix Service report attached and quotation attached.


Reasons for recommending the equipment as BER: Checked the centrifuge found centrifuge not get rotating and also sample cup broken. Problem with electric motor, speed regulator knob, brass ring and plastic pp tube. need to be replace these spares for further checking working condition of machine. Equipment installed on 30/06/2015 aged up to 9 years 4 months. Repairing cost is 71.13% both criteria for RBER met. As per tender clause 5.3.14.1 we recommending for condemnation.

AKHILESH P
 Name & Signature of CYRIX Authority with date *06/12/24*

Remarks of Junior Consultant (Biomedical) NHM:
Machine installed on 30/06/2015 and aged up to 9 years 4 months. Repairing cost is 71.13%. Both criteria RBER met.

Recommended for BER (Yes/ No): <i>Yes</i>	Signature of JCBM (NHM) <i>Ash</i>
Date: <i>06/12/24</i>	Consultant (Biomedical) NHM (Arogyakeralam), Palakkad

Date: _____
 Signature of Superintendent/Medical Officer (i/c) *06/12/24*



MEDICAL OFFICER
 URBAN PRIMARY HEALTH CENTRE,
 SOUTH PANAMANNA, OTTAPALAM

*Not mandatory #Based on the period of life and value as per the tender document * Attach Photograph



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 126217

CYRIX[®]

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility... <u>UPHC</u> Address... <u>Ottapalam</u> <u>Palakkad</u> Ph: <u>7034683055</u>	Call Registration Date : <u>27/11/2024</u> Caller ID : <u>149277</u> Date of Visit : <u>28/11/2024</u> Asset No : <u>0943361</u> EQPT Name : <u>Centrifuge</u> Manufacture... <u>REMI</u> Model : <u>C-854/8</u> S. No... <u>nil</u> Dept. <u>LAB</u>
---	---

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : Not working

Action Taken : Checked the centrifuge, found centrifuge not get rotating and also sample cup broken, problem with electric motor, speed regulator knob, brass ring and plastic pp tube damaged. Need to be replace these spares for further checking working condition of machine.

Completed Date : 28/11/2024 Time : 11:00 AM Spare Required

Spare Replaced Requested


Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Shivakrishnan.k</u>	<u>28/11/2024</u>	<u>10:00 AM</u>	<u>11:00 AM</u>

Customer Remark Completed

MEDICAL OFFICER
URBAN PRIMARY HEALTH CENTRE
SOUTH PANAMANNA, OTTAPALAM

Service Engineer Name : Shivakrishnan Customer Name : Babitha . E
 Signature: [Signature] Signature: [Signature]
 Date: 28/11/2024 Date: 28/11/24
 Contact Number: 623875513 Contact Number: 7034683055
 Designation: Lab technician
 Hospital Seal



REMI SALES & ENGINEERING LIMITED



48/1103, Puthyedath Road, Vytilla P.O, Ernakulam Kochi-682019
 Phone : 0484-682019, Mob: 8921409347 Email Id.: inservicekochi@remigroup.com/service@remilabworld.com

Corporate Off : Remi House, 3rd floor, 11 Cama Inld. Estate, Goregaon (East) Mumbai 400 063. INDIA
 • Board: +91 22 4058 9888 SALES SUPPORT NO.8679448682 • Email Id.: sales@remilabworld.com

QUOTATION CUM PROFORMA INVOICE

QN- DEC-02/24-25

Complaint No.		Customer Name & Address:		Quotation No	PRINTED
Dated	4/12/2024	CYRIX HEALTHCARE (P) LTD		Date	4/12/2024
ENGINEER NAME	vipin	30/641B, Petta, Puthura Ernakulam- 682038		REF	Phone
REF.	Phone	7593847154		DATE	4/12/2024
Email	akhilshcyrix@gmail.com				

Model no.	REMI :CBS4/8	SR NO. /MFG DT.	SLNO : HHLC-12635
-----------	--------------	-----------------	-------------------

Sr No.	Description	Part code	HSN/SAC	Rate Rs.	Qty	Amount Rs. (A)	GST % (B)	GST Amt (C)	Total Amt (A+C)
	ELECTRIC MOTOR (RT)	L22000718	848190	2808	1	2808	18%	505	3313
	IGNOB FOR SPEED REGULATOR (HRT)	L22000002-	842119	450	1	450	18%	81	531
	PLASTIC P P TUBE 15 ML(HRT)	L22000333	901890	60	8	480	18%	86	566
	BRASS RING 50 ML(HRT)		842119	260	8	2080	18%	374	2454
								TOTAL (D)	6865
	GAS CHARGEING	0	0	0	0	0	18%	0	0
	SERVICE CHARGE	0	0	0	0	0	18%	0	0
								TOTAL (E)	6865

Rupees in Words: **Rs: Six thousand eight hundred sixty five only** 6865

Terms & Conditions:

- Spares are suitable only to use with REMI instruments & not for any other make.
- We will accept return of spares only if they are unused and not damage. At our discretion, the returned spare will be replaced. The difference in prices, if any will have to be borne by you.
- To ensure supply of correct spares, we suggest you to send us defective spare in advance with your request for which you want to purchase new spares.
- 100 % Advance along with your order confirmation
- Payment can be made by Cheque/ NEFT in our Bank account. For Cash Payment you must collect Receipt from recipient of cash failing which REMI will not be held responsible for any discrepancy.
- Courier/ transportation charges are payable at destination & not included in our price.
- Validity: The quotation is valid for 30 days from the date of this quotation / Proforma Invoice
- Subject to Ernakulam Jurisdiction only

REMARKS : PAYMENT 100% ADVANCE

For REMI SALES & ENGINEERING LTD



BANK DETAILS FOR REMI SALES & ENGINEERING LTD.

- | | | |
|---------------------------|----------------------------------|-------------------------------|
| • A/C NO: 35699154671 | • Bank Name: STATE BANK OF INDIA | • BRANCH: ELAMAKKARA |
| • IFS CODE: SBIN0013223 | • MICR CODE:682002053 | • SWIFT CODE: SBININBB395 |
| • GST NO: 32AAACR0421M125 | • PAN NO: AAACR0421M | • CIN NO. U31100MH198PL022314 |



**REQUIRED QUOTATION # 0943361 - 149277 CENTRIFUGE MACHINE @ UPHC OTTAPALAM ,
PALAKKAD**

5 messages

Akhilesh P <akhileshcyrix@gmail.com>

To: REMI Group - Kochi - Vishnu S K <kochisales.hcd@remigroup.com>, instservicekochi <Instservicekochi@remigroup.com>, kochi@remigroup.com

Thu, 28 Nov, 2024 at 11:13 am

Cc: DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Blesson Jose <Zm2.klbemp@cyrix.in>

Dear sir,

Please share the price and stock availability of

1. ELETRIC MOTOR
2. KNOB FOR SPEED REGULATOR
3. TUBE SS 15 ML-8 NOS
4. BRASS RING 50 ML - 8 NOS

Deatils are furnished below

Hospital Name : UPHC OTTAPALAM

Department : LABORATORY

Equipment Name : CENTRIFUGE MACHINE

Make : REMI

Model : C854/8

SN : HHLC-12635

PFA

Thanks & Regards

AKHILESH P
DISTRICT INCHARGE - PALAKKAD
Bemp-Kerala
Mobile : +91 7593847154
Email : akhileshcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD
30/641B, Petta, Punithura
Ernakulam- 682038
www.cyrixhealthcare.com

Akhilesh P <akhileshcyrix@gmail.com>

To: REMI Group - Kochi - Vishnu S K <kochisales.hcd@remigroup.com>, instservicekochi <Instservicekochi@remigroup.com>, kochi@remigroup.com

Mon, 2 Dec, 2024 at 12:45 pm

Cc: DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Blesson Jose <Zm2.klbemp@cyrix.in>

Dear sir

Please update.

Thanks & Regards

AKHILESH P
DISTRICT INCHARGE - PALAKKAD
Bemp-Kerala
Mobile : +91 7593847154
Email : akhileshcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD
30/641B, Petta, Punithura
Ernakulam- 682038
www.cyrixhealthcare.com

[Quoted text hidden]

Akhilesh P <akhileshcyrix@gmail.com>

To: REMI Group - Kochi - Vishnu S K <kochisales.hcd@remigroup.com>, instservicekochi <Instservicekochi@remigroup.com>, Wed, 4 Dec, 2024 at 9:45 am
kochi@remigroup.com

Cc: DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Blesson Jose <Zm2.klbemp@cyrix.in>

Dear sir

Please update.

Thanks & Regards

AKHILESH P
DISTRICT INCHARGE - PALAKKAD
Bemp-Kerala
Mobile : +91 7593847154
Email : akhileshcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD
30/641B, Petta, Punithura
Ernakulam- 682038
www.cyrixhealthcare.com

[Quoted text hidden]

Akhilesh P <akhileshcyrix@gmail.com>

To: REMI Group - Kochi - Vishnu S K <kochisales.hcd@remigroup.com>, instservicekochi <Instservicekochi@remigroup.com>, Thu, 5 Dec, 2024 at 9:37 am
kochi@remigroup.com

Cc: DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Blesson Jose <Zm2.klbemp@cyrix.in>

Dear sir

Please update

Thanks & Regards

AKHILESH P
DISTRICT INCHARGE - PALAKKAD
Bemp-Kerala
Mobile : +91 7593847154
Email : akhileshcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD
30/641B, Petta, Punithura
Ernakulam- 682038
www.cyrixhealthcare.com

[Quoted text hidden]

instservicekochi <instservicekochi@remigroup.com>

To: akhileshcyrix@gmail.com

Cc: REMI Group - Kochi - Vishnu S K <kochisales.hcd@remigroup.com>

Thu, 5 Dec, 2024 at 10:01 am

Dear Sir,

Please find the attached PROFORMA INVOICE for your reference. **PLEASE RELEASE THE PAYMENT IN ADVANCE.**

Regards

REMI KOCHI

INST DIV.

[Quoted text hidden]

----- Forwarded message -----

From: instservicekochi <instservicekochi@remigroup.com>

To: <akhileshcyrix@gmail.com>

Cc: "REMI Group - Kochi - Vishnu S K" <kochisales.hcd@remigroup.com>, <kochi@remigroup.com>

Bcc:

Date: Wed, 4 Dec 2024 17:32:47 +0530

Subject: QUTATION CUM PERFORMA INVOICE FOR CYRIX HEALTH CARE LTD DT 4-12-2024


Dear Sir,

Please find the attached PROFORMA INVOICE for your reference. **PLEASE RELEASE THE PAYMENT IN ADVANCE.**

Regards

REMI KOCHI

INST DIV.



[Signature]
MEDICAL OFFICER
URBAN PRIMARY HEALTH CENTRE
SOUTH PANAMARNA, OTTAPALAM