





**PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: TRIBAL SPECIALITY HOSPITAL KOTTATHRA	Name of District: PALAKKAD
Name of Equipment: BIOCHEMISTRY ANALYZER	Equipment ID & Barcode: #149993- 0913283
Make: ERBA	Model: EM 360
Serial Number: 60452	Warranty details: No Warranty
Date of purchase/Year of manufacture/ Installation Date: In ticket master installation date - NIL Actual installation date as per stock book 24/04/2014 stock book attached for reference.	Present status of the equipment: Fully damaged
Date of breakdown: 01/12/2024 Date of registration of complaint through email/Tollfree: Tollfree	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: Checked the machine found result variations. Problem with RGT cooling assembly, assembly of photometer with packing, total internal tubings, RIPT assy, Big KNF pump, cuvette ring set and cuvette damaged. need to be replaced these spares for further checking working condition of equipment.	
Recommendations for repair (required service details): No recommendations	
Cost of spares (specify parts and cost): Cuvvet 60 nos = 324300/- Total internal tubing kit with connector = 75670/- Assembly of photometer with packing = 324875/- RGT cooling assy with 6mm plate and spacer = 95335/- RIPT assy = 72450/- Cuvvet ring set assembly = 52210/- Big knf pump assy = 52210/- GST 18% = 179034.3/- Total = 1173669.3/-	
#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 108.4%	Asset Value: 1082525/-
Abstract of Service Report provided by the OEM/Authorized Service Provider/CYRIX (Attached or Not): Cyrix Service report attached and quotation attached.	
Reasons for recommending the equipment as BER: Checked the machine found result variations. Problem with RGT cooling assembly, assembly of photometer with packing, total internal tubings, RIPT assy, Big KNF pump, cuvette ring set and cuvette damaged. need to be replaced these spares for further checking working condition of machine. In ticket master installation date is NIL As per stock book Equipment installed on 24/04/2014 aged up to 10 years 7 months. Repairing cost is 108.4% both criteria for RBER met. As per tender clause 5.3.14.1 we recommending for condemnation.	
AKHILESH P  Name & Signature of CYRIX Authority with date	
Remarks of Junior Consultant (Biomedical) NHM: Biochemistry analyser installed on 24/04/2014 and aged up to 10 years and 7 months. Quotation submitted by OEM. Repairing cost is 108.4% of asset value. Both criteria of RBER met.	
Recommended for BER (Yes/ No): Yes.	Consultant (Biomedical) NHM Palakkad 21/12/24 
Date: 20/12/24	
Date: 21/12/24	Superintendent Tribal Speciality Hospital Kottathara Palakkad 21/12/24 





BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 128472

CYRIX[®]

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

<p>Health Facility..... <u>UTSH Kottathara</u></p> <p>Address..... <u>Palakkad</u> <u>Kerala</u></p> <p>Ph:..... <u>08590497318</u></p>	<p>Call Registration Date : <u>01-12-2024</u></p> <p>Caller ID :..... <u>149993</u></p> <p>Date of Visit : <u>3-12-2024</u></p> <p>Asset No :..... <u>0913283</u></p> <p>EQPT Name :..... <u>Biochemistry analyzer</u></p> <p>Manufacture..... <u>Erba</u> Model : <u>EM360</u></p> <p>S. No..... <u>60452</u> Dept. <u>Laboratory</u></p>
---	--

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : unit not working

Action Taken : checked the machine Found result variations problem with RWT cooling assembly, assembly of photometer with packing, total internal tubings, RIPT assembly, Big ICMF Pump, current ring set and current damaged need to be replace these spare for further

Completed Date : 03-12-2024 Time : 11:30 AM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>ABIJITH-B</u>	<u>3-12-2024</u>	<u>9:05 AM</u>	<u>11:30 AM</u>

Customer Remark Completed Pending

Service Engineer Name: ABIJITH-B

Signature:

Date: 03-12-2024

Contact Number: 9567164651


Customer Name : Sujas R.

Signature:

Date: 3/12/24

Contact Number: 9447022719

Designation : Senior Lab Technician



Superintendent
Govt. Tribal Health Services
Kottathara, Poonithura, Alappuzha

QUOTATION

Bill To: CYRIX HEALTHCARE PVT LTD
 First Floor,petta Junction 30/641b, Poonithura,Ernakulam, Kerala,
 Pincode 682038 , Palakkad district , Kerala 628581


Ref. No : 2425QUOTE002491
 Date : Dec 17, 2024 11:44 AM
 Validity : 30 days
 TBM GSTIN : 27AAACTP2038C1ZT

Customer GSTIN No : 32AAFCC2499H2ZM

Ship to, Government Tribal Speciality Hospital,Attappadi
 Kottathara P.O. Palakkad Dist , Palakkad district , Kerala 628581

SUB : QUOTATION FOR SUPPLY/REPLACEMENT OF SPARES IN INSTRUMENT MODEL: EM360 SR.NO.: 60452

With reference to the above, following Spares / Consumables are required to be replaced / supplied. Price/s of the same Is / are given below, for approval:

SR. NO	Part Code	HSN Code	Description of Spares/Consumables	Qty	Unit Price	Discount%	GST%	Amount	
1	100279	70171000	CUVETTES	60	5405		18	324300	
2	104951	90278990	TOTAL INTERNAL TUBING KIT WITH CONNECTOR (XL600/300/V2/A4/EM360)	1	75670		18	75670	
3	103093	90278990	ASSEMBLY OF PHOTOMETER WITH PACKING	1	324875		18	324875	
4	114595	90278990	RGT COOLING ASSY WITH 6MM PLATE & SPACER_XLV2	1	95335		18	95335	
5	105281	90278990	R1PT ASSY(SPARE) FOR XL-600/300-	1	72450		18	72450	
6	182530	90278990	Cuvette Ring Set Assembly (WØ Cuvettes)	1	49795		18	49795	
7	182496	90278990	Big Knf Pump Assy.	1	52210		18	52210	
Merchant Name :TRANSASIA BIO-MEDICALS LTD								Total (Rs)	994635
SCAN AND PAY WITH UPI QR Merchant Name : TRANSASIA BIO MEDICALS LTD  Payee VPA : transasiabiomedicals@citibank								Discount (Rs.) (-)	
								GST (Rs.)	179034.3
								Rupees in words: Eleven lakhs Seventy Three thousand Six hundred Sixty Nine rupees and Three paise only	Grand Total (Rs.)

TERMS AND CONDITIONS :

1. **Payment** : Full payment should be made in advance through NEFT / RTGS/ Demand Draft in favour of "Transasia Bio-Medicals Ltd" to Bank A/C no.: 00600330002538 IFSC code: HDFC0000060, UPI ID : transasiabiomedicals@citibank. You can also pay us through UPI apps like BHIM/ Google Pay/ Paytm/PhonePe etc. by scanning the above mentioned code. Cash Payment is strictly prohibited. No employee of the TBM is authorized to collect cash and or issue receipt against collection of Cash.
2. **GST** : As applicable
3. **Other Cost / Local duties** : Customer's responsibility
4. **Delivery** : Within 1-2 weeks from the receipt of your work order with payment confirmation

Thanking You

RECEIVED & ACCEPTED

[Signature]
 Superintendent
 Govt. Tribal Speciality Hospital,
 Kottathara (P.O.) Attappady

Your faithfully,
For TRANSASIA BIO-MEDICALS LTD

TRANSASIA BIO-MEDICALS LTD.

Head Office: Transasia House, B. Chandivali Studio Road, Andheri (E), Mumbai - 400 072. Tel.: (022) 4030 9000 Fax: (022) 2857 3030 Mumbai : Toll Free No. 1800 103 8226,
Whatsapp No.: 8451048434, SMS No.: 9212433444

www.transasia.co.in

DOC NO : SC00-405 / ISS - 7

CUSTOMER SIGNATURE & STAMP:

CUSTOMER NAME: _____

CONTACT NO: _____

ISO 13485 & ISO 9001 Certified Company

Erba
LABORATORY



REQUIRED QUOTATION FOR FULLY AUTOBIOCHEMISTRY ANALYZER 0913283 # 149993 @ TRIBAL SPECIALITY HOSPITAL KOTTATHARA , PALAKKAD

6 messages

Akhilesh P <akhileshcyrix@gmail.com>

Fri, 6 Dec, 2024 at 9:04 am

To: SANDEEP DAS M <s.das@transasia.co.in>

Cc: Blesson Jose <Zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Dear Sir,

Please visit TRIBAL SPECIALITY HOSPITAL KOTTATHARA , PALAKKAD and share the price and stock availability of the required spare parts for the EM360 fully automatic analyzer."

Deatils are mentioned below

Hospital Name : TRIBAL SPECIALITY HOSPITAL KOTTATHARA , PALAKKAD

Equipment Name : FULLY AUTOMATIC BIOCHEMISTRY ANALYZER

Make : TRANSASIA

Model : EM360

SN : 60452

Thanks & Regards

AKHILESH P
DISTRICT INCHARGE - PALAKKAD
Bemp-Kerala
Mobile : +91 7593847154
Email : akhileshcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD
30/641B, Petta, Punithura
Ernakulam- 682038
www.cyrixhealthcare.com

Akhilesh P <akhileshcyrix@gmail.com>

Mon, 9 Dec, 2024 at 11:36 pm

To: SANDEEP DAS M <s.das@transasia.co.in>

Cc: Blesson Jose <Zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Dear sir

Please update.

Thanks & Regards

AKHILESH P
DISTRICT INCHARGE - PALAKKAD
Bemp-Kerala
Mobile : +91 7593847154
Email : akhileshcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD
30/641B, Petta, Punithura
Ernakulam- 682038
www.cyrixhealthcare.com

[Quoted text hidden]

Akhilesh P <akhileshcyrix@gmail.com>

Fri, 13 Dec, 2024 at 1:53 pm

To: SANDEEP DAS M <s.das@transasia.co.in>
Cc: Blesson Jose <Zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Dear sir,
Please update the status.

[Quoted text hidden]

[Quoted text hidden]

Akhilesh P <akhileshcyrix@gmail.com>
To: SANDEEP DAS M <s.das@transasia.co.in>
Cc: Blesson Jose <Zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Mon, 16 Dec, 2024 at 1:00 pm

Dear sir
Please update the status.

Thanks & Regards

AKHILESH P
DISTRICT INCHARGE - PALAKKAD
Bemp-Kerala
Mobile : +91 7593847154
Email : akhileshcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD
30/641B, Petta, Punithura
Ernakulam- 682038
www.cyrixhealthcare.com

[Quoted text hidden]

<postmaster@cyrix.in>
To: akhileshcyrix@gmail.com

Mon, 16 Dec, 2024 at 1:02 pm

Your email couldn't be forwarded from dm1.klbemp@cyrix.in to another email address.

DM Sarang-KLBEMP (dm1.klbemp@cyrix.in)

Your message wasn't delivered because the recipient's email provider rejected it.

Diagnostic information for administrators:

Generating server: PN0P287MB1953.INDP287.PROD.OUTLOOK.COM

dm1.klbemp@cyrix.in

Remote server returned '550 5.7.520 Access denied, Your organization does not allow external forwarding. Please contact your administrator for further assistance. AS(7555)'

Original message headers:

Resent-From: <dm1.klbemp@cyrix.in>
Received: from PN3PR01CA0144.INDP287.PROD.OUTLOOK.COM (2603:1096:c01:bf::11)
by PN0P287MB1953.INDP287.PROD.OUTLOOK.COM (2603:1096:c01:1b8::13) with
Microsoft SMTP Server (version=TLS1_2,
cipher=TLS_ECDHE_RSA_WITH_AES_256_GCM_SHA384) id 15.20.8251.21; Mon, 16 Dec
2024 07:30:41 +0000
Received: from PN1PEPF000062BA.INDP287.PROD.OUTLOOK.COM
(2603:1096:c01:bf:cafe::5b) by PN3PR01CA0144.outlook.office365.com
(2603:1096:c01:bf::11) with Microsoft SMTP Server (version=TLS1_3,
cipher=TLS_AES_256_GCM_SHA384) id 15.20.8251.21 via Frontend Transport; Mon,
16 Dec 2024 07:30:41 +0000
Authentication-Results: spf=pass (sender IP is 2607:f8b0:4864:20::741)

=?us-ascii?Q?+7ggX7GDnR19iewr3nClf3FT7N2zY.f\JfeguZVxKMISdy4DqHR1pgbkI/Y4XG?=
=?us-ascii?Q?9AmEPyn46Xk18GCSjuzJrZMvnG0ZapxbUxXCf6N0YP6BZx3vYmvtugModGQr?=
=?us-ascii?Q?ndlBms045vcHNloIDFR1SINF/fB2+9E9vi fWdQ4WrpVkJZaWmLz8jjS0RoU0F?=
=?us-ascii?Q?ZoB9YAq4Dhpc5tShgnGc3vnCqEmrehRDEm1EvcD1vVh00A9QEDz4/shh43JU?=
=?us-ascii?Q?iFxoKYKexHuIqVNv99Y1A1jRTKZbLe43JB3Ldnbnq6Tde/gFyi39jeootVB?=
=?us-ascii?Q?0veUTXJo5Vxd6vQ1b+tiuJ7K0Yz8aLxP5VFrhZ8y9PsRLTpVJCNVYwi3C9Up?=
=?us-ascii?Q?In0aurp5dbiHhXYzysWziJyUiwQ1NfQA0QRMBq82NBc0tnpxL6714d5iqfE?=
=?us-ascii?Q?xRrBAnXbsLnwXC112JMaJeJ5/RbT9uMQ8bM+xCIzYa7tmYazZjvCPeDkxpuc?=
=?us-ascii?Q?Hm/nh57Ewe8MF1eLlftzGNSCscVUEY8sUXKLAtHJUJESogJUnwhn1bD3M90?=
=?us-ascii?Q?1lR5TXn7j0M+ad3uiQ7ews+VP5HFTTXwRCr1zsIU37edHjckTDeYVuZPkSCH?=
=?us-ascii?Q?SgQH/dm4QmuoguGPvGf7iNrKQ50WLV6QiKwm2SCQEbFUwDyWF96yytFaeAxU?=
=?us-ascii?Q?tDRT9Ef+XnrXe209Hkj+/P05BwkDMYso6xLrw+/gjk04s/d5RrK7GwJJamkvd?=
=?us-ascii?Q?g9I0U623meFXsuA10N/wUAdRfP0k3BU/hcm09urUYOwnNGTc0M5S1tXVcMhL?=
=?us-ascii?Q?K1Iyr0TSfk9XN8Q//u//vuhk3VtYNfPJwYJBEd55pFW3PyPPkJwU1iRLmGy?=
=?us-ascii?Q?mIpzGV0L4PBnPD389LWI5rnQNJwL2Wit2sJyYk7GVbpmucVFNe3GvGM40hCB?=
=?us-ascii?Q?P2BH5SIqI/ibRK0W+42VnN5uYA4o+T5K2ozZA2fp/inQhSug/HcaSrdR76/v?=
=?us-ascii?Q?NVH/dqMWA03zBbHwYRRCjzTlqHH3Hqvo7YCj7+AtEZtapE7117/isjfDEGmV?=
=?us-ascii?Q?0yQHw6qbomde0Vn0MJ/yi7dF7A8Q3nfIgnNvMQjURh3BfBTAr4ehJFc/Hh2u?=
=?us-ascii?Q?oG1ppJem0z+BsbFRVqfDYimFxmR0mQPw1YVZ2cXFwWz//abcT4ZmkFWUkw9?=
=?us-ascii?Q?5eyPwbxEE11oHbfNJ4ZIKpWS6Gc/dpWTNyQpxE3n1Qsa1NEI5yHkQ=3D=3D?=
X-Forefront-Antispam-Report:

CIP:2607:f8b0:4864:20::741;CTRY:;LANG:en;SCL:1;SRV:;IPV:NLI;SFV:NSPM;H:mail-qk1-x741.
google.com;PTR:mail-qk1-x741.google.com;CAT:NONE;SFS:(13230040)(48200799018)(9140799003)(7093399012)(
61400799027)(376014)(8096899003);DIR:OUT;SFP:1102;
X-ExternalRecipientOutboundConnectors:bef10d5f-a5b4-435d-ab3d-0abc125487cd
X-MS-Exchange-ForwardingLoop:
dm1.klbemp@cyrix.in;bef10d5f-a5b4-435d-ab3d-0abc125487cd
[Quoted text hidden]

SANDEEP DAS M <s.das@transasia.co.in>
To: Akhilesh P <akhileshcyrix@gmail.com>
Cc: Blesson Jose <Zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Tue, 17 Dec, 2024 at 11:50 am

Dear Akhilesh ,

Please find the quotation for the same ans do the needful

Get Outlook for Android

From: Akhilesh P <akhileshcyrix@gmail.com>
Sent: Friday, December 6, 2024 9:04:48 AM


To: SANDEEP DAS M <s.das@transasia.co.in>
Cc: Blesson Jose <Zm2.klbemp@cyrix.in>; DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>
Subject: REQUIRED QUOTATION FOR FULLY AUTOBIOCHEMISTRY ANALYZER 0913283 # 149993 @ TRIBAL SPECIALITY HOSPITAL KOTTATHARA , PALAKKAD

[Quoted text hidden]

This is to certified that this Subst^{ns}
Equipment Register Vol. I contains 448
pages numbered serially from 1-448

Kottathara


~~Superintendent~~
[Signature]



[Handwritten signature]
28/2/18

Superintendent
Govt. Tribal Speciality Hospital
Kottathara (P.O), Attappady

Suresh Babu. M
Store Verification Officer
Dist. Medical Office (Health)
Palakkad



[Handwritten signature]

Superintendent
Govt. Tribal Speciality Hospital
Kottathara (P.O), Attappady

(974)67(254)09(21)5217

CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)



TOLL FREE NO:1800 - 425-7669



BAR CODE - (8004-890615225) 0913283

CE

TRANSASIA

AUTOMATED RANDOM ACCESS CLINICAL CHEMISTRY ANALYSER

MODEL : EM 360

SERIAL NUMBER : 00452

**POWER SOURCE : 230 V AC, 50 Hz
750 VA**

TRANSASIA BIOMEDICALS LTD

MADE IN INDIA

FOR IN VITRO DIAGNOSTIC USE

Signature

Superintendent
Govt. Tribal Speciality Hospital
Kottathara (P.O), Attappady

