

**PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: DH KANHNAGAD	Name of District: KASARGOD
Name of Equipment: PULSE OXIMETER	Equipment ID & Barcode: 1410654, 151082
Make: MASIMO	Model: RAD 97
Serial Number: 3000017224	Warranty details: No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 14/05/2018	Present status of the equipment: Fully damaged
Date of breakdown: 06/12/2024 Date of registration of complaint through email/Toll free): Toll free	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: Checked the machine, found Multisite reusable sensor, Patient cable m20, and Main RD to M-LNC adapter cable are defective. Enquired OEM for quotation.	

Recommendations for repair (required service details): Not recommending for repair

Cost of spares (specify parts and cost): Multisite reusable sensor: 10361, 4256 RD Rainbow SET patient cable, M20:13619, RD to M-LNC adapter cable 1.5FT: 4192
TOTAL: 28172

#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 62.32% Asset Value: 45202

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix service report and OEM quotation attached

Reasons for recommending the equipment as BER: Checked the machine, found Multisite reusable sensor, Patient cable m20, and Main RD to M-LNC adapter cable are defective. Machine installed on 14/05/2018 and aged up to 6 years and 6 months. Repairing cost of equipment is 62.32% with respect to asset value. Both criteria for RBER met. Considering the tender clause 5.3.14.1, We recommending the equipment for condemnation.

Name & Signature of CYRIX Authority with date
Jayesh Kurian 24/12/24

Remarks of Junior Consultant (Biomedical) NHM: checked and identified that patient cable, sensor and adapter cable are defective. Aged up to 6 years and above. its repairing cost 62.32% of asset value. As per tender condition recommended for BER as per Kmsch - approval

Recommended for BER (Yes/ No): Yes (Decision final from Kmsch)
Date: 24/12/24 Signature of JC BM (NHM)

Date: 30/12/2024 Signature of Superintendent / Medical Officer (i/c)

*Not mandatory *Attach Photograph

Based on the period of life and value as per the BER guidelines

കാസറഗോഡ് ജില്ലാ ആരോഗ്യവകുപ്പ്
കാഞ്ഞങ്ങാട്
കാസറഗോഡ് ജില്ല



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 128373

CYRIX[®]

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

<p>Health Facility..... <u>DH Kanhangad</u>.....</p> <p>Address..... <u>Kanhangad</u>.....</p> <p style="padding-left: 100px;"><u>Kasaragod (Dist)</u></p> <p>Ph:..... <u>7025641502</u>.....</p>	<p>Call Registration Date : <u>06/12/2024</u>.....</p> <p>Caller ID : <u>151082</u>.....</p> <p>Date of Visit : <u>07/12/2024</u>.....</p> <p>Asset No : <u>1410654</u>.....</p> <p>EQPT Name : <u>Pulse oximeter</u>.....</p> <p>Manufacture <u>Masimo</u> Model : <u>Rad 97</u>.....</p> <p>S. No. <u>3000017224</u> Dept.....</p>
---	--

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : SPO2 not getting

Action Taken : Checked the machine, found multisite reusable sensor, patient cable m20 and main PD to M-LNC, adapter cable are defective. Enquired OEM for quotation

Completed Date : 7/12/24 Time : 12pm Spare Required

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrux Engineer <u>Vasudevan.k</u>	Date <u>7/12/24</u>	Start Time <u>11:30am</u>	End Time <u>12pm</u>
-----------------------------------	---------------------	---------------------------	----------------------

Customer Remark Completed Pending

Service Engineer Name : <u>Vasudevan.k</u>	Customer Name : <u>KAMISHA K.V.</u>
Signature: <u>[Signature]</u>	Signature: <u>[Signature]</u>
Date: <u>07/12/2024</u>	Date: <u>7/12/24</u>
Contact Number: <u>9072522045</u>	Contact Number: <u>7025641500</u>

Hospital Seal


ഏകദേശം
 കാഞ്ഞിരം
 കാസർഗോഡ് ജില്ല

KANHANGAD - 671 315. INSTRUMENTS & EQUIPMENT


Model & Serial No.	Full address of supplier and Phone No.	Warranty	% of AMC and Amount	Installation Date	Phone No. Service Engineer or Installer	Service centre Address and Phone (if it is other than column 11)	End user	Sub Stock Page No.	Issue Date	Dated Sign. Name Designation of receiver
10	11	12	13	14	15	16	17	18	19	20
RFD-97 S-300017366	BIO MEDICAL EQUIPMENTS COMPANY - 391832 AA - VITE STRADIUM LINK ROAD APPTM. INTERNATIONAL STRADIUM, PIRAVATTOM Kochi - phone-0484- 2883267, 289215594	2 yr	3 yr	14-05-18	Shabir 9061592272	Same as 11	STCU	50 5B	14-05-18	Sh Sh
RFD-97 S-115 2000017324	"	2 yr	3 yr	14/05/18	Shabir 9061592272	"	NICU	1-A	14-05-18	Sh





ശ്രീ. എ. ജി. ശബരീ
 ഹെൽത്ത് കെയർ ഓഫീസർ
 കെ.എ.എ.ആർ. ഓഫീസ്
 കെ.എ.എ.ആർ. ഓഫീസ്



[Handwritten signature in green ink]
സംഗ്രഹം
ജി.ആ.പ. (ത)
അംഗീകൃതം






 30/12/14
 സുപ്രണ്ട
 ജില്ലാ ആശുപത്രി
 കാഞ്ഞങ്ങാട്
 കോസറക്കാട് ജില്ല
 കോസറക്കാട്

39/878 A2(YMJ Stadium Link Road) Opp. JLN International Stadium Metro Station,
Palarivattom Kochi PIN 682025 Kerala, India.
Email: sales@biomedicalengineeringcompany.com Tele: 0484 2887208, 0484
2887200, Mobile: 8921065594
TIN 32071362164, DL 20 B # KL-EKM111891 DL 21B # KL-EKM-111892
GST32AAGFB1151K1ZV

To,
The Purchase Manager
GYRIX HEALTHCARE PRIVATE
LIMITED
Ernakulam
682038

Quotation for Medical Equipment, Supplies, Accessories & Parts

Quote: Q1897425

Dated: 13-12-2024

SI No	Item	Qty	Rate	Tax %	Amount of Quantity (Tax Included)
1	4054 RD SET™ Y-I, Multisite Reusable sensor, 3 ft., Non-Sterile	1	9,251.00	12%	10,361.12
2	4256 RD rainbow® SET Patient Cable, M20?05, RD rainbow M20 Connector Patient Cable, 5 ft	1	11,542.00	18%	13,619.56
3	RD to M-LNC Adapter Cable 1.5FT	1	3,553.00	18%	4,192.54

Terms and Conditions

*** For Masimo Rad 97 , bearing serial number 3000017224 at DH Kanhangad

Taxes : Any variation on GST from above will be applicable at the time of billing

Additional cess / duties / tax will be extra, if applicable at the time of billing.

Payment : 50 % advance with PO, balance COD/bank transfer or as per contract.

Delivery : 6-8 weeks from date of confirmed order

A/c.No.99999388878006

HDFC Bank ,Kadavanthara Branch, IFSC code :HDFC0000295

GST : 32AAGFB1151K1ZV

Detailed Quotation

THANK YOU FOR YOUR BUSINESS



Jayesh Krishnan <jayeshcyrix@gmail.com>

PRICE ENQUIRY 1410654

7 messages

Anjitha T L <purchasecyrix2@gmail.com>

Thu, Dec 12, 2024 at 12:47 PM

To: Biomedical Engineering Company <becproductquotes@gmail.com>, jobby.bec@gmail.com, becproductquote@gmail.com, "BEC Healthcare (GE India, Non-GE)" <mail@bechealthcare.com>
Cc: "Cyrix Purchase Mgr." <purchasemgr@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, Blesson Jose <Zm2.klbemp@cyrix.in>, Jayesh Krishnan <jayeshcyrix@gmail.com>

Dear sir,

Please share the price and availability of SPO2 Probe for pulse oximeters with the following parts.

1. Masimo RD set multisite reusable sensor neonatal
2. Masimo RD patient cable M20-05
3. RD adapter cable RD to M-LNC 1.5ft

Make: Masimo
model: Rad 97
S/N: 3000017224

--

Thanks & Regards

Anjitha T L
Sr. Purchase Executive
PH:- 9947719111
purchasecyrix2@gmail.com
www.cyrixhealthcare.com

Cyrix Healthcare Pvt Ltd

| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India
KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH

Anjitha T L <purchasecyrix2@gmail.com>

Fri, Dec 13, 2024 at 5:28 PM

To: Biomedical Engineering Company <becproductquotes@gmail.com>, jobby.bec@gmail.com, becproductquote@gmail.com, "BEC Healthcare (GE India, Non-GE)" <mail@bechealthcare.com>
Cc: "Cyrix Purchase Mgr." <purchasemgr@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, Blesson Jose <Zm2.klbemp@cyrix.in>, Jayesh Krishnan <jayeshcyrix@gmail.com>

Reminder

[Quoted text hidden]

Jayesh Krishnan <jayeshcyrix@gmail.com>

Mon, Dec 16, 2024 at 9:35 AM

To: Biomedical Engineering Company <becproductquotes@gmail.com>
Cc: jobby.bec@gmail.com, Anjitha T L <purchasecyrix2@gmail.com>, becproductquote@gmail.com, "BEC Healthcare (GE India, Non-GE)" <mail@bechealthcare.com>, "Cyrix Purchase Mgr." <purchasemgr@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, Blesson Jose <Zm2.klbemp@cyrix.in>

Sir,

Kindly update at the earliest.

[Quoted text hidden]

--

Thanks & Regards

30/12/2024, 20:39

Gmail - PRICE ENQUIRY 1410654

JAYESH KRISHNAN,
Sr. Biomedical Engineer,
District In-charge
Kasargod.
PH:- 7593847156

jayeshcyrix@gmail.com

[Quoted text hidden]

Mail Delivery Subsystem <mailer-daemon@googlemail.com>
To: jayeshcyrix@gmail.com

Mon, Dec 16, 2024 at 9:35 AM



Address not found

Your message wasn't delivered to **becproductquote@gmail.com** because the address couldn't be found, or is unable to receive mail.

[LEARN MORE](#)

The response was:

550 5.1.1 The email account that you tried to reach does not exist. Please try double-checking the recipient's email address for typos or unnecessary spaces. For more information, go to https://support.google.com/mail/?p=NoSuchUser_00721157ae682-6f28923385asor23538517b3.16 - gsmt

Final-Recipient: rfc822; becproductquote@gmail.com

Action: failed

Status: 5.1.1

Diagnostic-Code: smtp; 550-5.1.1 The email account that you tried to reach does not exist. Please try

550-5.1.1 double-checking the recipient's email address for typos or

550-5.1.1 unnecessary spaces. For more information, go to

550 5.1.1 https://support.google.com/mail/?p=NoSuchUser_00721157ae682-6f28923385asor23538517b3.16 - gsmt

Last-Attempt-Date: Sun, 15 Dec 2024 20:05:42 -0800 (PST)

----- Forwarded message -----

From: Jayesh Krishnan <jayeshcyrix@gmail.com>

To: Biomedical Engineering Company <becproductquotes@gmail.com>

Cc: jobby.bec@gmail.com, Anjitha T L <purchasocyrix2@gmail.com>, becproductquote@gmail.com, "BEC Healthcare (GE India, Non-GE)" <mail@bechealthcare.com>, "Cyrix Purchase Mgr." <purchasemgr@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, Blesson Jose <Zm2.klbemp@cyrix.in>

Bcc:

Date: Mon, 16 Dec 2024 09:35:29 +0530

Subject: Re: PRICE ENQUIRY 1410654

----- Message truncated -----

Biomedical Engineering Company <becproductquotes@gmail.com>
To: Jayesh Krishnan <jayeshcyrix@gmail.com>

Mon, Dec 16, 2024 at 11:21 AM

30/12/2024, 20:39

Gmail - PRICE ENQUIRY 1410654

Cc: jobby.bec@gmail.com, Anjitha T L <purchasecyrix2@gmail.com>, becproductquote@gmail.com, "BEC Healthcare (GE India, Non-GE)" <mail@behealthcare.com>, "Cyrix Purchase Mgr." <purchasemgr@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, Blesson Jose <Zm2.klbemp@cyrix.in>

Kindly share the customer details

Warm regards,

Shaji

team BEC

[Quoted text hidden]

Anjitha T L <purchasecyrix2@gmail.com>

Mon, Dec 16, 2024 at 11:36 AM

To: Biomedical Engineering Company <becproductquotes@gmail.com>

Cc: Jayesh Krishnan <jayeshcyrix@gmail.com>, jobby.bec@gmail.com, becproductquote@gmail.com, "BEC Healthcare (GE India, Non-GE)" <mail@behealthcare.com>, "Cyrix Purchase Mgr." <purchasemgr@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, Blesson Jose <Zm2.klbemp@cyrix.in>

Dear sir,

Hospital Name: DH Kanhangad

[Quoted text hidden]

Biomedical Engineering Company <becproductquotes@gmail.com>

Mon, Dec 16, 2024 at 11:43 AM

To: Anjitha T L <purchasecyrix2@gmail.com>

Cc: jobby.bec@gmail.com, becproductquote@gmail.com, "BEC Healthcare (GE India, Non-GE)" <mail@behealthcare.com>, "Cyrix Purchase Mgr." <purchasemgr@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, Blesson Jose <Zm2.klbemp@cyrix.in>, Jayesh Krishnan <jayeshcyrix@gmail.com>

Madam,

Pls find attached as per discussed over phone

Warm regards,

Shaji

Team BEC

[Quoted text hidden]

 **Cyrix Masimo accessories for DH kanhangad.pdf**
239K