

PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: THQH PUTHUKKAD	Name of District: Thrissur
Name of Equipment: Fumigator	Equipment ID & Barcode: #147899- 0822207
Make: Goley	Model: AD-5L
Serial Number: NA	Warranty details: No Warranty
Date of purchase/Year of manufacture/ Installation Date: Ticket master installation date is- 05-09-2019 Actual installation date as per Stock book is: 29-01-2015	Present status of the equipment: Fully damaged
Date of breakdown: 20-11-2024	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Date of registration of complaint through email/Toll free): Toll free	
Action taken: Checked the machine and found fogger main housing, Vaccum motor, Tank, Spring wound timer and control board are defective. Need to replace these spares for further checking working condition of equipment.	
Recommendations for repair (required service details): Not Recommending	
Cost of spares (specify parts and cost): Fogger main housing - 9500/- Vaccum motor - 3000/- Tank - 1000/- Spring wound timer - 850/- Control board - 500/- <p style="text-align: center;">TOTAL - 15,512/-</p>	

#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 74.4%	Asset Value: 20833/-
--	----------------------

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix report and OEM quotation attached

Reasons for recommending the equipment as BER: Upon inspecting the machine, it was found that the fogger main housing, vacuum motor, tank, spring-wound timer, and control board are defective. According to the ticket master, the machine's installation date is 05-09-2015. However, the stock book (attached for reference) records the installation date as 29-01-2015, making the equipment 9 years and 10 months old. The repair cost is 74.4% of the asset value, thereby meeting both criteria for RBER. In accordance with tender clause 5.3.14.1, we recommend the equipment for condemnation.

[Signature]
BIJO T JOY
 Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM:
Repair cost is very high

[Signature]
Jr. Consultant (Biomedical)

Recommended for BER (Yes/ No): <i>yes</i>	<i>[Signature]</i> Signature of JC BM (NHM)
Date:	

Date:	<i>[Signature]</i>	Signature of Superintendent / Medical Officer (i/c)
-------	--------------------	---

*Not mandatory //Based on the period of life span value as per the BER guidelines * Attach Photograph





BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 138077

CYRIX[®]
HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility..... <u>THOH pathuvuud</u>	Call Registration Date : .. <u>20/11/24</u>
Address..... <u>Thrissur</u>	Caller ID : .. <u>14 7849</u>
.....	Date of Visit : .. <u>21/11/24</u>
Ph:	Asset No : .. <u>08.222.07</u>
	EQPT Name : .. <u>Fumigator</u>
	Manufacture..... <u>Goleg</u> Model : <u>AD-5L</u>
	S. No..... <u>Na</u> Dept. <u>Labour room</u>

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : Fumigator not working

Action Taken : checked the machine and found fogger main housing vacuum motor, tank, spring wound timer and controller board are defective need to replace these spares w.t. checking further working condition of the equipment.

Completed Date : .. 21/11/24 .. Time : .. Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time

Customer Remark Completed Pending

Service Engineer Name : Suhay
Signature: [Signature]
Date: 21/11/24
Contact Number: 7593847135

Customer Name : Shmy
Signature: [Signature]
Date: 21/11/24
Contact Number: 9497062131
Designation : SNO
Hospital Seal



**SUPERINTENDENT
TALUK HOSPITAL
PUDUKKAD**

REF: BEC/DHM/ 01/07-2024
16-07-2024

M/s. District Hospital
Mavelikkara

Pricing details of accessories for ULV fogger

S.No	Description	Qty	Pricing INR (exc.GST)	GST	Total Price (inc.GST)
1	Fogger main housing	1.00	3297.30	18.00	3890.81
2	Vaccum motor	1.00	5973.30	18.00	7048.49
3	Tank	1.00	1898.10	18.00	2239.76
4	Spring wound timer	1.00	1979.30	18.00	2335.57

TERMS & CONDITIONS

Taxes: As shown above

Payment : 50% advance with PO, balance COD/bank transfer against delivery

Warranty : NOT APPLICABLE

Delivery : 6-8 weeks from date of submitted quote

With Regards

Authorized Signatory

Biomedical Engineering Company (BEC)

Bank details :

M/s. Biomedical Engineering Company

HDFC Bank, Kadavanthara Branch ,IFSC code HDFC0000295

A/c: 99999388878006



supply details- thqh puthukad -0822207 Caller id - 147899

5 messages

Bijo T Joy <bijocyrix@gmail.com>
To: KMSCL <kmsclbmecentral1@gmail.com>
Cc: Blesson Jose <zm2.klbemp@cyrix.in>

Wed, 18 Dec, 2024 at 11:17 am

Dear sir,
There is a complaint reported from THQH PUDUKAD, Thrissur, Kerala, Regarding the complaint with , Goley,AD-5L fumigator machine ,Please share the supply details of the machine..

Equipment - Fumigator

Make : Goley

Model: AD-5L

Serial No: NA

Hospital- THQH Pudukad

District - Thrissur

Barcode - 0822207

Caller id - 147899

With Regards,

BIJO T JOY

DISTRICT INCHARGE THRISSUR,

CYRIX HEATHCARE PRIVATE LIMITED

30/641 B, Petta Junction | Poonithura | Kochi | Kerala-682038 Branches: Kerala | Karnataka || Tamil Nadu || Uttar Pradesh | Kuwait | Saudi Arabia

M.-8921870010 | E:- bijocyrix@gmail.com | W:- www.cyrixhealthcare.com | "

Bijo T Joy <bijocyrix@gmail.com>
To: KMSCL <kmsclbmecentral1@gmail.com>
Cc: Blesson Jose <zm2.klbemp@cyrix.in>

Sat, 21 Dec, 2024 at 7:31 am

reminder

With Regards,

BIJO T JOY

DISTRICT INCHARGE THRISSUR,

CYRIX HEATHCARE PRIVATE LIMITED

30/641 B, Petta Junction | Poonithura | Kochi | Kerala-682038 Branches: Kerala | Karnataka || Tamil Nadu || Uttar Pradesh | Kuwait | Saudi Arabia

M.-8921870010 | E:- bijocyrix@gmail.com | W:- www.cyrixhealthcare.com | "

[Quoted text hidden]

KMSCL <kmsclbmecentral1@gmail.com>
To: Bijo T Joy <bijocyrix@gmail.com>
Cc: Blesson Jose <zm2.klbemp@cyrix.in>

Sat, 21 Dec, 2024 at 1:53 pm

Supplier Name - **M/s Biomedical Engineering Company**

Contact Details - Mr. Ajith George Mathew

8921065594

9447012546

Email id - mail@bechealthcare.com

accounts@bechealthcare.com

sales@bechealthcare.com

[Quoted text hidden]

—
Regards,

SHALU U S

DISTRICT BIOMEDICAL ENGINEER

KMSCL

Contact No: +91 91889 10013

Bijo T Joy <bijocyrix@gmail.com>
To: KMSCL <kmsclbmecentral1@gmail.com>
Cc: Blesson Jose <zm2.klbemp@cyrix.in>

Thu, 26 Dec, 2024 at 10:15 am

Dear mam,

It seems like you're referring to a company that you have contact with, but they do not supply fumigator to THQH Puthukad hospital and no supply details available at hospital.

With Regards,

BIJO T JOY

DISTRICT INCHARGE THRISSUR,

CYRIX HEATHCARE PRIVATE LIMITED

30/641 B, Petta Junction | Poonithura | Kochi | Kerala-682038 Branches: Kerala | Karnataka || Tamil Nadu || Uttar Pradesh | Kuwait | Saudi Arabia
M:-8921870010 | E:- bijocyrix@gmail.com | W:- www.cyrixhealthcare.com | "

[Quoted text hidden]

KMSCL <kmsclbmececentral1@gmail.com>
To: Bijo T Joy <bijocyrix@gmail.com>
Cc: Blesson Jose <zm2.klbemp@cyrix.in>

Thu, 26 Dec, 2024 at 10:16 am

Noted with thanks.

[Quoted text hidden]

