

PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: THQH KUTTIADI	Name of District: KOZHIKODE
Name of Equipment: UPS	Equipment ID & Barcode: 152275 & 1120775
Make: ENERGIO	Model: RESQ1K3SB
Serial Number: NA	Warranty details: No Warranty
Date of purchase/Year of manufacture/ Installation Date: 21/02/2018	Present status of the equipment: Fully damaged
Date of breakdown: 12/12/2024	*AMC/ CAMC Period agreed
Date of registration of complaint through email/Toll free): Toll free	at the time of purchase: No AMC/CAMC
Action taken: Checked and found Power board defective. Enquired spares with oem.	

Recommendations for repair (required service details): Not recommending for repair
Cost of spares (specify parts and cost): 1) Power Board - 1 nos = 6000 /- Total = 6000 /- GST 18% = 1080 /- GRAND TOTAL = 7080 /-

#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 157.33 %	Asset Value: 4500 /-
---	----------------------

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix service report attached and vendor quotation attached.

Reasons for recommending the equipment as BER: Checked and found Power board defective. Equipment installed on 21/02/2018 (stock book attached for reference). Equipment out lived 6 years 10 months. Spare cost is 157 % with respect to asset value, both criteria for RBER met, As per tender clause 5.3.14.1 we recommending the equipment for condemnation.

JINEESH PP *Jineesh*
 Name & Signature of CYRIX Authority with date *31/12/2024*

Remarks of Junior Consultant (Biomedical) NHM:
checked & found that power board damaged. Equipment installed on 21/02/2018. spare is cost high. So recommended for condemnation.

Recommended for BER (Yes/ No): Yes	Consultant Bio Medical Engineer National Health Mission Kozhikode Signature of JE BM (NHM)
Date: <i>11/1/2025</i>	

Date: <i>4/1/2025</i>	Seal 	Signature of Superintendent / Medical Officer (i/c)
-----------------------	--	---

*Not mandatory #Based on the period of life and value as per the BER guidelines * Attach Photograph

*SUPERINTENDENT
TALUK HOSPITAL
KUTTIADI*



PKL POWER SYSTEMS

QUOTATION

service and maintenance of ups&batteries

10/684 VENGARI PO, KOZHICODE ,KERALA-673010
GSTIN : 32AVSPJ0663P1Z2
MAIL:pklpowersystems@gmail.com
PH:6282495532

DATE :26.12.2024
QUOTE :01/16
VALIDITY:15 DAYS

CUSTOMER INFO

CYRIX HEALTHCARE PRIVATE LIMITED
1ST FLOOR,30/641B,CYRIX HEALTHCARE PVT LTD,
KOCHI DHANUSHKODI ROAD,POONITHURA,
ERNAKULAM

DESCRIPTION OF WORK

ENERGIO MAKE UPS SERVICE

ITEMIZED COSTS	QTY	UNIT PRICE	AMOUNT		
POWER BOARD	1	6000.00	6000.00		
Thank you for your business	SUB TOTAL		6000.00		
	GST 18%		1080.00		
	TOTAL QUOTE		7080.00		
This quote is not a contract or a bill.It is our best guess at the total price for the service described above.The customer will be billed after indicating acceptance of this quote.payment will be prior to the delivery of service.please mail the signed quote to the address listed above.	<table border="1"><tr><td>Customer acceptance</td></tr><tr><td> </td></tr></table>			Customer acceptance	
				Customer acceptance	
If you have any questions,please contact Mr.jibiraj 999576918					



Request for spare quotation-152275-1120775 - UPS ,THQH KUTTIADI, KOZHIKODE

4 messages

Jineesh P P <jineeshcyrix@gmail.com>
To: pklpowersystems@gmail.com
Cc: Blesson Jose <zm2.klbemp@cyrix.in>, Arun K K <arunkkcyrix@gmail.com>

Sat, 21 Dec, 2024 at 6:52 pm

Dear Sir,

Please share price and spare availability of power board (main board).
Details as following ,

Barcode no: 1120775
Ticket id: 152275
Facility name: Thqh Kuttiadi

UPS (1KVA)
Make : Energio
Model: RESQ1K3SB
SN: NA

Thanks & Regards,

Jineesh PP
District Incharge - Kozhikode
KLBEMMP
Mob-7593847163
jineeshcyrix@gmail.com
www.cyrixhealthcare.com

Cyrix Healthcare Pvt Ltd

| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India

KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH

PKL POWER SYSTEM'S <pklpowersystems@gmail.com>
To: Jineesh P P <jineeshcyrix@gmail.com>
Cc: Blesson Jose <zm2.klbemp@cyrix.in>, Arun K K <arunkkcyrix@gmail.com>

Sun, 22 Dec, 2024 at 9:24 pm

Sir,
As per your request for spares for ups power board, we are sending the best price quote, please find the attachment.

Regards
Antu kv
Service co-ordinator
[Quoted text hidden]

PKL POWER SYSTEM'S <pklpowersystems@gmail.com>
To: Jineesh P P <jineeshcyrix@gmail.com>

Thu, 26 Dec, 2024 at 10:45 am

Sir,
We are sendind the revised service quote, please find the attachment.

On Sat, 21 Dec 2024, 6:53 pm Jineesh P P, <jineeshcyrix@gmail.com> wrote:
[Quoted text hidden]

PKL POWER SYSTEM'S <pklpowersystems@gmail.com>

Mon, 13 Jan, 2025 at 10:47 am


To: Jineesh P P <jineeshcyrix@gmail.com>

Sir, We are sending the revised service quote, please find the attachment.
[Quoted text hidden]

STOCK BOOK OF ULTRASONIC SCANNING 121
MECHING

Name of article _____

Date	From whom received or to whom issued	Receipt	Issued	Balance	Batch No Exp. Date Mfg. Date	Price per Unit	Remarks
	Brought Forward						
21/2/16	Received from store	1		1		10.50	
	EC-6 page 73		L.P (11mc)			11319106	
1	DC 90 US Scanning meter	1					
2	3c 5P Convex probe	1					
3	554383 Linear probe	1					
4	2p 2p Convex probe	1					
5	Sony printer UP-7896MD	1					
6	Energis 1kVA 36VDC SD model	1					
7	Rocket battery 7AH	3					



502
 SUPERINTENDENT
 TALUK HOSE
 KUTTIADY





- This unit may be operated by untrained technicians only!
- Even when switched off there is a hazardous voltage on the battery side!
- Lead acid battery in the inside of the enclosure!




CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)

TOLL FREE NO: 1800 - 425-7669

BAR CODE - (8004-890615225) 1120775

IS 16242 (Part 1) / IEC 62040-1




R-41067318

Model: RESQ1K3SB

Input: 220Vac-240Vac, 50/60Hz, 5.6A, 1ϕ

Output: 220Vac-240Vac, 50/60Hz, 4.3A, 1ϕ

Capacity: 1000VA/900W



83121708101300

MADE IN CHINA

Handwritten signature and date: 20/11/2011