

**PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: PHC ANDATHODE	Name of District: Thrissur
Name of Equipment: Oxygen Concentrator	Equipment ID & Barcode: 146307- 0840659
Make: Owgels oxy-med	Model: OM-501
Serial Number: NA	Warranty details: No Warranty
Date of purchase/Year of manufacture/ Installation Date: Ticket master installation date – 15-01-2019, As per the Stock book installation date- 26-12-2018	Present status of the equipment: Fully damaged
Date of breakdown: 12-11-2024 Date of registration of complaint through email/Toll free): Toll free	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: Upon inspecting the machine, it was discovered that the compressor, sieve bed, pressure regulator and solenoid valve are faulty. Need to replace these spares for further checking working condition of equipment.	
Recommendations for repair (required service details): Not Recommending	
Cost of spares (specify parts and cost): NA	
#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA	Asset Value: 33400/-
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix report and EOL letter attached.	
Reasons for recommending the equipment as BER: Upon inspecting the machine, it was discovered that the compressor, sieve bed, pressure regulator and solenoid valve are faulty. Ticket master installation date is 15-01-2019 ,As per the stock book Installation date is 26-12-2018 (stock book attached for reference) and aged up to 5 years and 11months. OEM informed that this was an obsolete model spares are not available. Obsolete letter issued. As per tender clause 5.3.14.2 We recommending the equipment for condemnation.	
<p align="right">BIJO T JOY <i>[Signature]</i> Name & Signature of CYRIX Authority with date</p>	
Remarks of Junior Consultant (Biomedical) NHM: <i>obsolete model</i>	<p align="center"><i>[Signature]</i> Jr. Consultant (Biomedical) NHM, Thrissur.</p>
Recommended for BER (Yes/ No): <i>Yes</i>	<p align="right"><i>[Signature]</i> 30/11/2024 Signature of JC BM (NHM)</p>
Date:	
Date: <i>2/12/2024</i>	<p align="center">Seal <i>[Seal: FAMILY HEALTH CENTER ANDATHODE, PHC, THRISSUR, PIN-679564]</i></p>

*Not mandatory


#Based on the period of life of its value as per the BER guidelines

[Signature]
OFFICER
Medical Officer
F.H.C. ANDATHODE
THRASSUR
PHN- 679564

Obsolete Letter for Oxy-med Owgels Model (3 Lpm / 5 Lpm)

Dear valued customers,

Medequip has discontinued support for the OXY-MED OWGELS 3Lpm and 5Lpm models, repair parts are no longer available for these models.



Please contact our OXY-MED Support staff or our Service team.

If you wish to exchange your old product with new updated Oxy-med -Mini Product, Please contact our OXY-MED support staff or our service team. Should you have any commercial or product related questions please contact Mr. Jayakumar for Further Information

Thanks & Regards

Jayakumar
Service Manager- India
Mobile- +918795660006
Email-jayakumar@medequip.co.in



Oxygen Concentrator

238

26-12-18

Aceta Medica store, Kundikal, Mn no 142

0840659

Rs 50.000/-

3 No

✓

3 No

15-1-19

1st set of P-10

3

NIL

