


PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: CHC KADAPLAMATTOM	Name of District: KOTTAYAM
Name of Equipment: Oxygen Concentrator	Equipment ID & Barcode: 145801 & 0532849
Make: Yuwell	Model: 8F-5A
Serial Number: 210501943	Warranty details: No warranty
Date of purchase/ Year of manufacture/ Installation Date: 26/07/2021	Present status of the equipment: Fully Damaged
Date of breakdown: 9/11/2024 (Toll free) (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: No AMC /CAMC
Action taken: Checked and found out the problem with Main pcb, and compressor. Need to replace these spares for further checking and working condition of the equipment. Enquired the spare with OEM	
Recommendations for repair (required service details) : Not recommending for repair.	
Cost of spares (specify parts and cost): NA	
# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA	Asset Value: 41600/-
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix service report and mail communication attached	
Reasons for recommending the equipment as RBER: Checked and found the problem with main pcb and compressor. Need to replace these spares for further checking and working condition of the equipment.. Enquired spares with oem . Oem not responded. so recommending the unit for condemnation	
Josephy George <i>[Signature]</i> 22/11/2024 Name & Signature of CYRIX Authority with date	
Remarks of Junior Consultant (Biomedical) NHM: <i>Asset value - 41600 - Cost of spares not submitted. - Life period -> 3 year 5 months - Cyrix reported that spare not available - as per the guidelines recommended as RBER</i>	
Recommended for BER (Yes/ No): <i>PLS</i>	 Signature of JC BM (NHM)
Date: <i>27/11/24</i>	
Date: <i>13/11/25</i> Medical Officer (i/c) <i>[Signature]</i> For M.O i/c.	Signature of Superintendent /

For MEDICAL OFFICER IN CHARGE
 CHC - KADAPLAMATTOM





BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 1013231

CYRIX®

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility..... <u>CHC Kadaplattom</u> Address..... <u>kottayam</u> Ph:..... <u>7909145767</u>	Call Registration Date : <u>9/11/2024</u> Caller ID : <u>145801</u> Date of Visit : <u>9/11/2024</u> Asset No : <u>0532849</u> EQPT Name : <u>Oxygen concentrator</u> Manufacture..... <u>Yuwell</u> Model : <u>BF-SA</u> S. No. <u>210501943</u> Dept. <u>Palliative care</u>
---	--

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : Not working

Action Taken : check the machine sieve bed, main board and Solinoid valve with gas exchanger unit not working. Need to replace these spare for further checking and working condition of the equipment. Enquired the spare with OEM.

Completed Date : 9/11/2024 Time : 4pm Spare Required

Spare Replaced Requested

1.	Description	Qty.	Part Number	PR Number
2.				
3.				

Cyrix Engineer <u>Josekutty George</u>	Date <u>9/11/2024</u>	Start Time <u>12 pm</u>	End Time <u>4 pm</u>
---	--------------------------	----------------------------	-------------------------

Customer Remark Completed Pending
Not working

Service Engineer Name : <u>Josekutty George</u> Signature: <u>[Signature]</u> Date: <u>9/11/2024</u> Contact Number: <u>8848230407</u>	 Customer Name : <u>AXIE.P.T</u> Signature: <u>[Signature]</u> Date: <u>9.11.24</u> Contact Number: <u>7012159100</u> Designation : <u>SNO</u> Hospital Seal <u>[Seal]</u>
---	---

[Signature]
DR. VIKAS
MEDICAL OFFICER IN CHARGE
CHC - KADAPLATTOM






STOCK REGISTER

Date	From whom received or to whom issued	Receipt	Issue	Balance	Initials of Receiver	Remarks
9/1/21	Rd from store (KMSCL)	2 → (811145-200)		2	W	(KMSCL)
"	RA from store (large)	10L	1 - (41102A)	1	W	
10/7/21	1/20 substock reg. B (small)	5L	2	Nil	W	
"	1/20 substock reg. B (small) (CP 244)	10L	1	Nil	W	1
26/7/21	Rd from store (KMSCL)	2 (40WELL)		2	W	(KMSCL)
26/7/21	1/20 substock reg. B (pg. 244)	med. equipment 5L	2	Nil	W	"
26/7/21	Rd from store (5L)	1		1	W	"
26/7/21	1/20 substock reg. B pg. 244 (SUMA-0 Batch No. JM15A)		1	Nil	W	
19.12.21	Rd from store (5L) NHM Supply.	1		1	W	
14/12/22	1/20 substock reg. B pg. 244		1	Nil		
9/5/22	Received from store	2	2	2	W	
	1 KMSCL supply Model No CR-P30, supplier name Cancer Medical technology					
	One Donation (World Malayalee council India Model: DA-1A control)					
	Mfg. by: Shenzhen Ooda Health Co. Ltd, Guangzhou, China (Cang?)					
9/5/22	Issued to primary palliative care	2	2	0	W	Received from (170)
19/4/23	Rd from the store	1		1	W	
19/4/23	1/20 substock reg. B (Ventox health care) (25890/-)	1		Nil	W	
	Verified	Riniy Joseph				
		Secondary Palliative Nurse				



DR. V. Jithu
16/12/24

MEDICAL OFFICER IN CHARGE
CHC - KADAPLAMATTOM



Dr. Vidhu
12/12/24
MEDICAL OFFICER IN CHARGE
CHC - KADAPLAMATTOM



Arun GC <arungcyrix@gmail.com>

**REGARDING QUOTATION - 0532849, 145801CHC KADAPLAMATTOMOXYGEN
CONCENTRATOR**

4 messages

Arun GC <arungcyrix@gmail.com>

Tue, Nov 12, 2024 at 6:59 PM

To: "yuwellindia@gmail.com" <yuwellindia@gmail.com>, "sales@yuwell.co.in" <sales@yuwell.co.in>, DM Abijith-KLBEMP <dm2.klbemp@cyrix.in>, Afsal Yunuze <zm1.klbemp@cyrix.in>

Dear Sir,

Kindly share the quotation for Oxygen concentrator(Make-yuwell Model-8F-10 SI no-210501943) spares

- 1)Main board
- 2)Compressor

Thanks & Regards,

ARUN GC

District incharge

8921992272

arungcyrix@gmail.com

www.cyrixhealthcare.com

Cyrix Healthcare Pvt Ltd

| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India

KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH

Arun GC <arungcyrix@gmail.com>

Wed, Nov 13, 2024 at 9:34 PM

To: "yuwellindia@gmail.com" <yuwellindia@gmail.com>, "sales@yuwell.co.in" <sales@yuwell.co.in>, DM Abijith-KLBEMP <dm2.klbemp@cyrix.in>, Afsal Yunuze <zm1.klbemp@cyrix.in>

Cc: josekutty <josekuttycyrix@gmail.com>

REMINDER

Arun G C

District Incharge

KLBEMMP, Kottayam,

Cyrix Health Care.

[Quoted text hidden]

Arun GC <arungcyrix@gmail.com>

Fri, Nov 15, 2024 at 9:30 PM

To: "yuwellindia@gmail.com" <yuwellindia@gmail.com>, "sales@yuwell.co.in" <sales@yuwell.co.in>, DM Abijith-KLBEMP <dm2.klbemp@cyrix.in>, Afsal Yunuze <zm1.klbemp@cyrix.in>

Cc: josekutty <josekuttycyrix@gmail.com>

REMINDER

Arun G C

District Incharge

KLBEMMP, Kottayam,

Cyrix Health Care.

[Quoted text hidden]

11/22/24, 9:27 AM

Gmail - REGARDING QUOTATION - 0532849, 145801CHC KADAPLAMATTOMOXYGEN CONCENTRATOR

Arun GC <arungcyrix@gmail.com>

Fri, Nov 22, 2024 at 9:27 AM

To: "yuwellindia@gmail.com" <yuwellindia@gmail.com>, "sales@yuwell.co.in" <sales@yuwell.co.in>, DM Abijith-KLBEMP <dm2.klbemp@cyrix.in>, Afsal Yunuze <zm1.klbemp@cyrix.in>

Cc: josekutty <josekuttycyrix@gmail.com>

Reminder

Arun G C
District Incharge
KLBEMMP, Kottayam,
Cyrix Health Care.

[Quoted text hidden]