



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



680

SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 134909

CYRIX®

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website :www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility..... <u>GMHC</u>	Call Registration Date : <u>21-11-2024</u>
Address..... <u>Peroorkada</u>	Caller ID : <u>14.80.6.0</u>
..... <u>Thiruvananthapuram.</u>	Date of Visit : <u>20-12-2024</u>
Ph : <u>9061405749</u>	Asset No : <u>0117047</u>
	EQPT Name : <u>Haemo. Diress</u>
	Manufacture : <u>Transasia Biomedical</u> : Model : <u>SR 20 A</u>
	S. No. <u>1045</u> : Dept. <u>Laboratory</u>

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : Not working

Action Taken : Checked and found that the AC motor and gear assembly defective. Need to replace the spare for further checking and working condition of the equipment. Enquired spares from vendor.

Completed Date : 20/12/24 Time : 11:00 AM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2. <u>NA</u>	<u>NA</u>	<u>NA</u>	<u>NA</u>
3.			

Cyrix Engineer	Date	Start Time	End Time
	<u>Akshay Manohar</u>	<u>20/12/24</u>	<u>10:00 AM</u>

Customer Remark Completed Pending

Service Engineer Name : Akshay Manohar
Signature: [Signature]
Date: 20/12/2024
Contact Number: 98472 99500

SUPERVISOR
Mental Health Centre
Peroorkada
Thiruvananthapuram
Ph: 98472 99500



SUNIL KUMAR J
[Signature] 20/12/24
9061405249
Lab Incharge CMC

PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)


Name of Hospital: GMHC PEROORKADA	Name of District: THIRUVANANTHAPURAM
Name of Equipment: HAEMO MIXER	Equipment ID & Barcode: 148060/0117047
Make: TRANSASIA BIOMEDICAL	Model: SR 20A
Serial Number: 1045	Warranty details: NA
Date of purchase/ Year of manufacture/ Installation Date: 25/02/2009	Present status of the equipment: Fully Damaged
Date of breakdown: 21/11/2024 (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: Checked and found that the AC Motor and Gear assembly defective. Need to replace the spare for further checking and working condition of the equipment. Enquired spares from vender.	
Recommendations for repair (required service details): Not recommending for repair.	
Cost of spares (specify parts and cost): 1. AC Motor = 3205/- 2. Gear Assembly = 1790/- Total = 4995/-	
# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 65.5%	Asset Value: 7620/-
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix Service report and Vendor quotation attached	
Reasons for recommending the equipment as BER: Checked and found that the AC Motor and Gear Assembly defective. The unit was installed on 25/02/2009 and covered up to 15+ years. The repair cost is 65.6%. As per the tender clause 5.3.14.2 both criteria met. So recommending the unit for condemnation.	
Karyep.p.v <i>[Signature]</i> Name & Signature of CYRIX Authority with date	

Remarks of Junior Consultant (Biomedical) NHM:
AC motor and gear assembly defective. Verified the estimate for repair. ∴ can be RBER as per tender norms. Submitted on 27-12-2024

Recommended for BER (Yes/ No):
 Date: 31.12.2024

[Signature]
MANEESHA MOHAN M.R.
 Junior Consultant (Biomedical)
 National Health Mission
 Thiruvananthapuram
 Signature of JC BM (NHM)

Date: _____



Seal

[Signature]
Superintendent
 Mental Health Centre
 PEROORKADA
 Thiruvananthapuram
 Pin: 695005

Fully Automated Blood Cell Counter

Date	Particulars	Receipt	Balance	Remarks
25/2/2009	18 piece fully Automated Blood Cell Counter Model - SY ^S NEX - KX - 21 [FHAK 80002] Transasia Bio. Medicals Ltd Mumbai - 72.	1	Nil	Rs. 3,70,000/-
			working in Lab.	
	Blood mixture (BM-11) SIZ - 2.0A (FBI CH 00 21)	11		Compliment-
	Caps - APC make 500 vials, 15 mint backup In bill (FUPNA 0064)	1		Compliment-

Carried over to page 42 of book III.

copy attached


Saunder

1/30



SUPERINTENDENT
 Mental Health Centre
 Peroorkada
 Thiruvananthapuram
 Pin: 695005

Estimate


 <p>CLOUD99 BIOTECH ROOM NO: 17 VELLAPALLY BUILDING VELLAPALLY LANE OPP. GENERAL HOSPITAL KOTTAYAM, KERALA PIN 686001 Phone no: 9947213555 Email: cloud99biotech@gmail.com GSTIN: 32AM1PT5633Q1Z0 State: 32-Kerala MSME NO: : UDYAM-KL-07--0030714 DL NO: KL/KTA/MD42/2023/000015</p>	Estimate No. C99B-QTN-95	Date 26-12-2024
	Place of supply 32-Kerala	

Estimate For
KASHTEP
 1ST FLOOR 30/541B CYRUX HEALTHCARE PVT LTD KOCHI DHANUSHKODI ROAD POONITHURA
 Contact No. : 7593847134
 GSTIN : 32AA4CC2499H2ZM
 State: 32-Kerala

#	Item name	HSN/ SAC	Quantity	Unit	Price/ Unit	GST	Amount
1	AC MOTOR	8501	1	Nos	₹ 2,716.10	₹ 488.90 (18%)	₹ 3,205.00
2	GEAR ASSEMBLY	8511	1	Nos	₹ 1,516.95	₹ 273.05 (18%)	₹ 1,790.00
Total						₹ 761.95	₹ 4,995.00

Estimate Amount in Words Four Thousand Nine Hundred Ninety Five Rupees only	Amounts Sub Total ₹ 4,995.00
Description HOSPITAL NAME: GMCH, PEROORKADA EQUIPMENT: HAMEO MIXER MAKE: TRANSASIA MODEL: SR20A BAR CODE: 0117047	Total ₹ 4,995.00

HSN/ SAC	Taxable amount	CGST		SGST		Total Tax Amount
		Rate	Amount	Rate	Amount	
8501	₹ 2,716.10	9%	₹ 244.45	9%	₹ 244.45	₹ 488.90
8511	₹ 1,516.95	9%	₹ 136.53	9%	₹ 136.53	₹ 273.05
Total	₹ 4,233.05		₹ 380.97		₹ 380.97	₹ 761.95

Bank Details Name : UNION BANK OF INDIA, KOTTAYAM Account No. : 361201010035646 IFSC code : UBIN0536121 Account holder's name : CLOUD99 BIOTECH	Terms and conditions Thanks for doing business with us!	For : CLOUD99 BIOTECH  Authorized Signatory
--	---	--

RE ENQUIRY - HAEMO MIXER - GMHC PEROORKADA (148060 / 0117047)

Messages

Thu, 26 Dec 2024 at 12:54 p

From: Kasyep P V <kasyepcyrix@gmail.com>
To: CLOUD 99 <cloud99biotech@gmail.com>
Cc: Afsal Yunuze <zm1.klbemp@cyrix.in>

Dear Sir,
Please provide a quotation for the Haemo mixer at GMHC Peroorkada
Equipment Name : Haemo Mixer
Make : Transasia
Model : SR 20A

Parts Required
1.AC Motor
2.Gear Assembly

Thanks & Regards,

KASYEP PV
District Incharge ,Trivandrum
CYRIX HEALTHCARE Pvt.Ltd
KLBEMP
Ph : +91 7593847134
kasyepcyrix@gmail.com

Thu, 26 Dec 2024 at 5:52 p

From: CLOUD 99 <cloud99biotech@gmail.com>
To: Kasyep P V <kasyepcyrix@gmail.com>

Dear Sir
Please find the attachment for a quotation for the Haemo mixer at GMHC Peroorkada

With Warm Regards


SUSAN JOHN
Cloud99 Biotech
Medical Engineering Works
Room No:17, Vellappally Building
Vellappally Lane
Opp. Kottayam District Hospital
Kottayam 686 001
Kerala, India
Ph: 9744473043
email: cloud99biotech@gmail.com

[Quoted text hidden]

MEDICAL
EQUIPMENT
MAINTENANCE
PROGRAMME

SR-20 A

CYRIX Biomedical Equipment Maintenance
Programme (BEMP)
TOLL FREE NO: 1800 - 425-7669
BAR CODE - (8004-890615225)
0117047



CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE
PROGRAMME (BEMP)

TOLL FREE NO: 1800 - 425-7669



KMSCL
KARNATAKA MEDICAL SOCIETY

**BAR CODE - (8004-890615225)
0117047**

